

Continuing Calibration Report for SequenceID : FC112724AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067881.D

Aliphatic C9-C12	9987946.000	60.000	3.186	6.461	166465.767	165791.359	-0.407
Aliphatic C12-C16	6942773.000	40.000	6.462	9.849	173569.325	170384.422	-1.869
Aliphatic C16-C21	10227182.000	60.000	9.850	13.205	170453.033	169621.553	-0.490
Aliphatic C21-C28	12690817.000	80.000	13.206	16.857	158635.213	157280.060	-0.862
Aliphatic C28-C40	16576371.000	120.000	16.858	21.700	138136.425	121111.336	-14.057
Aliphatic EPH	56425089.000	360.000	3.186	21.700	156736.358	150155.324	-4.383