

Continuing Calibration Report for SequenceID : FD112724AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048821.D

Aromatic C10-C12	8466989.000	40.000	4.087	5.803	211674.725	192017.544	-10.237
Aromatic C12-C16	12905784.000	60.000	5.804	8.410	215096.400	192458.977	-11.762
Aromatic C16-C21	15378400.000	80.000	8.411	12.673	192230.000	185804.816	-3.458
Aromatic C21-C36	30610203.000	180.000	12.674	18.080	170056.683	157231.056	-8.157
Aromatic EPH	67361376.000	360.000	4.087	18.080	187114.933	173317.266	-7.961