

SOLID EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P5002 **SAS No :** P5002 **SDG No:** P5002
Sample No : P5002-01MS **Datafile:** FC067876.D
Client ID : EX-7-TPH-1MS

COMPOUND	SPIKE ADDED mg/kg	SAMPLE CONCENTRATION mg/kg	MS/MSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	11.8	66.3	82.2	135		(40-140)
Aliphatic C12-C16	7.9	367	470	1305	*	(40-140)
Aliphatic C16-C21	11.8	350	448	823	*	(40-140)
Aliphatic C21-C28	15.7	119	158	245	*	(40-140)
Aliphatic C28-C40	23.6	74.9	108	140	*	(40-140)