

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P5002 **SAS No :** P5002 **SDG No:** P5002
Sample No : PB165286BS **Datafile:** FC067872.D
Client ID : PB165286BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10	4.54	45		(40-140)
Aliphatic C12-C16	6.7	4.73	71		(40-140)
Aliphatic C16-C21	10	8.31	83		(40-140)
Aliphatic C21-C28	13.3	11.0	83		(40-140)
Aliphatic C28-C40	20.0	18.2	91		(40-140)