

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P5002 **SAS No :** P5002 **SDG No:** P5002
Sample No : PB165286BSD **Datafile:** FC067873.D
Client ID : PB165286BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	10.0	4.55	46		0.25 (40-140)	50
Aliphatic C12-C16	6.7	4.73	71		0.0371 (40-140)	50
Aliphatic C16-C21	10.0	8.38	84		0.75 (40-140)	50
Aliphatic C21-C28	13.3	11.2	84		1.7 (40-140)	50
Aliphatic C28-C40	20.0	18.6	93		2.1 (40-140)	50