

## Prep Standard - Chemical Standard Summary

**Order ID :** P5002

**Test :** Gasoline Range Organics

**Prepbatch ID :**

**Sequence ID/Qc Batch ID:** FB112624,

**Standard ID :**

PP23534,PP23535,PP23538,PP24008,PP24009,PP24010,PP24011,PP24012,PP24013,PP24071,PP24072,

**Chemical ID :**

P11121,P9826,V11252,V14143,W3112,

## Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
231	10 PPM GRO STD 1ST SOURCE	<a href="#">PP23534</a>	07/29/2024	01/22/2025	Yogesh Patel	None	None	Ankita Jodhani
07/30/2024								

**FROM** 0.11100ml of P9826 + 9.89000ml of V14143 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
233	10 PPM GRO STD 2nd SOURCE	<a href="#">PP23535</a>	07/29/2024	01/22/2025	Yogesh Patel	None	None	Ankita Jodhani
07/30/2024								

**FROM** 0.11100ml of P11121 + 9.89000ml of V14143 = Final Quantity: 10.000 ml

## Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3619	25 PPM AAA-TFT Surg	<a href="#">PP23538</a>	07/29/2024	01/22/2025	Yogesh Patel	None	None	Ankita Jodhani
07/30/2024								

**FROM** 0.10000ml of V11252 + 9.90000ml of V14143 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
238	5 PPB ICC GRO STD	<a href="#">PP24008</a>	11/21/2024	01/22/2025	Yogesh Patel	None	None	Ankita Jodhani
11/21/2024								

**FROM** 5.00000ml of W3112 + 0.00100ml of PP23538 + 0.00250ml of PP23534 = Final Quantity: 5.004 ml

## Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
237	10 PPB ICC GRO STD	<a href="#">PP24009</a>	11/21/2024	01/22/2025	Yogesh Patel	None	None	Ankita Jodhani
								11/21/2024

**FROM** 5.00000ml of W3112 + 0.00200ml of PP23538 + 0.00500ml of PP23534 = Final Quantity: 5.007 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
239	20 PPB ICC GRO STD	<a href="#">PP24010</a>	11/21/2024	01/22/2025	Yogesh Patel	None	None	Ankita Jodhani
								11/21/2024

**FROM** 5.00000ml of W3112 + 0.00400ml of PP23538 + 0.01000ml of PP23534 = Final Quantity: 5.014 ml

## Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
235	50 PPB ICC GRO STD	<a href="#">PP24011</a>	11/21/2024	01/22/2025	Yogesh Patel	None	None	Ankita Jodhani
								11/21/2024

**FROM** 5.00000ml of W3112 + 0.01000ml of PP23538 + 0.02500ml of PP23534 = Final Quantity: 5.035 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
234	100 PPB ICC GRO STD	<a href="#">PP24012</a>	11/21/2024	01/22/2025	Yogesh Patel	None	None	Ankita Jodhani
								11/21/2024

**FROM** 5.00000ml of W3112 + 0.02000ml of PP23538 + 0.05000ml of PP23534 = Final Quantity: 5.070 ml

## Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
240	20 PPB ICV GRO STD	<a href="#">PP24013</a>	11/21/2024	01/22/2025	Yogesh Patel	None	None	Ankita Jodhani
								11/21/2024

**FROM** 5.00000ml of W3112 + 0.00400ml of PP23538 + 0.01000ml of PP23535 = Final Quantity: 5.014 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
241	20 PPB CCC GRO STD	<a href="#">PP24071</a>	11/26/2024	01/22/2025	Yogesh Patel	None	None	Ankita Jodhani
								12/02/2024

**FROM** 5.00000ml of W3112 + 0.00400ml of PP23538 + 0.01000ml of PP23534 = Final Quantity: 5.014 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
241	20 PPB CCC GRO STD	<a href="#">PP24072</a>	11/26/2024	01/22/2025	Yogesh Patel	None	None	Ankita Jodhani 12/02/2024
<u>FROM</u>	5.00000ml of W3112 + 0.00400ml of PP23538 + 0.01000ml of PP23534 = Final Quantity: 5.014 ml							

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30065 / GRO Mix (EPA)	A0161776	01/25/2025	07/25/2024 / yogesh	02/10/2021 / Sohil	P11121

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30065 / GRO Mix (EPA)	A0155991	01/25/2025	07/25/2024 / yogesh	09/11/2020 / DHAVAL	P9826

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30068 / VOA Mix, a, a, a-trifluorotoluene 2500ug/ml, P&T methanol, 1ml	A0158026	05/31/2028	11/27/2023 / yogesh	09/11/2020 / DHAVAL	V11252

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	01/22/2025	07/22/2024 / SAM	02/06/2024 / SAM	V14143

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112





110 Benner Circle  
Bellefonte, PA 16823-8812  
Tel: (800)356-1688  
Fax: (814)353-1309

www.restek.com

# CERTIFIED REFERENCE MATERIAL

## Certificate of Analysis



### FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

*This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.*

**Catalog No. :** 30065 **Lot No.:** A0161776

**Description :** Gasoline Range Organics Mix (EPA)

Gasoline Range Organics Mix (EPA) 500 - 1500µg/mL, P&T Methanol, 1mL/ampul

**Container Size :** 2 mL **Pkg Amt:** > 1 mL

**Expiration Date :** July 31, 2027 **Storage:** 0°C or colder

### CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	2-Methylpentane CAS # 107-83-5 (Lot MKCB1674V) Purity 99%	1,507.0 µg/mL	+/- 8.9511 µg/mL Gravimetric +/- 84.5158 µg/mL Unstressed +/- 86.4925 µg/mL Stressed
2	2,2,4-Trimethylpentane ( isooctane ) CAS # 540-84-1 (Lot SHBF8066V) Purity 99%	1,511.0 µg/mL	+/- 8.9749 µg/mL Gravimetric +/- 84.7402 µg/mL Unstressed +/- 86.7221 µg/mL Stressed
3	n-Heptane (C7) CAS # 142-82-5 (Lot SHBK8626) Purity 98%	498.8 µg/mL	+/- 2.9628 µg/mL Gravimetric +/- 27.9749 µg/mL Unstressed +/- 28.6292 µg/mL Stressed
4	Benzene CAS # 71-43-2 (Lot SHBK5679) Purity 99%	500.0 µg/mL	+/- 2.9698 µg/mL Gravimetric +/- 28.0411 µg/mL Unstressed +/- 28.6969 µg/mL Stressed
5	Toluene CAS # 108-88-3 (Lot MKCH9232) Purity 99%	1,510.0 µg/mL	+/- 8.9689 µg/mL Gravimetric +/- 84.6841 µg/mL Unstressed +/- 86.6647 µg/mL Stressed
6	Ethylbenzene CAS # 100-41-4 (Lot SHBL0706) Purity 99%	504.0 µg/mL	+/- 2.9936 µg/mL Gravimetric +/- 28.2654 µg/mL Unstressed +/- 28.9265 µg/mL Stressed
7	m-Xylene CAS # 108-38-3 (Lot SHBL0265) Purity 99%	1,005.0 µg/mL	+/- 5.9694 µg/mL Gravimetric +/- 56.3626 µg/mL Unstressed +/- 57.6808 µg/mL Stressed

8	o-Xylene		(Lot SHBK7739)	1,007.0	µg/mL	+/-	5.9813	µg/mL	Gravimetric
	CAS #	95-47-6				+/-	56.4747	µg/mL	Unstressed
	Purity	99%				+/-	57.7956	µg/mL	Stressed
9	1,2,4-Trimethylbenzene		(Lot WXBC4246V)	1,008.0	µg/mL	+/-	5.9872	µg/mL	Gravimetric
	CAS #	95-63-6				+/-	56.5308	µg/mL	Unstressed
	Purity	99%				+/-	57.8530	µg/mL	Stressed
Solvent:	P&T Methanol								
	CAS #	67-56-1							
	Purity	99%							

**Column:**  
105m x 0.53mm x 3.0µm  
Rtx-502.2 (cat.#10910)


**Carrier Gas:**  
hydrogen-constant pressure 11.0 psi.

**Temp. Program:**  
40°C (hold 2 min.) to 240°C  
@ 8°C/min. (hold 5 min.)

**Inj. Temp:**  
200°C

**Det. Temp:**  
250°C

**Det. Type:**  
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

*Cydnei L. Crust*  
Cydnei L. Crust - Mix Technician

**Date Mixed:** 15-Jun-2020

**Balance:** B251644995

*Fang-Yun Lo*  
Fang-Yun Lo - QC Analyst

**Date Passed:** 17-Jun-2020

Manufactured under Restek's ISO 9001:2015  
Registered Quality System  
Certificate #FM 80397

## General Certified Reference Material Notes

### Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

### Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

### Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified combined stressed uncertainty value ( includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k \sqrt{U_{gravimetric}^2 + U_{homogeneity}^2 + U_{storage\ stability}^2 + U_{shipping\ stability}^2}$$

$k$  is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at [www.restek.com/Contact-Us](http://www.restek.com/Contact-Us) for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at [www.restek.com/Contact-Us](http://www.restek.com/Contact-Us).
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

### Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

### Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

Methanol  
ULTRA RESI-ANALYZED  
For Purge and Trap Analysis



Material No.: 9077-02  
Batch No.: 22L0562016  
Manufactured Date: 2022-10-26  
Expiration Date: 2025-10-25  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
Assay (CH <sub>3</sub> OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid (μeq/g)	≤ 0.3	0.2
Titration Base (μeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory, Research, or Manufacturing Use  
Performance Tested for Use in EPA Methods  
500 Series for Drinking Water  
600 Series for Wastewater  
846 for Solid Waste

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Ethier  
Vice President Global Quality