

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133649

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	12/02/2024	09:35
2	P4972-01		1	14	100.00	>212.0	>212.0	12/02/2024	10:05
3	P4972-02		1	11	100.00	>212.0	>212.0	12/02/2024	10:35
4	P4972-03		1	11	100.00	>212.0	>212.0	12/02/2024	12:10
5	P4972-05		1	14	100.00	>212.0	>212.0	12/02/2024	12:40
6	P5006-18		1	13	100.00	>212.0	>212.0	12/02/2024	13:10
7	P5006-18DUP		1	13	100.00	>212.0	>212.0	12/02/2024	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB133649.

WorkList Name : fp-12-2 WorkList ID : 185881 Department : Wet-Chemistry Date : 12-02-2024 08:50:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4972-01	111524-A	Water	Flash Point	Cool 4 deg C	PSEG03	L41	11/22/2024	1010B
P4972-02	111524-B	Water	Flash Point	Cool 4 deg C	PSEG03	L41	11/22/2024	1010B
P4972-03	912-A	Water	Flash Point	Cool 4 deg C	PSEG03	L41	11/22/2024	1010B
P4972-05	628-F	Water	Flash Point	Cool 4 deg C	PSEG03	L41	11/22/2024	1010B
P5006-18	TW-WTS-02	Water	Flash Point	Cool 4 deg C	ENTA05	L51	11/25/2024	1010B

Date/Time 12/02/2024 09:25
 Raw Sample Received by: RM
 Raw Sample Relinquished by: JH WELC 1

Date/Time 12/02/2024 14:15
 Raw Sample Received by: RM
 Raw Sample Relinquished by: RM WELC