



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133640

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	11/26/2024
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	11/26/2024
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.484	0.5	97	90-110	11/26/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133643

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10	10	100	90-110	11/18/2024
Chloride	mg/L	3	3	100	90-110	11/18/2024
Fluoride	mg/L	2	2	100	90-110	11/18/2024
Nitrite	mg/L	3	3	100	90-110	11/18/2024
Nitrate	mg/L	2.5	2.5	100	90-110	11/18/2024
Sulfate	mg/L	14.6	15	97	90-110	11/18/2024
Orthophosphate as P	mg/L	5	5	100	90-110	11/18/2024
Sample ID: CCV1						
Bromide	mg/L	10	10	100	90-110	11/26/2024
Chloride	mg/L	2.9	3	97	90-110	11/26/2024
Fluoride	mg/L	2	2	100	90-110	11/26/2024
Nitrite	mg/L	3	3	100	90-110	11/26/2024
Nitrate	mg/L	2.5	2.5	100	90-110	11/26/2024
Sulfate	mg/L	14.2	15	95	90-110	11/26/2024
Orthophosphate as P	mg/L	5	5	100	90-110	11/26/2024
Sample ID: CCV2						
Bromide	mg/L	10	10	100	90-110	11/26/2024
Chloride	mg/L	2.9	3	97	90-110	11/26/2024
Fluoride	mg/L	1.9	2	95	90-110	11/26/2024
Nitrite	mg/L	3	3	100	90-110	11/26/2024
Nitrate	mg/L	2.5	2.5	100	90-110	11/26/2024
Sulfate	mg/L	14.1	15	94	90-110	11/26/2024
Orthophosphate as P	mg/L	4.9	5	98	90-110	11/26/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133647

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	11/26/2024
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	11/26/2024
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	11/26/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133649

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV						
Flash Point	o F	81.7	81	101	78-84	12/02/2024

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133703

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: TKN	ICV1 mg/L	4.7	5	94	90-110	12/03/2024
Sample ID: TKN	CCV1 mg/L	5	5	100	90-110	12/03/2024
Sample ID: TKN	CCV2 mg/L	4.9	5	98	90-110	12/03/2024