



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133640

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	11/26/2024
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	11/26/2024
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0027	0.01	11/26/2024

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133643

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	11/18/2024
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	11/18/2024
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	11/18/2024
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	11/18/2024
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	11/18/2024
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	11/18/2024
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	11/18/2024
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	11/26/2024
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	11/26/2024
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	11/26/2024
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	11/26/2024
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	11/26/2024
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	11/26/2024
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	11/26/2024
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	11/26/2024
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	11/26/2024
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	11/26/2024
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	11/26/2024
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	11/26/2024
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	11/26/2024
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	11/26/2024

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB133703

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TKN	mg/L	< 0.2500	0.2500	U	0.18	0.5	12/03/2024
Sample ID: CCB1 TKN	mg/L	< 0.2500	0.2500	U	0.18	0.5	12/03/2024
Sample ID: CCB2 TKN	mg/L	< 0.2500	0.2500	U	0.18	0.5	12/03/2024