

Matrix Spike Summary

Client:	ENTACT	SDG No.:	P5006
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	P5006-18
Client ID:	TW-WTS-02MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	7.50		1.50	J	10	1	60	*	11/26/2024
Chloride	mg/L	80-120	255	OR	261	OR	3	1	-200	*	11/26/2024
Fluoride	mg/L	80-120	0.67		1.90		2	1	-62	*	11/26/2024
Nitrite	mg/L	80-120	1.80		0.011	U	3	1	60	*	11/26/2024
Nitrate	mg/L	80-120	11.1	OR	9.80	OR	2.5	1	52	*	11/26/2024
Sulfate	mg/L	80-120	232	OR	229	OR	15	1	20	*	11/26/2024
Orthophosphate as P	mg/L	80-120	0.78	J	0.079	U	5	1	16	*	11/26/2024

Matrix Spike Summary

Client:	ENTACT	SDG No.:	P5006
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	P5006-18
Client ID:	TW-WTS-02MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	7.50		1.50	J	10	1	60	*	11/26/2024
Chloride	mg/L	80-120	254	OR	261	OR	3	1	-233	*	11/26/2024
Fluoride	mg/L	80-120	3.10		1.90		2	1	60	*	11/26/2024
Nitrite	mg/L	80-120	1.80		0.011	U	3	1	60	*	11/26/2024
Nitrate	mg/L	80-120	11.1	OR	9.80	OR	2.5	1	52	*	11/26/2024
Sulfate	mg/L	80-120	232	OR	229	OR	15	1	20	*	11/26/2024
Orthophosphate as P	mg/L	80-120	3.10		0.079	U	5	1	62	*	11/26/2024