

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133640

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	lb133640BS							
Hexavalent Chromium	mg/L	0.5	0.51		102	1	90-111	11/26/2024

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133643

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133643BSW							
Bromide	mg/L	10	10.0		100	1	90-110	11/26/2024
Chloride	mg/L	3	2.90		97	1	90-110	11/26/2024
Fluoride	mg/L	2	2.00		100	1	90-110	11/26/2024
Nitrite	mg/L	3	3.00		100	1	90-110	11/26/2024
Nitrate	mg/L	2.5	2.50		100	1	90-110	11/26/2024
Sulfate	mg/L	15	14.2		95	1	90-110	11/26/2024
Orthophosphate as P	mg/L	5	5.10		102	1	90-110	11/26/2024

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133658

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133658BS							
TSS	mg/L	550	536		98	1	90-110	11/27/2024

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133668

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133668BS							
BOD5	mg/L	198	194		98	1	84.6-115.4	11/27/2024

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: P5006

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB133703

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165319BS							
TKN	mg/L	5	4.60		92	1	90-110	12/03/2024