



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** P5006
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P5006 **SAS No.:** P5006
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.93	6.0	99	40 - 160	P	12/02/2024	11:34	LB133689
	Copper	22.8	20.0	114	40 - 160	P	12/02/2024	11:34	LB133689
	Lead	12.4	12.0	104	40 - 160	P	12/02/2024	11:34	LB133689
	Nickel	40.2	40.0	101	40 - 160	P	12/02/2024	11:34	LB133689
	Zinc	44.2	40.0	110	40 - 160	P	12/02/2024	11:34	LB133689
CRA	Mercury	0.17	0.2	83	40 - 160	CV	12/03/2024	11:07	LB133702