

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: ENTACT **SDG No.:** P5006
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P5006 **SAS No.:** P5006
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	6.75	1.0	675	-5	7	12/02/2024	11:38	LB133689
	Copper	5.49	2.0	274	-18	22	12/02/2024	11:38	LB133689
	Lead	7.95			-12	12	12/02/2024	11:38	LB133689
	Nickel	2.45	2.0	122	-38	42	12/02/2024	11:38	LB133689
	Zinc	5.77			-40	40	12/02/2024	11:38	LB133689
ICSAB01	Cadmium	1010	972	104	826	1120	12/02/2024	11:42	LB133689
	Copper	486	511	95	434	588	12/02/2024	11:42	LB133689
	Lead	55.3	49.0	113	37	61	12/02/2024	11:42	LB133689
	Nickel	995	954	104	810	1100	12/02/2024	11:42	LB133689
	Zinc	1070	952	112	809	1095	12/02/2024	11:42	LB133689