

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** P5006
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P5006 **SAS No.:** P5006
Matrix: Water **Sample ID:** P4972-05 **Client ID:** 628-FDUP
Percent Solids for Sample: NA **Duplicate ID** P4972-05DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	2.63	JD	2.60	JD	1		P
Copper	ug/L	20	220	D	199	D	10		P
Lead	ug/L	20	35.5	D	32.9	D	8		P
Nickel	ug/L	20	23.3	JD	21.4	JD	9		P
Zinc	ug/L	20	2650	D	2400	D	10		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** P5006
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P5006 **SAS No.:** P5006
Matrix: Water **Sample ID:** P4972-05MS **Client ID:** 628-FMSD
Percent Solids for Sample: NA **Duplicate ID** P4972-05MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	92.3	D	82.2	D	12		P
Copper	ug/L	20	360	D	325	D	10		P
Lead	ug/L	20	434	D	390	D	11		P
Nickel	ug/L	20	250	D	221	D	12		P
Zinc	ug/L	20	2720	D	2410	D	12		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** P5006
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P5006 **SAS No.:** P5006
Matrix: Water **Sample ID:** P5031-01 **Client ID:** 286127DUP
Percent Solids for Sample: NA **Duplicate ID** P5031-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.25		0.21		16		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>ENTACT</u>	Level:	<u>LOW</u>	SDG No.:	<u>P5006</u>
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P5006</u>
Matrix:	<u>Water</u>	Sample ID:	<u>P5031-01MS</u>	Client ID:	<u>286127MSD</u>
Percent Solids for Sample:	<u>NA</u>	Duplicate ID	<u>P5031-01MSD</u>	Percent Solids for Spike Sample:	<u>NA</u>

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	4.22		4.29		2		CV