

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ENTA05

Lab Code: CHEM **Case No.:** P5006 **SAS No.:** P5006 **SDG NO.:** P5006

Instrument ID: ECD_O **Date(s) Analyzed:** 11/18/2024 11/18/2024

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.57	4.47	4.67	108422000
		2	4.66	4.56	4.76	80468200
		3	4.73	4.63	4.83	233076000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.73	4.63	4.83	171958000
		2	5.26	5.16	5.36	94164000
		3	5.55	5.45	5.65	161766000
		4	5.71	5.61	5.81	76393000
		5	5.80	5.70	5.90	52874400
Aroclor-1262	500	1	7.76	7.66	7.86	323292000
		2	8.30	8.20	8.40	449988000
		3	8.61	8.51	8.71	288986000
		4	8.70	8.60	8.80	223132000
		5	9.34	9.24	9.44	152022000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ENTA05

Lab Code: CHEM **Case No.:** P5006 **SAS No.:** P5006 **SDG NO.:** P5006

Instrument ID: ECD_O **Date(s) Analyzed:** 11/18/2024 11/18/2024

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	3.85	3.75	3.95	35476800
		2	3.94	3.84	4.04	29137000
		3	4.01	3.91	4.11	89032800
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.01	3.91	4.11	63710600
		2	4.74	4.64	4.84	64827200
		3	4.91	4.81	5.01	33220400
		4	5.00	4.90	5.10	30291200
		5	5.17	5.07	5.27	31376600
Aroclor-1262	500	1	6.75	6.65	6.85	227544000
		2	7.25	7.15	7.35	426902000
		3	7.54	7.44	7.64	165111000
		4	7.60	7.50	7.70	317924000
		5	8.09	7.99	8.19	148021000