

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P5006 SAS No.: P5006 SDG NO.: P5006

Continuing Calib Date: 11/27/2024 Initial Calibration Date(s): 11/18/2024 11/18/2024

Continuing Calib Time: 13:39 Initial Calibration Time(s): 15:32 22:51

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.39       | 6.38      | 6.28      | 6.48  | -0.01      |
| Aroclor-1254-2       | (2) | 6.61       | 6.60      | 6.50      | 6.70  | -0.01      |
| Aroclor-1254-3       | (3) | 6.97       | 6.97      | 6.87      | 7.07  | 0.00       |
| Aroclor-1254-4       | (4) | 7.26       | 7.25      | 7.15      | 7.35  | -0.01      |
| Aroclor-1254-5       | (5) | 7.68       | 7.68      | 7.58      | 7.78  | 0.00       |
| Tetrachloro-m-xylene |     | 4.37       | 4.37      | 4.27      | 4.47  | 0.00       |
| Decachlorobiphenyl   |     | 10.10      | 10.09     | 9.99      | 10.19 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P5006 SAS No.: P5006 SDG NO.: P5006

Continuing Calib Date: 11/27/2024 Initial Calibration Date(s): 11/18/2024 11/18/2024

Continuing Calib Time: 13:39 Initial Calibration Time(s): 15:32 22:51

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.52       | 5.52      | 5.42      | 5.62 | 0.00       |
| Aroclor-1254-2       | (2) | 5.67       | 5.67      | 5.57      | 5.77 | 0.01       |
| Aroclor-1254-3       | (3) | 6.07       | 6.07      | 5.97      | 6.17 | 0.00       |
| Aroclor-1254-4       | (4) | 6.30       | 6.30      | 6.20      | 6.40 | 0.00       |
| Aroclor-1254-5       | (5) | 6.71       | 6.72      | 6.62      | 6.82 | 0.01       |
| Tetrachloro-m-xylene |     | 3.64       | 3.64      | 3.54      | 3.74 | 0.00       |
| Decachlorobiphenyl   |     | 8.63       | 8.63      | 8.53      | 8.73 | 0.00       |