

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P5006 SAS No.: P5006 SDG NO.: P5006

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL01 Date Analyzed: 11/27/2024

Lab Sample No.: AR1660CCC500 Data File : PO108250.D Time Analyzed: 09:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.525	5.425	5.625	539.290	500.000	7.9
Aroclor-1016-2	5.548	5.447	5.647	531.110	500.000	6.2
Aroclor-1016-3	5.611	5.510	5.710	511.980	500.000	2.4
Aroclor-1016-4	5.708	5.607	5.807	528.790	500.000	5.8
Aroclor-1016-5	6.004	5.904	6.104	529.230	500.000	5.8
Aroclor-1260-1	7.138	7.037	7.237	533.860	500.000	6.8
Aroclor-1260-2	7.396	7.295	7.495	563.180	500.000	12.6
Aroclor-1260-3	7.759	7.658	7.858	562.890	500.000	12.6
Aroclor-1260-4	7.984	7.884	8.084	569.090	500.000	13.8
Aroclor-1260-5	8.300	8.198	8.398	575.220	500.000	15.0
Decachlorobiphenyl	10.094	9.991	10.191	63.230	50.000	26.5
Tetrachloro-m-xylene	4.371	4.270	4.470	54.330	50.000	8.7

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P5006 SAS No.: P5006 SDG NO.: P5006

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL01 Date Analyzed: 11/27/2024

Lab Sample No.: AR1660CCC500 Data File : PO108250.D Time Analyzed: 09:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.718	4.620	4.820	568.510	500.000	13.7
Aroclor-1016-2	4.737	4.639	4.839	574.820	500.000	15.0
Aroclor-1016-3	4.912	4.815	5.015	566.780	500.000	13.4
Aroclor-1016-4	4.954	4.856	5.056	553.980	500.000	10.8
Aroclor-1016-5	5.167	5.070	5.270	558.820	500.000	11.8
Aroclor-1260-1	6.198	6.100	6.300	550.970	500.000	10.2
Aroclor-1260-2	6.385	6.288	6.488	558.020	500.000	11.6
Aroclor-1260-3	6.539	6.441	6.641	538.140	500.000	7.6
Aroclor-1260-4	7.009	6.912	7.112	536.270	500.000	7.3
Aroclor-1260-5	7.250	7.154	7.354	567.560	500.000	13.5
Decachlorobiphenyl	8.630	8.535	8.735	58.850	50.000	17.7
Tetrachloro-m-xylene	3.637	3.539	3.739	57.110	50.000	14.2