

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P5006 SAS No.: P5006 SDG NO.: P5006

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL02 Date Analyzed: 11/27/2024

Lab Sample No.: AR1660CCC500 Data File : PO108262.D Time Analyzed: 12:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.525	5.425	5.625	474.050	500.000	-5.2
Aroclor-1016-2	5.547	5.447	5.647	471.160	500.000	-5.8
Aroclor-1016-3	5.609	5.510	5.710	462.330	500.000	-7.5
Aroclor-1016-4	5.707	5.607	5.807	474.140	500.000	-5.2
Aroclor-1016-5	6.003	5.904	6.104	468.360	500.000	-6.3
Aroclor-1260-1	7.137	7.037	7.237	456.300	500.000	-8.7
Aroclor-1260-2	7.395	7.295	7.495	483.560	500.000	-3.3
Aroclor-1260-3	7.759	7.658	7.858	475.990	500.000	-4.8
Aroclor-1260-4	7.985	7.884	8.084	477.060	500.000	-4.6
Aroclor-1260-5	8.301	8.198	8.398	489.880	500.000	-2.0
Decachlorobiphenyl	10.094	9.991	10.191	49.970	50.000	-0.1
Tetrachloro-m-xylene	4.369	4.270	4.470	46.370	50.000	-7.3

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P5006 SAS No.: P5006 SDG NO.: P5006

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL02 Date Analyzed: 11/27/2024

Lab Sample No.: AR1660CCC500 Data File : PO108262.D Time Analyzed: 12:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.717	4.620	4.820	492.950	500.000	-1.4
Aroclor-1016-2	4.736	4.639	4.839	496.530	500.000	-0.7
Aroclor-1016-3	4.912	4.815	5.015	495.220	500.000	-1.0
Aroclor-1016-4	4.954	4.856	5.056	492.950	500.000	-1.4
Aroclor-1016-5	5.167	5.070	5.270	491.980	500.000	-1.6
Aroclor-1260-1	6.198	6.100	6.300	469.140	500.000	-6.2
Aroclor-1260-2	6.385	6.288	6.488	467.890	500.000	-6.4
Aroclor-1260-3	6.539	6.441	6.641	461.660	500.000	-7.7
Aroclor-1260-4	7.009	6.912	7.112	455.940	500.000	-8.8
Aroclor-1260-5	7.251	7.154	7.354	445.120	500.000	-11.0
Decachlorobiphenyl	8.631	8.535	8.735	42.120	50.000	-15.8
Tetrachloro-m-xylene	3.637	3.539	3.739	51.190	50.000	2.4