

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P5006 SAS No.: P5006 SDG NO.: P5006

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL03 Date Analyzed: 11/27/2024

Lab Sample No.: AR1660CCC500 Data File : PO108272.D Time Analyzed: 18:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.525	5.425	5.625	446.960	500.000	-10.6
Aroclor-1016-2	5.546	5.447	5.647	443.230	500.000	-11.4
Aroclor-1016-3	5.609	5.510	5.710	430.270	500.000	-13.9
Aroclor-1016-4	5.707	5.607	5.807	433.840	500.000	-13.2
Aroclor-1016-5	6.003	5.904	6.104	433.390	500.000	-13.3
Aroclor-1260-1	7.137	7.037	7.237	467.620	500.000	-6.5
Aroclor-1260-2	7.395	7.295	7.495	477.190	500.000	-4.6
Aroclor-1260-3	7.759	7.658	7.858	444.880	500.000	-11.0
Aroclor-1260-4	7.984	7.884	8.084	448.170	500.000	-10.4
Aroclor-1260-5	8.300	8.198	8.398	462.210	500.000	-7.6
Decachlorobiphenyl	10.093	9.991	10.191	46.440	50.000	-7.1
Tetrachloro-m-xylene	4.369	4.270	4.470	49.910	50.000	-0.2

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: P5006 SAS No.: P5006 SDG NO.: P5006

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/18/2024 11/18/2024

Client Sample No.: CCAL03 Date Analyzed: 11/27/2024

Lab Sample No.: AR1660CCC500 Data File : PO108272.D Time Analyzed: 18:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.718	4.620	4.820	508.450	500.000	1.7
Aroclor-1016-2	4.736	4.639	4.839	511.630	500.000	2.3
Aroclor-1016-3	4.912	4.815	5.015	512.220	500.000	2.4
Aroclor-1016-4	4.954	4.856	5.056	505.030	500.000	1.0
Aroclor-1016-5	5.167	5.070	5.270	514.140	500.000	2.8
Aroclor-1260-1	6.198	6.100	6.300	501.920	500.000	0.4
Aroclor-1260-2	6.385	6.288	6.488	489.640	500.000	-2.1
Aroclor-1260-3	6.539	6.441	6.641	474.360	500.000	-5.1
Aroclor-1260-4	7.010	6.912	7.112	446.250	500.000	-10.8
Aroclor-1260-5	7.251	7.154	7.354	430.890	500.000	-13.8
Decachlorobiphenyl	8.630	8.535	8.735	35.870	50.000	-28.3
Tetrachloro-m-xylene	3.637	3.539	3.739	51.540	50.000	3.1