

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120224\
 Data File : BF140681.D
 Acq On : 02 Dec 2024 10:22
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 02 11:01:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	78	0.00
2	1,4-Dioxane	40.000	39.241	1.9	77	-0.02
3	Pyridine	40.000	41.355	-3.4	73	-0.02
4	n-Nitrosodimethylamine	40.000	36.032	9.9	69	-0.02
5 S	2-Fluorophenol	80.000	78.340	2.1	76	0.00
6	Aniline	40.000	42.754	-6.9	72	0.00
7 S	Phenol-d6	80.000	77.993	2.5	73	0.00
8	2-Chlorophenol	40.000	39.590	1.0	75	0.00
9	Benzaldehyde	40.000	37.711	5.7	81	0.00
10 C	Phenol	40.000	39.907	0.2	73	0.00
11	bis(2-Chloroethyl)ether	40.000	38.644	3.4	73	0.00
12	1,3-Dichlorobenzene	40.000	39.012	2.5	77	0.00
13 C	1,4-Dichlorobenzene	40.000	38.799	3.0	76	0.00
14	1,2-Dichlorobenzene	40.000	39.430	1.4	76	0.00
15	Benzyl Alcohol	40.000	39.344	1.6	72	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.120	4.7	72	0.00
17	2-Methylphenol	40.000	39.183	2.0	74	0.00
18	Hexachloroethane	40.000	38.253	4.4	74	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.674	5.8	71	0.00
20	3+4-Methylphenols	40.000	39.158	2.1	73	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	73	0.00
22	Acetophenone	40.000	39.582	1.0	74	0.00
23 S	Nitrobenzene-d5	80.000	77.900	2.6	71	0.00
24	Nitrobenzene	40.000	39.523	1.2	73	0.00
25	Isophorone	40.000	39.291	1.8	71	0.00
26 C	2-Nitrophenol	40.000	40.392	-1.0	74	0.00
27	2,4-Dimethylphenol	40.000	42.539	-6.3	75	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.982	2.5	72	0.00
29 C	2,4-Dichlorophenol	40.000	40.122	-0.3	74	0.00
30	1,2,4-Trichlorobenzene	40.000	39.084	2.3	73	0.00
31	Naphthalene	40.000	39.627	0.9	74	0.00
32	Benzoic acid	40.000	34.616	13.5	62	0.00
33	4-Chloroaniline	40.000	42.251	-5.6	74	0.00
34 C	Hexachlorobutadiene	40.000	39.335	1.7	73	0.00
35	Caprolactam	40.000	40.334	-0.8	72	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.440	1.4	70	0.00
37	2-Methylnaphthalene	40.000	40.053	-0.1	74	0.00
38	1-Methylnaphthalene	40.000	39.495	1.3	73	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	73	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.254	1.9	74	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.441	1.4	74	0.00
42 S	2,4,6-Tribromophenol	80.000	78.824	1.5	71	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.903	2.7	71	0.00
44	2,4,5-Trichlorophenol	40.000	39.418	1.5	71	0.00
45 S	2-Fluorobiphenyl	80.000	77.648	2.9	73	0.00
46	1,1'-Biphenyl	40.000	39.480	1.3	74	0.00
47	2-Chloronaphthalene	40.000	39.690	0.8	74	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120224\
 Data File : BF140681.D
 Acq On : 02 Dec 2024 10:22
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 02 11:01:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.310	1.7	71	0.00
49	Acenaphthylene	40.000	39.157	2.1	73	0.00
50	Dimethylphthalate	40.000	38.837	2.9	72	0.00
51	2,6-Dinitrotoluene	40.000	39.791	0.5	72	0.00
52 C	Acenaphthene	40.000	39.516	1.2	73	0.00
53	3-Nitroaniline	40.000	40.697	-1.7	72	0.00
54 P	2,4-Dinitrophenol	40.000	35.042	12.4	61	0.00
55	Dibenzofuran	40.000	39.170	2.1	73	0.00
56 P	4-Nitrophenol	40.000	34.381	14.0	60	0.00
57	2,4-Dinitrotoluene	40.000	40.854	-2.1	73	0.00
58	Fluorene	40.000	39.467	1.3	74	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.989	2.5	70	0.00
60	Diethylphthalate	40.000	38.859	2.9	72	-0.01
61	4-Chlorophenyl-phenylether	40.000	39.860	0.4	74	0.00
62	4-Nitroaniline	40.000	40.706	-1.8	73	0.00
63	Azobenzene	40.000	39.095	2.3	73	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	73	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.032	-5.1	72	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.004	2.5	73	0.00
67	4-Bromophenyl-phenylether	40.000	39.735	0.7	75	0.00
68	Hexachlorobenzene	40.000	38.740	3.1	73	0.00
69	Atrazine	40.000	35.580	11.1	78	0.00
70 C	Pentachlorophenol	40.000	36.131	9.7	60	0.00
71	Phenanthrene	40.000	39.392	1.5	74	0.00
72	Anthracene	40.000	39.774	0.6	74	0.00
73	Carbazole	40.000	40.242	-0.6	75	0.00
74	Di-n-butylphthalate	40.000	38.680	3.3	72	0.00
75 C	Fluoranthene	40.000	40.355	-0.9	76	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	77	0.01
77	Benzydine	40.000	33.951	15.1	66	0.00
78	Pyrene	40.000	40.567	-1.4	78	0.00
79 S	Terphenyl-d14	80.000	78.587	1.8	75	0.00
80	Butylbenzylphthalate	40.000	40.363	-0.9	76	0.00
81	Benzo(a)anthracene	40.000	39.870	0.3	79	0.01
82	3,3'-Dichlorobenzidine	40.000	38.858	2.9	72	0.01
83	Chrysene	40.000	39.658	0.9	77	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.009	-0.0	77	0.00
85 c	Di-n-octyl phthalate	40.000	39.662	0.8	77	0.04
86 I	Perylene-d12	20.000	20.000	0.0	76	0.04
87	Indeno(1,2,3-cd)pyrene	40.000	38.823	2.9	73	0.05
88	Benzo(b)fluoranthene	40.000	39.598	1.0	79	0.04
89	Benzo(k)fluoranthene	40.000	38.123	4.7	72	0.04
90 C	Benzo(a)pyrene	40.000	39.153	2.1	75	0.04
91	Dibenzo(a,h)anthracene	40.000	38.424	3.9	73	0.05
92	Benzo(g,h,i)perylene	40.000	39.067	2.3	74	0.05

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF120224\
Data File : BF140681.D
Acq On : 02 Dec 2024 10:22
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 02 11:01:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 21 15:23:48 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0