

Surrogate Summary

SW-846

SDG No.: P5006

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P5006-18	TW-WTS-02	2-Fluorophenol	75	27.8	37		15 (10)	110 (139)
		Phenol-d6	75	17.9	24		15 (10)	110 (134)
		Nitrobenzene-d5	50	40.5	81		30 (49)	130 (133)
		2-Fluorobiphenyl	50	42.1	84		30 (52)	130 (132)
		2,4,6-Tribromophenol	75	58.9	79		15 (44)	110 (137)
		Terphenyl-d14	50	41.7	83		30 (48)	130 (125)
PB165296BL	PB165296BL	2-Fluorophenol	150	146	97		15 (10)	110 (139)
		Phenol-d6	150	141	94		15 (10)	110 (134)
		Nitrobenzene-d5	100	94.5	94		30 (49)	130 (133)
		2-Fluorobiphenyl	100	96.0	96		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	136	91		15 (44)	110 (137)
		Terphenyl-d14	100	94.7	95		30 (48)	130 (125)
PB165296BS	PB165296BS	2-Fluorophenol	150	129	86		15 (10)	110 (139)
		Phenol-d6	150	129	86		15 (10)	110 (134)
		Nitrobenzene-d5	100	86.7	87		30 (49)	130 (133)
		2-Fluorobiphenyl	100	87.6	88		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	134	89		15 (44)	110 (137)
		Terphenyl-d14	100	87.5	87		30 (48)	130 (125)
PB165296BSD	PB165296BSD	2-Fluorophenol	150	147	98		15 (10)	110 (139)
		Phenol-d6	150	145	96		15 (10)	110 (134)
		Nitrobenzene-d5	100	96.7	97		30 (49)	130 (133)
		2-Fluorobiphenyl	100	95.4	95		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	149	99		15 (44)	110 (137)
		Terphenyl-d14	100	101	101		30 (48)	130 (125)