

Hit Summary Sheet SW-846

SDG No.: P5006
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	SPLP-C4-2							
P5006-01	SPLP-C4-2	WATER	Acetone	31.5		1.40	25.0	ug/L
P5006-01	SPLP-C4-2	WATER	cis-1,2-Dichloroethene	8.00		0.25	5.00	ug/L
P5006-01	SPLP-C4-2	WATER	Benzene	2.00	J	0.16	5.00	ug/L
P5006-01	SPLP-C4-2	WATER	Ethyl Benzene	6.70		0.16	5.00	ug/L
P5006-01	SPLP-C4-2	WATER	o-Xylene	1.90	J	0.14	5.00	ug/L
			Total Voc :	50.1				
			Total Concentration:	50.1				
Client ID:	SPLP-C4-2RE							
P5006-01RE	SPLP-C4-2RE	WATER	Acetone	29.4		1.40	25.0	ug/L
P5006-01RE	SPLP-C4-2RE	WATER	Methyl Acetate	1.00	J	0.60	5.00	ug/L
P5006-01RE	SPLP-C4-2RE	WATER	cis-1,2-Dichloroethene	7.80		0.25	5.00	ug/L
P5006-01RE	SPLP-C4-2RE	WATER	Benzene	2.10	J	0.16	5.00	ug/L
P5006-01RE	SPLP-C4-2RE	WATER	Ethyl Benzene	6.90		0.16	5.00	ug/L
P5006-01RE	SPLP-C4-2RE	WATER	o-Xylene	1.90	J	0.14	5.00	ug/L
			Total Voc :	49.1				
			Total Concentration:	49.1				
Client ID:	SPLP-C4-3							
P5006-02	SPLP-C4-3	WATER	Acetone	19.5	J	1.40	25.0	ug/L
P5006-02	SPLP-C4-3	WATER	cis-1,2-Dichloroethene	1.60	J	0.25	5.00	ug/L
P5006-02	SPLP-C4-3	WATER	Ethyl Benzene	3.10	J	0.16	5.00	ug/L
			Total Voc :	24.2				
			Total Concentration:	24.2				
Client ID:	SPLP-C5-1							
P5006-03	SPLP-C5-1	WATER	Acetone	42.4		1.40	25.0	ug/L
P5006-03	SPLP-C5-1	WATER	2-Butanone	6.40	J	1.30	25.0	ug/L
P5006-03	SPLP-C5-1	WATER	cis-1,2-Dichloroethene	9.70		0.25	5.00	ug/L
P5006-03	SPLP-C5-1	WATER	Benzene	3.10	J	0.16	5.00	ug/L
P5006-03	SPLP-C5-1	WATER	Toluene	1.40	J	0.18	5.00	ug/L
P5006-03	SPLP-C5-1	WATER	Ethyl Benzene	10.3		0.16	5.00	ug/L
P5006-03	SPLP-C5-1	WATER	m/p-Xylenes	3.30	J	0.31	10.0	ug/L
P5006-03	SPLP-C5-1	WATER	o-Xylene	4.20	J	0.14	5.00	ug/L
P5006-03	SPLP-C5-1	WATER	Isopropylbenzene	1.20	J	0.13	5.00	ug/L
			Total Voc :	82.0				
			Total Concentration:	82.0				
Client ID:	SPLP-C5-1RE							
P5006-03RE	SPLP-C5-1RE	WATER	Acetone	51.5		1.40	25.0	ug/L
P5006-03RE	SPLP-C5-1RE	WATER	2-Butanone	7.80	J	1.30	25.0	ug/L

Hit Summary Sheet SW-846

SDG No.: P5006

Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
P5006-03RE	SPLP-C5-1RE	WATER	cis-1,2-Dichloroethene	11.0		0.25	5.00	ug/L
P5006-03RE	SPLP-C5-1RE	WATER	Benzene	3.30	J	0.16	5.00	ug/L
P5006-03RE	SPLP-C5-1RE	WATER	Toluene	1.50	J	0.18	5.00	ug/L
P5006-03RE	SPLP-C5-1RE	WATER	Ethyl Benzene	11.5		0.16	5.00	ug/L
P5006-03RE	SPLP-C5-1RE	WATER	m/p-Xylenes	3.60	J	0.31	10.0	ug/L
P5006-03RE	SPLP-C5-1RE	WATER	o-Xylene	4.40	J	0.14	5.00	ug/L
P5006-03RE	SPLP-C5-1RE	WATER	Isopropylbenzene	1.30	J	0.13	5.00	ug/L
Total Voc :				95.9				
Total Concentration:				95.9				
Client ID:	SPLP-C5-2							
P5006-04	SPLP-C5-2	WATER	Acetone	27.7		1.40	25.0	ug/L
P5006-04	SPLP-C5-2	WATER	Methyl Acetate	1.10	J	0.60	5.00	ug/L
P5006-04	SPLP-C5-2	WATER	cis-1,2-Dichloroethene	4.70	J	0.25	5.00	ug/L
P5006-04	SPLP-C5-2	WATER	Benzene	2.60	J	0.16	5.00	ug/L
P5006-04	SPLP-C5-2	WATER	Toluene	1.10	J	0.18	5.00	ug/L
P5006-04	SPLP-C5-2	WATER	Ethyl Benzene	12.7		0.16	5.00	ug/L
P5006-04	SPLP-C5-2	WATER	m/p-Xylenes	4.50	J	0.31	10.0	ug/L
P5006-04	SPLP-C5-2	WATER	o-Xylene	3.50	J	0.14	5.00	ug/L
P5006-04	SPLP-C5-2	WATER	Isopropylbenzene	1.50	J	0.13	5.00	ug/L
Total Voc :				59.4				
Total Concentration:				59.4				
Client ID:	SPLP-C5-2RE							
P5006-04RE	SPLP-C5-2RE	WATER	Acetone	31.0		1.40	25.0	ug/L
P5006-04RE	SPLP-C5-2RE	WATER	cis-1,2-Dichloroethene	4.90	J	0.25	5.00	ug/L
P5006-04RE	SPLP-C5-2RE	WATER	Benzene	2.60	J	0.16	5.00	ug/L
P5006-04RE	SPLP-C5-2RE	WATER	Toluene	1.20	J	0.18	5.00	ug/L
P5006-04RE	SPLP-C5-2RE	WATER	Ethyl Benzene	13.7		0.16	5.00	ug/L
P5006-04RE	SPLP-C5-2RE	WATER	m/p-Xylenes	4.80	J	0.31	10.0	ug/L
P5006-04RE	SPLP-C5-2RE	WATER	o-Xylene	3.70	J	0.14	5.00	ug/L
P5006-04RE	SPLP-C5-2RE	WATER	Isopropylbenzene	1.70	J	0.13	5.00	ug/L
Total Voc :				63.6				
Total Concentration:				63.6				
Client ID:	SPLP-C5-3							
P5006-05	SPLP-C5-3	WATER	Acetone	16.6	J	1.40	25.0	ug/L
P5006-05	SPLP-C5-3	WATER	cis-1,2-Dichloroethene	1.30	J	0.25	5.00	ug/L
P5006-05	SPLP-C5-3	WATER	Benzene	1.60	J	0.16	5.00	ug/L
P5006-05	SPLP-C5-3	WATER	Ethyl Benzene	43.5		0.16	5.00	ug/L
P5006-05	SPLP-C5-3	WATER	m/p-Xylenes	12.5		0.31	10.0	ug/L

Hit Summary Sheet
SW-846

SDG No.: P5006
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
P5006-05	SPLP-C5-3	WATER	o-Xylene	7.60		0.14	5.00	ug/L
P5006-05	SPLP-C5-3	WATER	Isopropylbenzene	5.20		0.13	5.00	ug/L
			Total Voc :	88.3				
			Total Concentration:	88.3				
Client ID:	SPLP-C5-3RE							
P5006-05RE	SPLP-C5-3RE	WATER	Acetone	17.6	J	1.40	25.0	ug/L
P5006-05RE	SPLP-C5-3RE	WATER	cis-1,2-Dichloroethene	1.10	J	0.25	5.00	ug/L
P5006-05RE	SPLP-C5-3RE	WATER	Benzene	1.60	J	0.16	5.00	ug/L
P5006-05RE	SPLP-C5-3RE	WATER	Ethyl Benzene	43.5		0.16	5.00	ug/L
P5006-05RE	SPLP-C5-3RE	WATER	m/p-Xylenes	12.1		0.31	10.0	ug/L
P5006-05RE	SPLP-C5-3RE	WATER	o-Xylene	7.60		0.14	5.00	ug/L
P5006-05RE	SPLP-C5-3RE	WATER	Isopropylbenzene	5.60		0.13	5.00	ug/L
			Total Voc :	89.1				
			Total Concentration:	89.1				
Client ID:	SPLP-C6-1							
P5006-06	SPLP-C6-1	WATER	Acetone	22.7	J	1.40	25.0	ug/L
P5006-06	SPLP-C6-1	WATER	cis-1,2-Dichloroethene	3.80	J	0.25	5.00	ug/L
P5006-06	SPLP-C6-1	WATER	Ethyl Benzene	2.10	J	0.16	5.00	ug/L
			Total Voc :	28.6				
			Total Concentration:	28.6				
Client ID:	SPLP-C6-1RE							
P5006-06RE	SPLP-C6-1RE	WATER	Acetone	26.7		1.40	25.0	ug/L
P5006-06RE	SPLP-C6-1RE	WATER	cis-1,2-Dichloroethene	3.80	J	0.25	5.00	ug/L
P5006-06RE	SPLP-C6-1RE	WATER	Ethyl Benzene	2.20	J	0.16	5.00	ug/L
			Total Voc :	32.7				
			Total Concentration:	32.7				
Client ID:	SPLP-C6-2							
P5006-07	SPLP-C6-2	WATER	Acetone	17.8	J	1.40	25.0	ug/L
P5006-07	SPLP-C6-2	WATER	Ethyl Benzene	3.50	J	0.16	5.00	ug/L
P5006-07	SPLP-C6-2	WATER	o-Xylene	1.10	J	0.14	5.00	ug/L
			Total Voc :	22.4				
			Total Concentration:	22.4				
Client ID:	SPLP-C6-3							
P5006-08	SPLP-C6-3	WATER	Acetone	17.5	J	1.40	25.0	ug/L
P5006-08	SPLP-C6-3	WATER	cis-1,2-Dichloroethene	1.00	J	0.25	5.00	ug/L
P5006-08	SPLP-C6-3	WATER	Ethyl Benzene	8.50		0.16	5.00	ug/L
P5006-08	SPLP-C6-3	WATER	o-Xylene	1.90	J	0.14	5.00	ug/L
P5006-08	SPLP-C6-3	WATER	Isopropylbenzene	1.20	J	0.13	5.00	ug/L
			Total Voc :	30.1				

Hit Summary Sheet SW-846

SDG No.: P5006
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Total Concentration:				30.1				
Client ID:	SPLP-C6-3RE							
P5006-08RE	SPLP-C6-3RE	WATER	Acetone	20.5	J	1.40	25.0	ug/L
P5006-08RE	SPLP-C6-3RE	WATER	cis-1,2-Dichloroethene	1.00	J	0.25	5.00	ug/L
P5006-08RE	SPLP-C6-3RE	WATER	Ethyl Benzene	9.00		0.16	5.00	ug/L
P5006-08RE	SPLP-C6-3RE	WATER	m/p-Xylenes	2.20	J	0.31	10.0	ug/L
P5006-08RE	SPLP-C6-3RE	WATER	o-Xylene	2.20	J	0.14	5.00	ug/L
P5006-08RE	SPLP-C6-3RE	WATER	Isopropylbenzene	1.20	J	0.13	5.00	ug/L
Total Voc :				36.1				
Total Concentration:				36.1				
Client ID:	SPLP-C10-1							
P5006-09	SPLP-C10-1	WATER	Acetone	21.6	J	1.40	25.0	ug/L
P5006-09	SPLP-C10-1	WATER	cis-1,2-Dichloroethene	6.20		0.25	5.00	ug/L
P5006-09	SPLP-C10-1	WATER	Benzene	23.8		0.16	5.00	ug/L
P5006-09	SPLP-C10-1	WATER	Trichloroethene	1.30	J	0.32	5.00	ug/L
P5006-09	SPLP-C10-1	WATER	Toluene	80.3		0.18	5.00	ug/L
P5006-09	SPLP-C10-1	WATER	Ethyl Benzene	210	E	0.16	5.00	ug/L
P5006-09	SPLP-C10-1	WATER	m/p-Xylenes	80.4		0.31	10.0	ug/L
P5006-09	SPLP-C10-1	WATER	o-Xylene	62.8		0.14	5.00	ug/L
P5006-09	SPLP-C10-1	WATER	Isopropylbenzene	9.70		0.13	5.00	ug/L
Total Voc :				496				
Total Concentration:				496				
Client ID:	SPLP-C10-1DL							
P5006-09DL	SPLP-C10-1DL	WATER	Acetone	27.5	JD	7.00	130	ug/L
P5006-09DL	SPLP-C10-1DL	WATER	cis-1,2-Dichloroethene	6.10	JD	1.30	25.0	ug/L
P5006-09DL	SPLP-C10-1DL	WATER	Benzene	24.0	JD	0.80	25.0	ug/L
P5006-09DL	SPLP-C10-1DL	WATER	Toluene	81.2	D	0.90	25.0	ug/L
P5006-09DL	SPLP-C10-1DL	WATER	Ethyl Benzene	210	D	0.80	25.0	ug/L
P5006-09DL	SPLP-C10-1DL	WATER	m/p-Xylenes	83.1	D	1.60	50.0	ug/L
P5006-09DL	SPLP-C10-1DL	WATER	o-Xylene	62.9	D	0.70	25.0	ug/L
P5006-09DL	SPLP-C10-1DL	WATER	Isopropylbenzene	10.6	JD	0.65	25.0	ug/L
Total Voc :				505				
Total Concentration:				505				
Client ID:	SPLP-C10-2							
P5006-10	SPLP-C10-2	WATER	Vinyl Chloride	1.50	J	0.34	5.00	ug/L
P5006-10	SPLP-C10-2	WATER	Acetone	23.1	J	1.40	25.0	ug/L
P5006-10	SPLP-C10-2	WATER	cis-1,2-Dichloroethene	24.6		0.25	5.00	ug/L
P5006-10	SPLP-C10-2	WATER	Benzene	160	E	0.16	5.00	ug/L

A

B

C

D

Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
P5006-10	SPLP-C10-2	WATER	Toluene	320	E	0.18	5.00	ug/L
P5006-10	SPLP-C10-2	WATER	Ethyl Benzene	590	E	0.16	5.00	ug/L
P5006-10	SPLP-C10-2	WATER	m/p-Xylenes	230		0.31	10.0	ug/L
P5006-10	SPLP-C10-2	WATER	o-Xylene	160	E	0.14	5.00	ug/L
P5006-10	SPLP-C10-2	WATER	Isopropylbenzene	23.4		0.13	5.00	ug/L
			Total Voc :	1530				
			Total Concentration:	1530				
Client ID:	SPLP-C10-2DL							
P5006-10DL	SPLP-C10-2DL	WATER	cis-1,2-Dichloroethene	22.2	JD	5.00	100	ug/L
P5006-10DL	SPLP-C10-2DL	WATER	Benzene	160	D	3.20	100	ug/L
P5006-10DL	SPLP-C10-2DL	WATER	Toluene	310	D	3.60	100	ug/L
P5006-10DL	SPLP-C10-2DL	WATER	Ethyl Benzene	570	D	3.20	100	ug/L
P5006-10DL	SPLP-C10-2DL	WATER	m/p-Xylenes	220	D	6.20	200	ug/L
P5006-10DL	SPLP-C10-2DL	WATER	o-Xylene	160	D	2.80	100	ug/L
P5006-10DL	SPLP-C10-2DL	WATER	Isopropylbenzene	24.3	JD	2.60	100	ug/L
			Total Voc :	1470				
			Total Concentration:	1470				
Client ID:	SPLP-C10-3							
P5006-11	SPLP-C10-3	WATER	Acetone	15.1	J	1.40	25.0	ug/L
P5006-11	SPLP-C10-3	WATER	cis-1,2-Dichloroethene	7.60		0.25	5.00	ug/L
P5006-11	SPLP-C10-3	WATER	Benzene	77.6		0.16	5.00	ug/L
P5006-11	SPLP-C10-3	WATER	Toluene	170	E	0.18	5.00	ug/L
P5006-11	SPLP-C10-3	WATER	Ethyl Benzene	540	E	0.16	5.00	ug/L
P5006-11	SPLP-C10-3	WATER	m/p-Xylenes	180		0.31	10.0	ug/L
P5006-11	SPLP-C10-3	WATER	o-Xylene	130		0.14	5.00	ug/L
P5006-11	SPLP-C10-3	WATER	Isopropylbenzene	24.3		0.13	5.00	ug/L
			Total Voc :	1140				
			Total Concentration:	1140				
Client ID:	SPLP-C10-3DL							
P5006-11DL	SPLP-C10-3DL	WATER	Benzene	72.8	JD	3.20	100	ug/L
P5006-11DL	SPLP-C10-3DL	WATER	Toluene	150	D	3.60	100	ug/L
P5006-11DL	SPLP-C10-3DL	WATER	Ethyl Benzene	480	D	3.20	100	ug/L
P5006-11DL	SPLP-C10-3DL	WATER	m/p-Xylenes	160	JD	6.20	200	ug/L
P5006-11DL	SPLP-C10-3DL	WATER	o-Xylene	120	D	2.80	100	ug/L
P5006-11DL	SPLP-C10-3DL	WATER	Isopropylbenzene	23.2	JD	2.60	100	ug/L
			Total Voc :	1010				
			Total Concentration:	1010				
Client ID:	SPLP-C11-1							

Hit Summary Sheet
SW-846

SDG No.: P5006
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
P5006-12	SPLP-C11-1	WATER	Acetone	16.1	J	1.40	25.0	ug/L
P5006-12	SPLP-C11-1	WATER	cis-1,2-Dichloroethene	1.40	J	0.25	5.00	ug/L
P5006-12	SPLP-C11-1	WATER	Benzene	19.9		0.16	5.00	ug/L
P5006-12	SPLP-C11-1	WATER	Toluene	16.9		0.18	5.00	ug/L
P5006-12	SPLP-C11-1	WATER	Ethyl Benzene	180	E	0.16	5.00	ug/L
P5006-12	SPLP-C11-1	WATER	m/p-Xylenes	37.6		0.31	10.0	ug/L
P5006-12	SPLP-C11-1	WATER	o-Xylene	43.2		0.14	5.00	ug/L
P5006-12	SPLP-C11-1	WATER	Isopropylbenzene	13.5		0.13	5.00	ug/L
Total Voc :				329				
Total Concentration:				329				
Client ID:	SPLP-C11-1DL							
P5006-12DL	SPLP-C11-1DL	WATER	Acetone	22.2	JD	7.00	130	ug/L
P5006-12DL	SPLP-C11-1DL	WATER	Benzene	20.3	JD	0.80	25.0	ug/L
P5006-12DL	SPLP-C11-1DL	WATER	Toluene	17.3	JD	0.90	25.0	ug/L
P5006-12DL	SPLP-C11-1DL	WATER	Ethyl Benzene	180	D	0.80	25.0	ug/L
P5006-12DL	SPLP-C11-1DL	WATER	m/p-Xylenes	37.1	JD	1.60	50.0	ug/L
P5006-12DL	SPLP-C11-1DL	WATER	o-Xylene	44.9	D	0.70	25.0	ug/L
P5006-12DL	SPLP-C11-1DL	WATER	Isopropylbenzene	15.9	JD	0.65	25.0	ug/L
Total Voc :				338				
Total Concentration:				338				
Client ID:	SPLP-C11-2							
P5006-13	SPLP-C11-2	WATER	Acetone	19.2	J	1.40	25.0	ug/L
P5006-13	SPLP-C11-2	WATER	cis-1,2-Dichloroethene	1.10	J	0.25	5.00	ug/L
P5006-13	SPLP-C11-2	WATER	Benzene	15.6		0.16	5.00	ug/L
P5006-13	SPLP-C11-2	WATER	Toluene	12.0		0.18	5.00	ug/L
P5006-13	SPLP-C11-2	WATER	Ethyl Benzene	140		0.16	5.00	ug/L
P5006-13	SPLP-C11-2	WATER	m/p-Xylenes	28.7		0.31	10.0	ug/L
P5006-13	SPLP-C11-2	WATER	o-Xylene	32.5		0.14	5.00	ug/L
P5006-13	SPLP-C11-2	WATER	Isopropylbenzene	11.4		0.13	5.00	ug/L
Total Voc :				261				
Total Concentration:				261				
Client ID:	SPLP-C11-2RE							
P5006-13RE	SPLP-C11-2RE	WATER	Acetone	20.2	J	1.40	25.0	ug/L
P5006-13RE	SPLP-C11-2RE	WATER	cis-1,2-Dichloroethene	1.20	J	0.25	5.00	ug/L
P5006-13RE	SPLP-C11-2RE	WATER	Benzene	16.6		0.16	5.00	ug/L
P5006-13RE	SPLP-C11-2RE	WATER	Toluene	12.3		0.18	5.00	ug/L
P5006-13RE	SPLP-C11-2RE	WATER	Ethyl Benzene	150		0.16	5.00	ug/L
P5006-13RE	SPLP-C11-2RE	WATER	m/p-Xylenes	30.4		0.31	10.0	ug/L

Hit Summary Sheet
SW-846

SDG No.: P5006
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
P5006-13RE	SPLP-C11-2RE	WATER	o-Xylene	33.8		0.14	5.00	ug/L
P5006-13RE	SPLP-C11-2RE	WATER	Isopropylbenzene	12.0		0.13	5.00	ug/L
Total Voc :				277				
Total Concentration:				277				
Client ID:	SPLP-C11-3							
P5006-14	SPLP-C11-3	WATER	Acetone	17.8	J	1.40	25.0	ug/L
P5006-14	SPLP-C11-3	WATER	Benzene	2.50	J	0.16	5.00	ug/L
P5006-14	SPLP-C11-3	WATER	Toluene	2.60	J	0.18	5.00	ug/L
P5006-14	SPLP-C11-3	WATER	Ethyl Benzene	45.9		0.16	5.00	ug/L
P5006-14	SPLP-C11-3	WATER	m/p-Xylenes	9.70	J	0.31	10.0	ug/L
P5006-14	SPLP-C11-3	WATER	o-Xylene	10.8		0.14	5.00	ug/L
P5006-14	SPLP-C11-3	WATER	Isopropylbenzene	4.40	J	0.13	5.00	ug/L
Total Voc :				93.7				
Total Concentration:				93.7				
Client ID:	SPLP-C11-3RE							
P5006-14RE	SPLP-C11-3RE	WATER	Acetone	17.5	J	1.40	25.0	ug/L
P5006-14RE	SPLP-C11-3RE	WATER	Benzene	2.50	J	0.16	5.00	ug/L
P5006-14RE	SPLP-C11-3RE	WATER	Toluene	2.40	J	0.18	5.00	ug/L
P5006-14RE	SPLP-C11-3RE	WATER	Ethyl Benzene	44.8		0.16	5.00	ug/L
P5006-14RE	SPLP-C11-3RE	WATER	m/p-Xylenes	8.90	J	0.31	10.0	ug/L
P5006-14RE	SPLP-C11-3RE	WATER	o-Xylene	10.5		0.14	5.00	ug/L
P5006-14RE	SPLP-C11-3RE	WATER	Isopropylbenzene	4.30	J	0.13	5.00	ug/L
Total Voc :				90.9				
Total Concentration:				90.9				
Client ID:	SPLP-C12-1							
P5006-15	SPLP-C12-1	WATER	Acetone	31.7		1.40	25.0	ug/L
P5006-15	SPLP-C12-1	WATER	cis-1,2-Dichloroethene	2.60	J	0.25	5.00	ug/L
P5006-15	SPLP-C12-1	WATER	Benzene	2.20	J	0.16	5.00	ug/L
P5006-15	SPLP-C12-1	WATER	Toluene	2.50	J	0.18	5.00	ug/L
P5006-15	SPLP-C12-1	WATER	Ethyl Benzene	21.4		0.16	5.00	ug/L
P5006-15	SPLP-C12-1	WATER	m/p-Xylenes	6.50	J	0.31	10.0	ug/L
P5006-15	SPLP-C12-1	WATER	o-Xylene	7.90		0.14	5.00	ug/L
P5006-15	SPLP-C12-1	WATER	Isopropylbenzene	2.00	J	0.13	5.00	ug/L
Total Voc :				76.8				
Total Concentration:				76.8				
Client ID:	SPLP-C12-1RE							
P5006-15RE	SPLP-C12-1RE	WATER	Acetone	33.1		1.40	25.0	ug/L
P5006-15RE	SPLP-C12-1RE	WATER	cis-1,2-Dichloroethene	2.70	J	0.25	5.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: P5006
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
P5006-15RE	SPLP-C12-1RE	WATER	Benzene	2.20	J	0.16	5.00	ug/L
P5006-15RE	SPLP-C12-1RE	WATER	Toluene	2.50	J	0.18	5.00	ug/L
P5006-15RE	SPLP-C12-1RE	WATER	Ethyl Benzene	21.4		0.16	5.00	ug/L
P5006-15RE	SPLP-C12-1RE	WATER	m/p-Xylenes	6.20	J	0.31	10.0	ug/L
P5006-15RE	SPLP-C12-1RE	WATER	o-Xylene	7.40		0.14	5.00	ug/L
P5006-15RE	SPLP-C12-1RE	WATER	Isopropylbenzene	2.20	J	0.13	5.00	ug/L
Total Voc :				77.7				
Total Concentration:				77.7				
Client ID:	SPLP-C12-2							
P5006-16	SPLP-C12-2	WATER	Acetone	19.9	J	1.40	25.0	ug/L
P5006-16	SPLP-C12-2	WATER	cis-1,2-Dichloroethene	1.50	J	0.25	5.00	ug/L
P5006-16	SPLP-C12-2	WATER	Benzene	4.40	J	0.16	5.00	ug/L
P5006-16	SPLP-C12-2	WATER	Toluene	11.1		0.18	5.00	ug/L
P5006-16	SPLP-C12-2	WATER	Ethyl Benzene	88.4		0.16	5.00	ug/L
P5006-16	SPLP-C12-2	WATER	m/p-Xylenes	26.1		0.31	10.0	ug/L
P5006-16	SPLP-C12-2	WATER	o-Xylene	26.0		0.14	5.00	ug/L
P5006-16	SPLP-C12-2	WATER	Isopropylbenzene	7.00		0.13	5.00	ug/L
Total Voc :				184				
Total Concentration:				184				
Client ID:	SPLP-C12-2RE							
P5006-16RE	SPLP-C12-2RE	WATER	Acetone	22.0	J	1.40	25.0	ug/L
P5006-16RE	SPLP-C12-2RE	WATER	cis-1,2-Dichloroethene	1.50	J	0.25	5.00	ug/L
P5006-16RE	SPLP-C12-2RE	WATER	Benzene	4.50	J	0.16	5.00	ug/L
P5006-16RE	SPLP-C12-2RE	WATER	Toluene	11.9		0.18	5.00	ug/L
P5006-16RE	SPLP-C12-2RE	WATER	Ethyl Benzene	92.5		0.16	5.00	ug/L
P5006-16RE	SPLP-C12-2RE	WATER	m/p-Xylenes	26.9		0.31	10.0	ug/L
P5006-16RE	SPLP-C12-2RE	WATER	o-Xylene	27.1		0.14	5.00	ug/L
P5006-16RE	SPLP-C12-2RE	WATER	Isopropylbenzene	7.50		0.13	5.00	ug/L
Total Voc :				194				
Total Concentration:				194				
Client ID:	SPLP-C12-3							
P5006-17	SPLP-C12-3	WATER	Acetone	11.4	J	1.40	25.0	ug/L
P5006-17	SPLP-C12-3	WATER	Benzene	3.20	J	0.16	5.00	ug/L
P5006-17	SPLP-C12-3	WATER	Toluene	8.10		0.18	5.00	ug/L
P5006-17	SPLP-C12-3	WATER	Ethyl Benzene	65.0		0.16	5.00	ug/L
P5006-17	SPLP-C12-3	WATER	m/p-Xylenes	19.1		0.31	10.0	ug/L
P5006-17	SPLP-C12-3	WATER	o-Xylene	18.3		0.14	5.00	ug/L
P5006-17	SPLP-C12-3	WATER	Isopropylbenzene	5.10		0.13	5.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: P5006

Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Total Voc :				130				
Total Concentration:				130				
Client ID:	SPLP-C12-3RE							
P5006-17RE	SPLP-C12-3RE	WATER	Acetone	14.1	J	1.40	25.0	ug/L
P5006-17RE	SPLP-C12-3RE	WATER	Benzene	3.50	J	0.16	5.00	ug/L
P5006-17RE	SPLP-C12-3RE	WATER	Toluene	9.10		0.18	5.00	ug/L
P5006-17RE	SPLP-C12-3RE	WATER	Ethyl Benzene	75.9		0.16	5.00	ug/L
P5006-17RE	SPLP-C12-3RE	WATER	m/p-Xylenes	21.2		0.31	10.0	ug/L
P5006-17RE	SPLP-C12-3RE	WATER	o-Xylene	20.8		0.14	5.00	ug/L
P5006-17RE	SPLP-C12-3RE	WATER	Isopropylbenzene	6.30		0.13	5.00	ug/L
Total Voc :				151				
Total Concentration:				151				



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C4-2		SDG No.:	P5006	
Lab Sample ID:	P5006-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044076.D	1		12/02/24 19:04	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	31.5		1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	8.00		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	2.00	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C4-2	SDG No.:	P5006
Lab Sample ID:	P5006-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044076.D	1		12/02/24 19:04	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	6.70		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	10.0	ug/L
95-47-6	o-Xylene	1.90	J	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.8		70 (74) - 130 (125)	108%	SPK: 50
1868-53-7	Dibromofluoromethane	35.6		70 (75) - 130 (124)	71%	SPK: 50
2037-26-5	Toluene-d8	51.4		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.3		70 (77) - 130 (121)	109%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	114000	5.544			
540-36-3	1,4-Difluorobenzene	228000	6.757			
3114-55-4	Chlorobenzene-d5	210000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	88400	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C4-2		SDG No.:	P5006	
Lab Sample ID:	P5006-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044076.D	1		12/02/24 19:04	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C4-2RE		SDG No.:	P5006	
Lab Sample ID:	P5006-01RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044088.D	1		12/03/24 13:24	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	29.4		1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	1.00	J	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	7.80		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	2.10	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C4-2RE		SDG No.:	P5006	
Lab Sample ID:	P5006-01RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044088.D	1		12/03/24 13:24	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	6.90		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	10.0	ug/L
95-47-6	o-Xylene	1.90	J	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.1		70 (74) - 130 (125)	102%	SPK: 50
1868-53-7	Dibromofluoromethane	37.2		70 (75) - 130 (124)	74%	SPK: 50
2037-26-5	Toluene-d8	48.1		70 (86) - 130 (113)	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.4		70 (77) - 130 (121)	97%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	117000	5.544			
540-36-3	1,4-Difluorobenzene	224000	6.757			
3114-55-4	Chlorobenzene-d5	190000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	82000	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C4-2RE		SDG No.:	P5006	
Lab Sample ID:	P5006-01RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044088.D	1		12/03/24 13:24	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C4-3		SDG No.:	P5006	
Lab Sample ID:	P5006-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044089.D	1		12/03/24 13:47	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	19.5	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.60	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	0.16	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C4-3		SDG No.:	P5006	
Lab Sample ID:	P5006-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044089.D	1		12/03/24 13:47	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	3.10	J	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	10.0	ug/L
95-47-6	o-Xylene	0.14	U	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.3		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	38.5		70 (75) - 130 (124)	77%	SPK: 50
2037-26-5	Toluene-d8	50.1		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.0		70 (77) - 130 (121)	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	113000	5.544			
540-36-3	1,4-Difluorobenzene	220000	6.751			
3114-55-4	Chlorobenzene-d5	195000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	89500	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C4-3		SDG No.:	P5006	
Lab Sample ID:	P5006-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044089.D	1		12/03/24 13:47	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C5-1	SDG No.:	P5006
Lab Sample ID:	P5006-03	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044091.D	1		12/03/24 14:33	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	42.4		1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	6.40	J	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	9.70		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	3.10	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	1.40	J	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C5-1	SDG No.:	P5006
Lab Sample ID:	P5006-03	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044091.D	1		12/03/24 14:33	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	10.3		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	3.30	J	0.31	10.0	ug/L
95-47-6	o-Xylene	4.20	J	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	1.20	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.8		70 (74) - 130 (125)	104%	SPK: 50
1868-53-7	Dibromofluoromethane	35.8		70 (75) - 130 (124)	72%	SPK: 50
2037-26-5	Toluene-d8	49.8		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.5		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	115000	5.544			
540-36-3	1,4-Difluorobenzene	227000	6.757			
3114-55-4	Chlorobenzene-d5	195000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	88100	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C5-1		SDG No.:	P5006	
Lab Sample ID:	P5006-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044091.D	1		12/03/24 14:33	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C5-1RE		SDG No.:	P5006	
Lab Sample ID:	P5006-03RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044120.D	1		12/04/24 16:29	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	51.5		1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	7.80	J	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	11.0		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	3.30	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	1.50	J	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C5-1RE	SDG No.:	P5006
Lab Sample ID:	P5006-03RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044120.D	1		12/04/24 16:29	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	11.5		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	3.60	J	0.31	10.0	ug/L
95-47-6	o-Xylene	4.40	J	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	1.30	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.8		70 (74) - 130 (125)	104%	SPK: 50
1868-53-7	Dibromofluoromethane	32.9	*	70 (75) - 130 (124)	66%	SPK: 50
2037-26-5	Toluene-d8	49.7		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.9		70 (77) - 130 (121)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	108000	5.544			
540-36-3	1,4-Difluorobenzene	207000	6.757			
3114-55-4	Chlorobenzene-d5	184000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	80100	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C5-2		SDG No.:	P5006	
Lab Sample ID:	P5006-04		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044114.D	1		12/04/24 14:10	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	27.7		1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	1.10	J	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	4.70	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	2.60	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	1.10	J	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C5-2	SDG No.:	P5006
Lab Sample ID:	P5006-04	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044114.D	1		12/04/24 14:10	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	12.7		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	4.50	J	0.31	10.0	ug/L
95-47-6	o-Xylene	3.50	J	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	1.50	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.2		70 (74) - 130 (125)	102%	SPK: 50
1868-53-7	Dibromofluoromethane	34.3	*	70 (75) - 130 (124)	69%	SPK: 50
2037-26-5	Toluene-d8	49.2		70 (86) - 130 (113)	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.5		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	108000	5.544			
540-36-3	1,4-Difluorobenzene	209000	6.757			
3114-55-4	Chlorobenzene-d5	184000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	81300	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C5-2		SDG No.:	P5006	
Lab Sample ID:	P5006-04		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044114.D	1		12/04/24 14:10	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C5-2RE	SDG No.:	P5006
Lab Sample ID:	P5006-04RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044129.D	1		12/04/24 19:57	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	31.0		1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	4.90	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	2.60	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	1.20	J	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C5-2RE		SDG No.:	P5006	
Lab Sample ID:	P5006-04RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044129.D	1		12/04/24 19:57	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	13.7		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	4.80	J	0.31	10.0	ug/L
95-47-6	o-Xylene	3.70	J	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	1.70	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.3		70 (74) - 130 (125)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	35.0		70 (75) - 130 (124)	70%	SPK: 50
2037-26-5	Toluene-d8	49.8		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.1		70 (77) - 130 (121)	96%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	102000	5.55			
540-36-3	1,4-Difluorobenzene	200000	6.757			
3114-55-4	Chlorobenzene-d5	169000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	68400	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C5-2RE		SDG No.:	P5006	
Lab Sample ID:	P5006-04RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044129.D	1		12/04/24 19:57	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C5-3		SDG No.:	P5006	
Lab Sample ID:	P5006-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044093.D	1		12/03/24 15:19	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	16.6	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.30	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	1.60	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C5-3	SDG No.:	P5006
Lab Sample ID:	P5006-05	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044093.D	1		12/03/24 15:19	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	43.5		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	12.5		0.31	10.0	ug/L
95-47-6	o-Xylene	7.60		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	5.20		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.4		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	37.2		70 (75) - 130 (124)	74%	SPK: 50
2037-26-5	Toluene-d8	50.0		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.4		70 (77) - 130 (121)	107%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	106000	5.544			
540-36-3	1,4-Difluorobenzene	209000	6.757			
3114-55-4	Chlorobenzene-d5	186000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	88200	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C5-3		SDG No.:	P5006	
Lab Sample ID:	P5006-05		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044093.D	1		12/03/24 15:19	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C5-3RE		SDG No.:	P5006	
Lab Sample ID:	P5006-05RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044121.D	1		12/04/24 16:52	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	17.6	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.10	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	1.60	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C5-3RE	SDG No.:	P5006
Lab Sample ID:	P5006-05RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044121.D	1		12/04/24 16:52	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	43.5		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	12.1		0.31	10.0	ug/L
95-47-6	o-Xylene	7.60		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	5.60		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.3		70 (74) - 130 (125)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	35.1		70 (75) - 130 (124)	70%	SPK: 50
2037-26-5	Toluene-d8	49.8		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.9		70 (77) - 130 (121)	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	106000	5.55			
540-36-3	1,4-Difluorobenzene	207000	6.757			
3114-55-4	Chlorobenzene-d5	181000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	79100	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C6-1		SDG No.:	P5006	
Lab Sample ID:	P5006-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044094.D	1		12/03/24 15:41	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	22.7	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	3.80	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	0.16	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C6-1	SDG No.:	P5006
Lab Sample ID:	P5006-06	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044094.D	1		12/03/24 15:41	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	2.10	J	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	10.0	ug/L
95-47-6	o-Xylene	0.14	U	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.4		70 (74) - 130 (125)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	36.5		70 (75) - 130 (124)	73%	SPK: 50
2037-26-5	Toluene-d8	50.7		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.1		70 (77) - 130 (121)	100%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	114000	5.544			
540-36-3	1,4-Difluorobenzene	221000	6.757			
3114-55-4	Chlorobenzene-d5	192000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	79600	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C6-1		SDG No.:	P5006	
Lab Sample ID:	P5006-06		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044094.D	1		12/03/24 15:41	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C6-1RE		SDG No.:	P5006	
Lab Sample ID:	P5006-06RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044122.D	1		12/04/24 17:15	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	26.7		1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	3.80	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	0.16	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C6-1RE	SDG No.:	P5006
Lab Sample ID:	P5006-06RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044122.D	1		12/04/24 17:15	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	2.20	J	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	10.0	ug/L
95-47-6	o-Xylene	0.14	U	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.7		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	34.4	*	70 (75) - 130 (124)	69%	SPK: 50
2037-26-5	Toluene-d8	48.8		70 (86) - 130 (113)	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.2		70 (77) - 130 (121)	98%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	105000	5.55			
540-36-3	1,4-Difluorobenzene	207000	6.757			
3114-55-4	Chlorobenzene-d5	177000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	72600	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C6-1RE		SDG No.:	P5006	
Lab Sample ID:	P5006-06RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044122.D	1		12/04/24 17:15	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C6-2		SDG No.:	P5006	
Lab Sample ID:	P5006-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044095.D	1		12/03/24 16:04	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	17.8	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	0.16	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C6-2		SDG No.:	P5006	
Lab Sample ID:	P5006-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044095.D	1		12/03/24 16:04	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	3.50	J	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	10.0	ug/L
95-47-6	o-Xylene	1.10	J	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.4		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	38.0		70 (75) - 130 (124)	76%	SPK: 50
2037-26-5	Toluene-d8	50.6		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.2		70 (77) - 130 (121)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	112000	5.55			
540-36-3	1,4-Difluorobenzene	223000	6.757			
3114-55-4	Chlorobenzene-d5	201000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	87400	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C6-2		SDG No.:	P5006	
Lab Sample ID:	P5006-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044095.D	1		12/03/24 16:04	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C6-3		SDG No.:	P5006	
Lab Sample ID:	P5006-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044096.D	1		12/03/24 16:27	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	17.5	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.00	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	0.16	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C6-3	SDG No.:	P5006
Lab Sample ID:	P5006-08	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044096.D	1		12/03/24 16:27	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	8.50		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	10.0	ug/L
95-47-6	o-Xylene	1.90	J	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	1.20	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.4		70 (74) - 130 (125)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	37.1		70 (75) - 130 (124)	74%	SPK: 50
2037-26-5	Toluene-d8	49.4		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.6		70 (77) - 130 (121)	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	118000	5.544			
540-36-3	1,4-Difluorobenzene	225000	6.757			
3114-55-4	Chlorobenzene-d5	199000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	90200	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C6-3		SDG No.:	P5006	
Lab Sample ID:	P5006-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044096.D	1		12/03/24 16:27	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C6-3RE		SDG No.:	P5006	
Lab Sample ID:	P5006-08RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044123.D	1		12/04/24 17:38	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	20.5	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.00	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	0.16	U	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	0.18	U	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C6-3RE	SDG No.:	P5006
Lab Sample ID:	P5006-08RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044123.D	1		12/04/24 17:38	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	9.00		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	2.20	J	0.31	10.0	ug/L
95-47-6	o-Xylene	2.20	J	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	1.20	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.6		70 (74) - 130 (125)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	34.7	*	70 (75) - 130 (124)	69%	SPK: 50
2037-26-5	Toluene-d8	49.7		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.9		70 (77) - 130 (121)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	108000	5.55			
540-36-3	1,4-Difluorobenzene	212000	6.757			
3114-55-4	Chlorobenzene-d5	187000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	86000	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C6-3RE		SDG No.:	P5006	
Lab Sample ID:	P5006-08RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044123.D	1		12/04/24 17:38	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C10-1		SDG No.:	P5006	
Lab Sample ID:	P5006-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044097.D	1		12/03/24 16:51	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	21.6	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	6.20		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	23.8		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	1.30	J	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	80.3		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C10-1	SDG No.:	P5006
Lab Sample ID:	P5006-09	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044097.D	1		12/03/24 16:51	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	210	E	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	80.4		0.31	10.0	ug/L
95-47-6	o-Xylene	62.8		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	9.70		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.3		70 (74) - 130 (125)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	36.8		70 (75) - 130 (124)	74%	SPK: 50
2037-26-5	Toluene-d8	50.1		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.1		70 (77) - 130 (121)	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	114000	5.544			
540-36-3	1,4-Difluorobenzene	225000	6.757			
3114-55-4	Chlorobenzene-d5	195000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	94500	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C10-1		SDG No.:	P5006	
Lab Sample ID:	P5006-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044097.D	1		12/03/24 16:51	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C10-1DL	SDG No.:	P5006
Lab Sample ID:	P5006-09DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044116.D	5		12/04/24 14:56	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	UD	1.10	25.0	ug/L
74-87-3	Chloromethane	1.80	UD	1.80	25.0	ug/L
75-01-4	Vinyl Chloride	1.70	UD	1.70	25.0	ug/L
74-83-9	Bromomethane	6.80	UD	6.80	25.0	ug/L
75-00-3	Chloroethane	2.80	UD	2.80	25.0	ug/L
75-69-4	Trichlorofluoromethane	1.70	UD	1.70	25.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	1.30	UD	1.30	25.0	ug/L
75-35-4	1,1-Dichloroethene	1.30	UD	1.30	25.0	ug/L
67-64-1	Acetone	27.5	JD	7.00	130	ug/L
75-15-0	Carbon Disulfide	1.60	UD	1.60	25.0	ug/L
1634-04-4	Methyl tert-butyl Ether	0.80	UD	0.80	25.0	ug/L
79-20-9	Methyl Acetate	3.00	UD	3.00	25.0	ug/L
75-09-2	Methylene Chloride	1.60	UD	1.60	25.0	ug/L
156-60-5	trans-1,2-Dichloroethene	1.30	UD	1.30	25.0	ug/L
75-34-3	1,1-Dichloroethane	1.20	UD	1.20	25.0	ug/L
110-82-7	Cyclohexane	8.10	UD	8.10	25.0	ug/L
78-93-3	2-Butanone	6.50	UD	6.50	130	ug/L
56-23-5	Carbon Tetrachloride	1.30	UD	1.30	25.0	ug/L
156-59-2	cis-1,2-Dichloroethene	6.10	JD	1.30	25.0	ug/L
74-97-5	Bromochloromethane	0.90	UD	0.90	25.0	ug/L
67-66-3	Chloroform	1.30	UD	1.30	25.0	ug/L
71-55-6	1,1,1-Trichloroethane	0.95	UD	0.95	25.0	ug/L
108-87-2	Methylcyclohexane	0.95	UD	0.95	25.0	ug/L
71-43-2	Benzene	24.0	JD	0.80	25.0	ug/L
107-06-2	1,2-Dichloroethane	1.20	UD	1.20	25.0	ug/L
79-01-6	Trichloroethene	1.60	UD	1.60	25.0	ug/L
78-87-5	1,2-Dichloropropane	0.95	UD	0.95	25.0	ug/L
75-27-4	Bromodichloromethane	1.20	UD	1.20	25.0	ug/L
108-10-1	4-Methyl-2-Pentanone	3.80	UD	3.80	130	ug/L
108-88-3	Toluene	81.2	D	0.90	25.0	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C10-1DL	SDG No.:	P5006
Lab Sample ID:	P5006-09DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044116.D	5		12/04/24 14:56	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	1.10	UD	1.10	25.0	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.90	UD	0.90	25.0	ug/L
79-00-5	1,1,2-Trichloroethane	1.10	UD	1.10	25.0	ug/L
591-78-6	2-Hexanone	5.70	UD	5.70	130	ug/L
124-48-1	Dibromochloromethane	0.90	UD	0.90	25.0	ug/L
106-93-4	1,2-Dibromoethane	0.80	UD	0.80	25.0	ug/L
127-18-4	Tetrachloroethene	1.30	UD	1.30	25.0	ug/L
108-90-7	Chlorobenzene	0.65	UD	0.65	25.0	ug/L
100-41-4	Ethyl Benzene	210	D	0.80	25.0	ug/L
179601-23-1	m/p-Xylenes	83.1	D	1.60	50.0	ug/L
95-47-6	o-Xylene	62.9	D	0.70	25.0	ug/L
100-42-5	Styrene	0.80	UD	0.80	25.0	ug/L
75-25-2	Bromoform	1.10	UD	1.10	25.0	ug/L
98-82-8	Isopropylbenzene	10.6	JD	0.65	25.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	1.40	UD	1.40	25.0	ug/L
541-73-1	1,3-Dichlorobenzene	1.20	UD	1.20	25.0	ug/L
106-46-7	1,4-Dichlorobenzene	1.40	UD	1.40	25.0	ug/L
95-50-1	1,2-Dichlorobenzene	0.95	UD	0.95	25.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	2.30	UD	2.30	25.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	2.10	UD	2.10	25.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	2.60	UD	2.60	25.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.5		70 (74) - 130 (125)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	45.4		70 (75) - 130 (124)	91%	SPK: 50
2037-26-5	Toluene-d8	50.0		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.9		70 (77) - 130 (121)	100%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	107000	5.55			
540-36-3	1,4-Difluorobenzene	210000	6.757			
3114-55-4	Chlorobenzene-d5	182000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	79000	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C10-1DL		SDG No.:	P5006	
Lab Sample ID:	P5006-09DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044116.D	5		12/04/24 14:56	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C10-2	SDG No.:	P5006
Lab Sample ID:	P5006-10	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044098.D	1		12/03/24 17:14	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	1.50	J	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	23.1	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	24.6		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	160	E	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	320	E	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C10-2		SDG No.:	P5006	
Lab Sample ID:	P5006-10		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044098.D	1		12/03/24 17:14	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	590	E	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	230		0.31	10.0	ug/L
95-47-6	o-Xylene	160	E	0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	23.4		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.3		70 (74) - 130 (125)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	35.0		70 (75) - 130 (124)	70%	SPK: 50
2037-26-5	Toluene-d8	50.5		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.5		70 (77) - 130 (121)	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	103000	5.55			
540-36-3	1,4-Difluorobenzene	204000	6.757			
3114-55-4	Chlorobenzene-d5	182000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	84700	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C10-2		SDG No.:	P5006	
Lab Sample ID:	P5006-10		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044098.D	1		12/03/24 17:14	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C10-2DL	SDG No.:	P5006
Lab Sample ID:	P5006-10DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044118.D	20		12/04/24 15:43	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	4.20	UD	4.20	100	ug/L
74-87-3	Chloromethane	7.00	UD	7.00	100	ug/L
75-01-4	Vinyl Chloride	6.80	UD	6.80	100	ug/L
74-83-9	Bromomethane	27.2	UD	27.2	100	ug/L
75-00-3	Chloroethane	11.2	UD	11.2	100	ug/L
75-69-4	Trichlorofluoromethane	6.80	UD	6.80	100	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	5.00	UD	5.00	100	ug/L
75-35-4	1,1-Dichloroethene	5.20	UD	5.20	100	ug/L
67-64-1	Acetone	27.8	UD	27.8	500	ug/L
75-15-0	Carbon Disulfide	6.40	UD	6.40	100	ug/L
1634-04-4	Methyl tert-butyl Ether	3.20	UD	3.20	100	ug/L
79-20-9	Methyl Acetate	12.0	UD	12.0	100	ug/L
75-09-2	Methylene Chloride	6.40	UD	6.40	100	ug/L
156-60-5	trans-1,2-Dichloroethene	5.00	UD	5.00	100	ug/L
75-34-3	1,1-Dichloroethane	4.60	UD	4.60	100	ug/L
110-82-7	Cyclohexane	32.4	UD	32.4	100	ug/L
78-93-3	2-Butanone	26.0	UD	26.0	500	ug/L
56-23-5	Carbon Tetrachloride	5.00	UD	5.00	100	ug/L
156-59-2	cis-1,2-Dichloroethene	22.2	JD	5.00	100	ug/L
74-97-5	Bromochloromethane	3.60	UD	3.60	100	ug/L
67-66-3	Chloroform	5.20	UD	5.20	100	ug/L
71-55-6	1,1,1-Trichloroethane	3.80	UD	3.80	100	ug/L
108-87-2	Methylcyclohexane	3.80	UD	3.80	100	ug/L
71-43-2	Benzene	160	D	3.20	100	ug/L
107-06-2	1,2-Dichloroethane	4.80	UD	4.80	100	ug/L
79-01-6	Trichloroethene	6.40	UD	6.40	100	ug/L
78-87-5	1,2-Dichloropropane	3.80	UD	3.80	100	ug/L
75-27-4	Bromodichloromethane	4.80	UD	4.80	100	ug/L
108-10-1	4-Methyl-2-Pentanone	15.0	UD	15.0	500	ug/L
108-88-3	Toluene	310	D	3.60	100	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C10-2DL		SDG No.:	P5006	
Lab Sample ID:	P5006-10DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044118.D	20		12/04/24 15:43	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	4.20	UD	4.20	100	ug/L
10061-01-5	cis-1,3-Dichloropropene	3.60	UD	3.60	100	ug/L
79-00-5	1,1,2-Trichloroethane	4.20	UD	4.20	100	ug/L
591-78-6	2-Hexanone	22.6	UD	22.6	500	ug/L
124-48-1	Dibromochloromethane	3.60	UD	3.60	100	ug/L
106-93-4	1,2-Dibromoethane	3.20	UD	3.20	100	ug/L
127-18-4	Tetrachloroethene	5.00	UD	5.00	100	ug/L
108-90-7	Chlorobenzene	2.60	UD	2.60	100	ug/L
100-41-4	Ethyl Benzene	570	D	3.20	100	ug/L
179601-23-1	m/p-Xylenes	220	D	6.20	200	ug/L
95-47-6	o-Xylene	160	D	2.80	100	ug/L
100-42-5	Styrene	3.20	UD	3.20	100	ug/L
75-25-2	Bromoform	4.20	UD	4.20	100	ug/L
98-82-8	Isopropylbenzene	24.3	JD	2.60	100	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	5.40	UD	5.40	100	ug/L
541-73-1	1,3-Dichlorobenzene	4.80	UD	4.80	100	ug/L
106-46-7	1,4-Dichlorobenzene	5.40	UD	5.40	100	ug/L
95-50-1	1,2-Dichlorobenzene	3.80	UD	3.80	100	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	9.20	UD	9.20	100	ug/L
120-82-1	1,2,4-Trichlorobenzene	8.40	UD	8.40	100	ug/L
87-61-6	1,2,3-Trichlorobenzene	10.2	UD	10.2	100	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.0		70 (74) - 130 (125)	104%	SPK: 50
1868-53-7	Dibromofluoromethane	46.2		70 (75) - 130 (124)	92%	SPK: 50
2037-26-5	Toluene-d8	50.1		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.1		70 (77) - 130 (121)	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	106000	5.55			
540-36-3	1,4-Difluorobenzene	202000	6.757			
3114-55-4	Chlorobenzene-d5	175000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	74900	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C10-2DL		SDG No.:	P5006	
Lab Sample ID:	P5006-10DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044118.D	20		12/04/24 15:43	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C10-3		SDG No.:	P5006	
Lab Sample ID:	P5006-11		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044099.D	1		12/03/24 17:36	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	15.1	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	7.60		0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	77.6		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	170	E	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C10-3	SDG No.:	P5006
Lab Sample ID:	P5006-11	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044099.D	1		12/03/24 17:36	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	540	E	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	180		0.31	10.0	ug/L
95-47-6	o-Xylene	130		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	24.3		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.7		70 (74) - 130 (125)	103%	SPK: 50
1868-53-7	Dibromofluoromethane	37.6		70 (75) - 130 (124)	75%	SPK: 50
2037-26-5	Toluene-d8	49.7		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.4		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	105000	5.544			
540-36-3	1,4-Difluorobenzene	204000	6.757			
3114-55-4	Chlorobenzene-d5	175000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	82500	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C10-3		SDG No.:	P5006	
Lab Sample ID:	P5006-11		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044099.D	1		12/03/24 17:36	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C10-3DL	SDG No.:	P5006
Lab Sample ID:	P5006-11DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044119.D	20		12/04/24 16:06	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	4.20	UD	4.20	100	ug/L
74-87-3	Chloromethane	7.00	UD	7.00	100	ug/L
75-01-4	Vinyl Chloride	6.80	UD	6.80	100	ug/L
74-83-9	Bromomethane	27.2	UD	27.2	100	ug/L
75-00-3	Chloroethane	11.2	UD	11.2	100	ug/L
75-69-4	Trichlorofluoromethane	6.80	UD	6.80	100	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	5.00	UD	5.00	100	ug/L
75-35-4	1,1-Dichloroethene	5.20	UD	5.20	100	ug/L
67-64-1	Acetone	27.8	UD	27.8	500	ug/L
75-15-0	Carbon Disulfide	6.40	UD	6.40	100	ug/L
1634-04-4	Methyl tert-butyl Ether	3.20	UD	3.20	100	ug/L
79-20-9	Methyl Acetate	12.0	UD	12.0	100	ug/L
75-09-2	Methylene Chloride	6.40	UD	6.40	100	ug/L
156-60-5	trans-1,2-Dichloroethene	5.00	UD	5.00	100	ug/L
75-34-3	1,1-Dichloroethane	4.60	UD	4.60	100	ug/L
110-82-7	Cyclohexane	32.4	UD	32.4	100	ug/L
78-93-3	2-Butanone	26.0	UD	26.0	500	ug/L
56-23-5	Carbon Tetrachloride	5.00	UD	5.00	100	ug/L
156-59-2	cis-1,2-Dichloroethene	5.00	UD	5.00	100	ug/L
74-97-5	Bromochloromethane	3.60	UD	3.60	100	ug/L
67-66-3	Chloroform	5.20	UD	5.20	100	ug/L
71-55-6	1,1,1-Trichloroethane	3.80	UD	3.80	100	ug/L
108-87-2	Methylcyclohexane	3.80	UD	3.80	100	ug/L
71-43-2	Benzene	72.8	JD	3.20	100	ug/L
107-06-2	1,2-Dichloroethane	4.80	UD	4.80	100	ug/L
79-01-6	Trichloroethene	6.40	UD	6.40	100	ug/L
78-87-5	1,2-Dichloropropane	3.80	UD	3.80	100	ug/L
75-27-4	Bromodichloromethane	4.80	UD	4.80	100	ug/L
108-10-1	4-Methyl-2-Pentanone	15.0	UD	15.0	500	ug/L
108-88-3	Toluene	150	D	3.60	100	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C10-3DL	SDG No.:	P5006
Lab Sample ID:	P5006-11DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044119.D	20		12/04/24 16:06	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	4.20	UD	4.20	100	ug/L
10061-01-5	cis-1,3-Dichloropropene	3.60	UD	3.60	100	ug/L
79-00-5	1,1,2-Trichloroethane	4.20	UD	4.20	100	ug/L
591-78-6	2-Hexanone	22.6	UD	22.6	500	ug/L
124-48-1	Dibromochloromethane	3.60	UD	3.60	100	ug/L
106-93-4	1,2-Dibromoethane	3.20	UD	3.20	100	ug/L
127-18-4	Tetrachloroethene	5.00	UD	5.00	100	ug/L
108-90-7	Chlorobenzene	2.60	UD	2.60	100	ug/L
100-41-4	Ethyl Benzene	480	D	3.20	100	ug/L
179601-23-1	m/p-Xylenes	160	JD	6.20	200	ug/L
95-47-6	o-Xylene	120	D	2.80	100	ug/L
100-42-5	Styrene	3.20	UD	3.20	100	ug/L
75-25-2	Bromoform	4.20	UD	4.20	100	ug/L
98-82-8	Isopropylbenzene	23.2	JD	2.60	100	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	5.40	UD	5.40	100	ug/L
541-73-1	1,3-Dichlorobenzene	4.80	UD	4.80	100	ug/L
106-46-7	1,4-Dichlorobenzene	5.40	UD	5.40	100	ug/L
95-50-1	1,2-Dichlorobenzene	3.80	UD	3.80	100	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	9.20	UD	9.20	100	ug/L
120-82-1	1,2,4-Trichlorobenzene	8.40	UD	8.40	100	ug/L
87-61-6	1,2,3-Trichlorobenzene	10.2	UD	10.2	100	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.0		70 (74) - 130 (125)	106%	SPK: 50
1868-53-7	Dibromofluoromethane	45.1		70 (75) - 130 (124)	90%	SPK: 50
2037-26-5	Toluene-d8	49.5		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.4		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	105000	5.55			
540-36-3	1,4-Difluorobenzene	207000	6.757			
3114-55-4	Chlorobenzene-d5	180000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	78900	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C10-3DL		SDG No.:	P5006	
Lab Sample ID:	P5006-11DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044119.D	20		12/04/24 16:06	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-1		SDG No.:	P5006	
Lab Sample ID:	P5006-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044100.D	1		12/03/24 18:00	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	16.1	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.40	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	19.9		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	16.9		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-1		SDG No.:	P5006	
Lab Sample ID:	P5006-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044100.D	1		12/03/24 18:00	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	180	E	0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	37.6		0.31	10.0	ug/L
95-47-6	o-Xylene	43.2		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	13.5		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.1		70 (74) - 130 (125)	104%	SPK: 50
1868-53-7	Dibromofluoromethane	34.1	*	70 (75) - 130 (124)	68%	SPK: 50
2037-26-5	Toluene-d8	50.6		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.7		70 (77) - 130 (121)	111%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	102000	5.55			
540-36-3	1,4-Difluorobenzene	200000	6.757			
3114-55-4	Chlorobenzene-d5	178000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	86200	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-1		SDG No.:	P5006	
Lab Sample ID:	P5006-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044100.D	1		12/03/24 18:00	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-1DL		SDG No.:	P5006	
Lab Sample ID:	P5006-12DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044117.D	5		12/04/24 15:20	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	UD	1.10	25.0	ug/L
74-87-3	Chloromethane	1.80	UD	1.80	25.0	ug/L
75-01-4	Vinyl Chloride	1.70	UD	1.70	25.0	ug/L
74-83-9	Bromomethane	6.80	UD	6.80	25.0	ug/L
75-00-3	Chloroethane	2.80	UD	2.80	25.0	ug/L
75-69-4	Trichlorofluoromethane	1.70	UD	1.70	25.0	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	1.30	UD	1.30	25.0	ug/L
75-35-4	1,1-Dichloroethene	1.30	UD	1.30	25.0	ug/L
67-64-1	Acetone	22.2	JD	7.00	130	ug/L
75-15-0	Carbon Disulfide	1.60	UD	1.60	25.0	ug/L
1634-04-4	Methyl tert-butyl Ether	0.80	UD	0.80	25.0	ug/L
79-20-9	Methyl Acetate	3.00	UD	3.00	25.0	ug/L
75-09-2	Methylene Chloride	1.60	UD	1.60	25.0	ug/L
156-60-5	trans-1,2-Dichloroethene	1.30	UD	1.30	25.0	ug/L
75-34-3	1,1-Dichloroethane	1.20	UD	1.20	25.0	ug/L
110-82-7	Cyclohexane	8.10	UD	8.10	25.0	ug/L
78-93-3	2-Butanone	6.50	UD	6.50	130	ug/L
56-23-5	Carbon Tetrachloride	1.30	UD	1.30	25.0	ug/L
156-59-2	cis-1,2-Dichloroethene	1.30	UD	1.30	25.0	ug/L
74-97-5	Bromochloromethane	0.90	UD	0.90	25.0	ug/L
67-66-3	Chloroform	1.30	UD	1.30	25.0	ug/L
71-55-6	1,1,1-Trichloroethane	0.95	UD	0.95	25.0	ug/L
108-87-2	Methylcyclohexane	0.95	UD	0.95	25.0	ug/L
71-43-2	Benzene	20.3	JD	0.80	25.0	ug/L
107-06-2	1,2-Dichloroethane	1.20	UD	1.20	25.0	ug/L
79-01-6	Trichloroethene	1.60	UD	1.60	25.0	ug/L
78-87-5	1,2-Dichloropropane	0.95	UD	0.95	25.0	ug/L
75-27-4	Bromodichloromethane	1.20	UD	1.20	25.0	ug/L
108-10-1	4-Methyl-2-Pentanone	3.80	UD	3.80	130	ug/L
108-88-3	Toluene	17.3	JD	0.90	25.0	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C11-1DL	SDG No.:	P5006
Lab Sample ID:	P5006-12DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044117.D	5		12/04/24 15:20	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	1.10	UD	1.10	25.0	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.90	UD	0.90	25.0	ug/L
79-00-5	1,1,2-Trichloroethane	1.10	UD	1.10	25.0	ug/L
591-78-6	2-Hexanone	5.70	UD	5.70	130	ug/L
124-48-1	Dibromochloromethane	0.90	UD	0.90	25.0	ug/L
106-93-4	1,2-Dibromoethane	0.80	UD	0.80	25.0	ug/L
127-18-4	Tetrachloroethene	1.30	UD	1.30	25.0	ug/L
108-90-7	Chlorobenzene	0.65	UD	0.65	25.0	ug/L
100-41-4	Ethyl Benzene	180	D	0.80	25.0	ug/L
179601-23-1	m/p-Xylenes	37.1	JD	1.60	50.0	ug/L
95-47-6	o-Xylene	44.9	D	0.70	25.0	ug/L
100-42-5	Styrene	0.80	UD	0.80	25.0	ug/L
75-25-2	Bromoform	1.10	UD	1.10	25.0	ug/L
98-82-8	Isopropylbenzene	15.9	JD	0.65	25.0	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	1.40	UD	1.40	25.0	ug/L
541-73-1	1,3-Dichlorobenzene	1.20	UD	1.20	25.0	ug/L
106-46-7	1,4-Dichlorobenzene	1.40	UD	1.40	25.0	ug/L
95-50-1	1,2-Dichlorobenzene	0.95	UD	0.95	25.0	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	2.30	UD	2.30	25.0	ug/L
120-82-1	1,2,4-Trichlorobenzene	2.10	UD	2.10	25.0	ug/L
87-61-6	1,2,3-Trichlorobenzene	2.60	UD	2.60	25.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.9		70 (74) - 130 (125)	106%	SPK: 50
1868-53-7	Dibromofluoromethane	44.7		70 (75) - 130 (124)	89%	SPK: 50
2037-26-5	Toluene-d8	49.9		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.9		70 (77) - 130 (121)	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	101000	5.544			
540-36-3	1,4-Difluorobenzene	197000	6.757			
3114-55-4	Chlorobenzene-d5	175000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	75600	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-1DL		SDG No.:	P5006	
Lab Sample ID:	P5006-12DL		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044117.D	5		12/04/24 15:20	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C11-2	SDG No.:	P5006
Lab Sample ID:	P5006-13	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044101.D	1		12/03/24 18:23	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	19.2	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.10	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	15.6		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	12.0		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-2		SDG No.:	P5006	
Lab Sample ID:	P5006-13		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044101.D	1		12/03/24 18:23	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	140		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	28.7		0.31	10.0	ug/L
95-47-6	o-Xylene	32.5		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	11.4		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.6		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	36.4		70 (75) - 130 (124)	73%	SPK: 50
2037-26-5	Toluene-d8	50.3		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.4		70 (77) - 130 (121)	109%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	107000	5.544			
540-36-3	1,4-Difluorobenzene	214000	6.757			
3114-55-4	Chlorobenzene-d5	192000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	90500	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-2		SDG No.:	P5006	
Lab Sample ID:	P5006-13		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044101.D	1		12/03/24 18:23	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C11-2RE	SDG No.:	P5006
Lab Sample ID:	P5006-13RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044124.D	1		12/04/24 18:02	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	20.2	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.20	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	16.6		0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	12.3		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C11-2RE	SDG No.:	P5006
Lab Sample ID:	P5006-13RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044124.D	1		12/04/24 18:02	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	150		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	30.4		0.31	10.0	ug/L
95-47-6	o-Xylene	33.8		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	12.0		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.1		70 (74) - 130 (125)	104%	SPK: 50
1868-53-7	Dibromofluoromethane	35.7		70 (75) - 130 (124)	71%	SPK: 50
2037-26-5	Toluene-d8	50.4		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.7		70 (77) - 130 (121)	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	102000	5.55			
540-36-3	1,4-Difluorobenzene	197000	6.757			
3114-55-4	Chlorobenzene-d5	178000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	81300	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-2RE		SDG No.:	P5006	
Lab Sample ID:	P5006-13RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044124.D	1		12/04/24 18:02	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-3		SDG No.:	P5006	
Lab Sample ID:	P5006-14		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044102.D	1		12/03/24 18:46	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	17.8	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	2.50	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	2.60	J	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C11-3	SDG No.:	P5006
Lab Sample ID:	P5006-14	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044102.D	1		12/03/24 18:46	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	45.9		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	9.70	J	0.31	10.0	ug/L
95-47-6	o-Xylene	10.8		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	4.40	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.4		70 (74) - 130 (125)	105%	SPK: 50
1868-53-7	Dibromofluoromethane	36.0		70 (75) - 130 (124)	72%	SPK: 50
2037-26-5	Toluene-d8	50.2		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.7		70 (77) - 130 (121)	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	105000	5.544			
540-36-3	1,4-Difluorobenzene	206000	6.757			
3114-55-4	Chlorobenzene-d5	184000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	84900	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-3		SDG No.:	P5006	
Lab Sample ID:	P5006-14		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044102.D	1		12/03/24 18:46	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-3RE		SDG No.:	P5006	
Lab Sample ID:	P5006-14RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044125.D	1		12/04/24 18:25	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	17.5	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	2.50	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	2.40	J	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-3RE		SDG No.:	P5006	
Lab Sample ID:	P5006-14RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044125.D	1		12/04/24 18:25	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	44.8		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	8.90	J	0.31	10.0	ug/L
95-47-6	o-Xylene	10.5		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	4.30	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.5		70 (74) - 130 (125)	103%	SPK: 50
1868-53-7	Dibromofluoromethane	35.4		70 (75) - 130 (124)	71%	SPK: 50
2037-26-5	Toluene-d8	50.3		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.7		70 (77) - 130 (121)	107%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	108000	5.544			
540-36-3	1,4-Difluorobenzene	208000	6.757			
3114-55-4	Chlorobenzene-d5	188000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	86400	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C11-3RE		SDG No.:	P5006	
Lab Sample ID:	P5006-14RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044125.D	1		12/04/24 18:25	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-1		SDG No.:	P5006	
Lab Sample ID:	P5006-15		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044103.D	1		12/03/24 19:09	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	31.7		1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	2.60	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	2.20	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	2.50	J	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-1		SDG No.:	P5006	
Lab Sample ID:	P5006-15		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044103.D	1		12/03/24 19:09	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	21.4		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	6.50	J	0.31	10.0	ug/L
95-47-6	o-Xylene	7.90		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	2.00	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.8		70 (74) - 130 (125)	108%	SPK: 50
1868-53-7	Dibromofluoromethane	33.6	*	70 (75) - 130 (124)	67%	SPK: 50
2037-26-5	Toluene-d8	50.3		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		70 (77) - 130 (121)	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	109000	5.544			
540-36-3	1,4-Difluorobenzene	218000	6.757			
3114-55-4	Chlorobenzene-d5	193000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	90000	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-1		SDG No.:	P5006	
Lab Sample ID:	P5006-15		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044103.D	1		12/03/24 19:09	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-1RE		SDG No.:	P5006	
Lab Sample ID:	P5006-15RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044126.D	1		12/04/24 18:48	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	33.1		1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	2.70	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	2.20	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	2.50	J	0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C12-1RE	SDG No.:	P5006
Lab Sample ID:	P5006-15RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044126.D	1		12/04/24 18:48	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	21.4		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	6.20	J	0.31	10.0	ug/L
95-47-6	o-Xylene	7.40		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	2.20	J	0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.9		70 (74) - 130 (125)	108%	SPK: 50
1868-53-7	Dibromofluoromethane	32.3	*	70 (75) - 130 (124)	65%	SPK: 50
2037-26-5	Toluene-d8	49.5		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.3		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	104000	5.544			
540-36-3	1,4-Difluorobenzene	209000	6.757			
3114-55-4	Chlorobenzene-d5	183000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	81700	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-1RE		SDG No.:	P5006	
Lab Sample ID:	P5006-15RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044126.D	1		12/04/24 18:48	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C12-2	SDG No.:	P5006
Lab Sample ID:	P5006-16	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044104.D	1		12/03/24 19:32	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	19.9	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.50	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	4.40	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	11.1		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-2		SDG No.:	P5006	
Lab Sample ID:	P5006-16		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044104.D	1		12/03/24 19:32	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	88.4		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	26.1		0.31	10.0	ug/L
95-47-6	o-Xylene	26.0		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	7.00		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.9		70 (74) - 130 (125)	108%	SPK: 50
1868-53-7	Dibromofluoromethane	33.2	*	70 (75) - 130 (124)	66%	SPK: 50
2037-26-5	Toluene-d8	50.0		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.8		70 (77) - 130 (121)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	108000	5.544			
540-36-3	1,4-Difluorobenzene	214000	6.757			
3114-55-4	Chlorobenzene-d5	187000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	88000	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-2		SDG No.:	P5006	
Lab Sample ID:	P5006-16		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044104.D	1		12/03/24 19:32	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-2RE		SDG No.:	P5006	
Lab Sample ID:	P5006-16RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044127.D	1		12/04/24 19:11	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	22.0	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	1.50	J	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	4.50	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	11.9		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-2RE		SDG No.:	P5006	
Lab Sample ID:	P5006-16RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044127.D	1		12/04/24 19:11	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	92.5		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	26.9		0.31	10.0	ug/L
95-47-6	o-Xylene	27.1		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	7.50		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.6		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	33.2	*	70 (75) - 130 (124)	66%	SPK: 50
2037-26-5	Toluene-d8	50.6		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.8		70 (77) - 130 (121)	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	100000	5.544			
540-36-3	1,4-Difluorobenzene	201000	6.757			
3114-55-4	Chlorobenzene-d5	178000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	78300	12.024			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-2RE		SDG No.:	P5006	
Lab Sample ID:	P5006-16RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044127.D	1		12/04/24 19:11	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-3		SDG No.:	P5006	
Lab Sample ID:	P5006-17		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044105.D	1		12/03/24 19:55	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	11.4	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	3.20	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	8.10		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C12-3	SDG No.:	P5006
Lab Sample ID:	P5006-17	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044105.D	1		12/03/24 19:55	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	65.0		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	19.1		0.31	10.0	ug/L
95-47-6	o-Xylene	18.3		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	5.10		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.5		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	36.6		70 (75) - 130 (124)	73%	SPK: 50
2037-26-5	Toluene-d8	50.2		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.6		70 (77) - 130 (121)	107%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	119000	5.55			
540-36-3	1,4-Difluorobenzene	232000	6.757			
3114-55-4	Chlorobenzene-d5	206000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	96300	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-3		SDG No.:	P5006	
Lab Sample ID:	P5006-17		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044105.D	1		12/03/24 19:55	VX120324

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	11/25/24
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	11/26/24
Client Sample ID:	SPLP-C12-3RE	SDG No.:	P5006
Lab Sample ID:	P5006-17RE	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	SPLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044128.D	1		12/04/24 19:34	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	5.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	5.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	5.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	5.00	ug/L
67-64-1	Acetone	14.1	J	1.40	25.0	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	5.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	5.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	5.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	5.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	5.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	5.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	5.00	ug/L
71-43-2	Benzene	3.50	J	0.16	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	5.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	5.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	25.0	ug/L
108-88-3	Toluene	9.10		0.18	5.00	ug/L

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-3RE		SDG No.:	P5006	
Lab Sample ID:	P5006-17RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044128.D	1		12/04/24 19:34	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	5.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	25.0	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	5.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	5.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	5.00	ug/L
100-41-4	Ethyl Benzene	75.9		0.16	5.00	ug/L
179601-23-1	m/p-Xylenes	21.2		0.31	10.0	ug/L
95-47-6	o-Xylene	20.8		0.14	5.00	ug/L
100-42-5	Styrene	0.16	U	0.16	5.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	5.00	ug/L
98-82-8	Isopropylbenzene	6.30		0.13	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.0		70 (74) - 130 (125)	102%	SPK: 50
1868-53-7	Dibromofluoromethane	37.0		70 (75) - 130 (124)	74%	SPK: 50
2037-26-5	Toluene-d8	49.0		70 (86) - 130 (113)	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.5		70 (77) - 130 (121)	99%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	105000	5.543			
540-36-3	1,4-Difluorobenzene	200000	6.757			
3114-55-4	Chlorobenzene-d5	170000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	72600	12.018			

Report of Analysis

Client:	ENTACT		Date Collected:	11/25/24	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	11/26/24	
Client Sample ID:	SPLP-C12-3RE		SDG No.:	P5006	
Lab Sample ID:	P5006-17RE		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	SPLP VOA	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044128.D	1		12/04/24 19:34	VX120424

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	P5006	OrderDate:	11/26/2024 11:21:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Jarod Stanfield	Location:	L51,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P5006-01	SPLP-C4-2	Water			11/25/24			11/26/24
			SPLP VOA	8260D			12/02/24	
P5006-01RE	SPLP-C4-2RE	Water			11/25/24			11/26/24
			SPLP VOA	8260D			12/03/24	
P5006-02	SPLP-C4-3	Water			11/25/24			11/26/24
			SPLP VOA	8260D			12/03/24	
P5006-03	SPLP-C5-1	Water			11/25/24			11/26/24
			SPLP VOA	8260D			12/03/24	
P5006-03RE	SPLP-C5-1RE	Water			11/25/24			11/26/24
			SPLP VOA	8260D			12/04/24	
P5006-04	SPLP-C5-2	Water			11/25/24			11/26/24
			SPLP VOA	8260D			12/04/24	
P5006-04RE	SPLP-C5-2RE	Water			11/25/24			11/26/24
			SPLP VOA	8260D			12/04/24	
P5006-05	SPLP-C5-3	Water			11/25/24			11/26/24
			SPLP VOA	8260D			12/03/24	
P5006-05RE	SPLP-C5-3RE	Water			11/25/24			11/26/24
			SPLP VOA	8260D			12/04/24	
P5006-06	SPLP-C6-1	Water			11/25/24			11/26/24
			SPLP VOA	8260D			12/03/24	
P5006-06RE	SPLP-C6-1RE	Water			11/25/24			11/26/24
			SPLP VOA	8260D			12/04/24	
P5006-07	SPLP-C6-2	Water			11/25/24			11/26/24

LAB CHRONICLE

			SPLP VOA	8260D	12/03/24	
P5006-08	SPLP-C6-3	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/03/24	
P5006-08RE	SPLP-C6-3RE	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/04/24	
P5006-09	SPLP-C10-1	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/03/24	
P5006-09DL	SPLP-C10-1DL	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/04/24	
P5006-10	SPLP-C10-2	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/03/24	
P5006-10DL	SPLP-C10-2DL	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/04/24	
P5006-11	SPLP-C10-3	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/03/24	
P5006-11DL	SPLP-C10-3DL	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/04/24	
P5006-12	SPLP-C11-1	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/03/24	
P5006-12DL	SPLP-C11-1DL	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/04/24	
P5006-13	SPLP-C11-2	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/03/24	
P5006-13RE	SPLP-C11-2RE	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/04/24	
P5006-14	SPLP-C11-3	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/03/24	
P5006-14RE	SPLP-C11-3RE	Water			11/25/24	11/26/24
			SPLP VOA	8260D	12/04/24	

LAB CHRONICLE

P5006-15	SPLP-C12-1	Water			11/25/24		11/26/24
			SPLP VOA	8260D		12/03/24	
P5006-15RE	SPLP-C12-1RE	Water			11/25/24		11/26/24
			SPLP VOA	8260D		12/04/24	
P5006-16	SPLP-C12-2	Water			11/25/24		11/26/24
			SPLP VOA	8260D		12/03/24	
P5006-16RE	SPLP-C12-2RE	Water			11/25/24		11/26/24
			SPLP VOA	8260D		12/04/24	
P5006-17	SPLP-C12-3	Water			11/25/24		11/26/24
			SPLP VOA	8260D		12/03/24	
P5006-17RE	SPLP-C12-3RE	Water			11/25/24		11/26/24
			SPLP VOA	8260D		12/04/24	
P5006-18	TW-WTS-02	Water			11/25/24		11/26/24
			VOCMS Group4	8260-Low		11/27/24	