



Lab Name:	<u>CHEMTECH</u>			Contract:	<u>ENTA05</u>		
Lab Code:	<u>CHEM</u>	Case No.:	<u>P5006</u>	SAS No.:	<u>P5006</u>	SDG NO.:	<u>P5006</u>
Lab File ID:	<u>VX044080.D</u>			Date Analyzed:	<u>12/03/2024</u>		
Instrument ID:	<u>MSVOA_X</u>			Time Analyzed:	<u>09:44</u>		
GC Column:	DB-624UI	ID:	0.18 (mm)	Heated Purge:	(Y/N)	N	

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	121618	5.54	213358	6.75	176784	10.05
UPPER LIMIT	243236	6.044	426716	7.251	353568	10.549
LOWER LIMIT	60809	5.044	106679	6.251	88392	9.549
EPA SAMPLE NO.						
SPLP-C4-2RE	117073	5.54	223904	6.76	189590	10.05
SPLP-C4-3	112757	5.54	220233	6.75	195286	10.05
SPLP-C5-1	114961	5.54	227218	6.76	195298	10.05
SPLP-C5-3	106174	5.54	209275	6.76	186004	10.05
SPLP-C6-1	114452	5.54	221169	6.76	192410	10.05
SPLP-C6-2	112488	5.55	223144	6.76	201427	10.05
SPLP-C6-3	118097	5.54	224861	6.76	199085	10.05
SPLP-C10-1	114433	5.54	224500	6.76	195198	10.05
SPLP-C10-2	103374	5.55	203524	6.76	182011	10.06
SPLP-C10-3	105039	5.54	203705	6.76	175202	10.06
SPLP-C11-1	101739	5.55	199677	6.76	178071	10.06
SPLP-C11-2	107038	5.54	213687	6.76	192034	10.06
SPLP-C11-3	104678	5.54	205665	6.76	184006	10.06
SPLP-C12-1	109453	5.54	218071	6.76	193451	10.05
SPLP-C12-2	108046	5.54	214154	6.76	187128	10.06
SPLP-C12-3	118829	5.55	231895	6.76	206380	10.05
VX1203WBL01	113150	5.54	222913	6.76	189304	10.05
VX1203WBS01	121696	5.54	217179	6.75	182356	10.05

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P5006 SAS No.: P5006 SDG NO.: P5006
 Lab File ID: VX044080.D Date Analyzed: 12/03/2024
 Instrument ID: MSVOA_X Time Analyzed: 09:44
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	121618	5.54	213358	6.75	176784	10.05
UPPER LIMIT	243236	6.044	426716	7.251	353568	10.549
LOWER LIMIT	60809	5.044	106679	6.251	88392	9.549
EPA SAMPLE NO.						

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.