

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P5006 SAS No.: P5006 SDG NO.: P5006
 Lab File ID: VX044080.D Date Analyzed: 12/03/2024
 Instrument ID: MSVOA_X Time Analyzed: 09:44
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	80742	12.018				
UPPER LIMIT	161484	12.518				
LOWER LIMIT	40371	11.518				
EPA SAMPLE NO.						
SPLP-C4-2RE	81966	12.02				
SPLP-C4-3	89527	12.02				
SPLP-C5-1	88091	12.02				
SPLP-C5-3	88193	12.02				
SPLP-C6-1	79632	12.02				
SPLP-C6-2	87397	12.02				
SPLP-C6-3	90167	12.02				
SPLP-C10-1	94536	12.02				
SPLP-C10-2	84668	12.02				
SPLP-C10-3	82520	12.02				
SPLP-C11-1	86182	12.02				
SPLP-C11-2	90537	12.02				
SPLP-C11-3	84942	12.02				
SPLP-C12-1	89950	12.02				
SPLP-C12-2	87999	12.02				
SPLP-C12-3	96338	12.02				
VX1203WBL01	79760	12.02				
VX1203WBS01	84290	12.02				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P5006 SAS No.: P5006 SDG NO.: P5006
 Lab File ID: VX044080.D Date Analyzed: 12/03/2024
 Instrument ID: MSVOA_X Time Analyzed: 09:44
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	80742	12.018				
UPPER LIMIT	161484	12.518				
LOWER LIMIT	40371	11.518				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.