

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>ENTA05</u>				
Lab Code:	<u>CHEM</u>	Case No.:	<u>P5006</u>	SAS No.:	<u>P5006</u>	SDG NO.:	<u>P5006</u>
Lab File ID:	<u>VX044109.D</u>	Date Analyzed:	<u>12/04/2024</u>				
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>12:12</u>				
GC Column:	DB-624UI	ID:	0.18 (mm)	Heated Purge:	(Y/N)	N	

	IS4 AREA #	RT #				
12 HOUR STD	85782	12.018				
UPPER LIMIT	171564	12.518				
LOWER LIMIT	42891	11.518				
EPA SAMPLE NO.						
SPLP-C5-1RE	80127	12.02				
SPLP-C5-2	81312	12.02				
SPLP-C5-2RE	68386	12.02				
SPLP-C5-3RE	79148	12.02				
SPLP-C6-1RE	72597	12.02				
SPLP-C6-3RE	85975	12.02				
SPLP-C10-1DL	79045	12.02				
SPLP-C10-2DL	74917	12.02				
SPLP-C10-3DL	78940	12.02				
SPLP-C11-1DL	75611	12.02				
SPLP-C11-2RE	81304	12.02				
SPLP-C11-3RE	86376	12.02				
SPLP-C12-1RE	81734	12.02				
SPLP-C12-2RE	78276	12.02				
SPLP-C12-3RE	72597	12.02				
VX1204WBL01	77704	12.02				
VX1204WBS01	84893	12.02				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: P5006 SAS No.: P5006 SDG NO.: P5006
 Lab File ID: VX044109.D Date Analyzed: 12/04/2024
 Instrument ID: MSVOA_X Time Analyzed: 12:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	85782	12.018				
UPPER LIMIT	171564	12.518				
LOWER LIMIT	42891	11.518				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.