

Surrogate Summary

SDG No.: P5006

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5006-18	TW-WTS-02	1,2-Dichloroethane-d4	50	51.1	102	70 (74)	130 (125)
		Dibromofluoromethane	50	47.4	95	70 (75)	130 (124)
		Toluene-d8	50	51.9	104	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.9	98	70 (77)	130 (121)
VX1127WBL01	VX1127WBL01	1,2-Dichloroethane-d4	50	50.8	102	70 (74)	130 (125)
		Dibromofluoromethane	50	46.6	93	70 (75)	130 (124)
		Toluene-d8	50	48.9	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.2	94	70 (77)	130 (121)
VX1127WBS01	VX1127WBS01	1,2-Dichloroethane-d4	50	54.7	109	70 (74)	130 (125)
		Dibromofluoromethane	50	53.5	107	70 (75)	130 (124)
		Toluene-d8	50	53.1	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.6	107	70 (77)	130 (121)
VX1127WBSD0	VX1127WBSD01	1,2-Dichloroethane-d4	50	51.5	103	70 (74)	130 (125)
		Dibromofluoromethane	50	46.2	92	70 (75)	130 (124)
		Toluene-d8	50	47.1	94	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.5	97	70 (77)	130 (121)

