

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJW4Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133720 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV012				ID: CCV082				ID: CCV083		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	94	0	400000	390000	99	0	400000	100	1
Antimony	1000	1000	102	0	5000	5000	100	0	5000	99	0
Arsenic	1000	920	92	1	5000	4900	97	0	4800	97	0
Barium	520	520	101	0	10000	10000	101	1	10000	100	0
Beryllium	510	540	106	0	500	510	101	0	510	102	0
Cadmium	510	480	93	0	2500	2500	101	0	2500	101	0
Calcium	10000	10000	103	0	400000	410000	104	0	400000	101	0
Chromium	520	570	109	0	15500	15000	99	0	16000	102	2
Cobalt	520	500	95	0	2500	2500	101	0	2500	101	0
Copper	510	500	97	0	15000	15000	100	0	15000	100	0
Iron	10000	11000	107	0	400000	400000	101	0	410000	102	0
Lead	1000	980	98	0	25000	25000	100	0	25000	100	0
Magnesium	6000	5600	93	1	400000	420000	105	0	410000	103	0
Manganese	520	540	104	0	15000	15000	101	0	15000	99	0
Nickel	530	510	96	0	2500	2500	101	0	2500	100	0
Potassium	9900	9100	92	0	150000	150000	101	0	150000	102	0
Selenium	1000	970	97	1	5000	4900	98	0	4900	97	0
Silver	250	240	98	0	1250	1300	102	0	1300	102	2
Sodium	10000	10000	101	1	400000	410000	102	0	400000	100	0
Thallium	1000	1000	102	1	5000	5000	99	0	4900	98	0
Vanadium	500	520	104	1	2500	2500	100	0	2500	100	0
Zinc	1000	1000	100	1	15000	15000	101	0	15000	99	2

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJW4Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133720 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV084				ID: CCV085		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	100	1	400000	101	1
Antimony					5000	4800	97	0	4700	94	0
Arsenic					5000	4700	95	0	4700	94	1
Barium					10000	9800	98	0	9800	98	0
Beryllium					500	510	101	1	510	101	1
Cadmium					2500	2500	99	0	2400	97	0
Calcium					400000	400000	99	0	390000	97	0
Chromium					15500	16000	101	1	16000	100	0
Cobalt					2500	2500	99	0	2400	97	0
Copper					15000	15000	99	0	15000	99	0
Iron					400000	400000	101	0	400000	101	0
Lead					25000	25000	98	0	24000	97	0
Magnesium					400000	380000	96	0	380000	95	0
Manganese					15000	15000	97	0	14000	94	0
Nickel					2500	2400	98	0	2400	96	0
Potassium					150000	150000	101	0	150000	101	1
Selenium					5000	4700	95	0	4600	93	1
Silver					1250	1200	99	1	1200	97	0
Sodium					400000	390000	98	0	390000	96	0
Thallium					5000	4800	95	0	4700	93	0
Vanadium					2500	2400	97	0	2400	97	0
Zinc					15000	14000	94	1	14000	91	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJW4Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133883 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV014				ID: CCV089				ID: CCV090		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	97	1	400000	410000	102	0	400000	100	0
Antimony	1000	910	91	0	5000	4900	98	0	4800	97	0
Arsenic	1000	1000	104	0	5000	5000	100	1	4900	99	0
Barium	520	520	100	1	10000	10000	104	0	10000	101	0
Beryllium	510	520	102	1	500	510	103	0	500	101	1
Cadmium	510	480	93	0	2500	2500	102	0	2500	100	0
Calcium	10000	10000	102	2	400000	410000	104	0	410000	101	0
Chromium	520	550	106	0	15500	16000	102	0	16000	101	0
Cobalt	520	500	96	0	2500	2600	102	1	2500	101	0
Copper	510	500	98	2	15000	15000	102	0	15000	100	0
Iron	10000	11000	106	1	400000	420000	104	0	410000	102	0
Lead	1000	970	97	0	25000	25000	101	0	25000	100	0
Magnesium	6000	6000	101	1	400000	420000	105	0	410000	102	0
Manganese	520	530	101	1	15000	15000	102	0	15000	100	0
Nickel	530	510	96	0	2500	2600	103	0	2500	101	0
Potassium	9900	9400	95	4	150000	160000	104	0	150000	102	0
Selenium	1000	930	93	0	5000	4900	97	0	4800	95	1
Silver	250	250	99	0	1250	1300	102	0	1300	100	0
Sodium	10000	10000	101	2	400000	410000	102	0	400000	100	0
Thallium	1000	980	98	0	5000	5000	100	0	4900	98	0
Vanadium	500	510	101	1	2500	2600	102	0	2500	101	0
Zinc	1000	1100	110	1	15000	15000	99	1	15000	99	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJW4Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133883 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV091				ID: CCV092		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	99	1	400000	100	1
Antimony					5000	4800	96	1	4800	97	0
Arsenic					5000	4800	97	0	4900	98	0
Barium					10000	10000	100	1	10000	100	1
Beryllium					500	500	100	2	500	100	1
Cadmium					2500	2500	99	0	2500	100	0
Calcium					400000	400000	100	1	410000	102	1
Chromium					15500	15000	100	0	16000	101	1
Cobalt					2500	2500	101	0	2500	101	0
Copper					15000	15000	99	1	15000	100	1
Iron					400000	400000	101	1	410000	102	1
Lead					25000	25000	99	0	25000	99	0
Magnesium					400000	400000	99	1	400000	101	1
Manganese					15000	15000	97	1	15000	98	1
Nickel					2500	2500	100	0	2500	101	0
Potassium					150000	150000	100	1	150000	102	1
Selenium					5000	4700	94	0	4800	95	0
Silver					1250	1200	99	0	1300	100	0
Sodium					400000	390000	99	1	400000	99	1
Thallium					5000	4900	98	0	4900	98	0
Vanadium					2500	2500	100	1	2500	101	1
Zinc					15000	14000	95	1	14000	93	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJW4Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133883 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV093				ID: CCV094		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	99	0	400000	99	0
Antimony					5000	4800	96	0	4800	97	0
Arsenic					5000	4900	98	0	5000	100	0
Barium					10000	10000	100	0	10000	100	0
Beryllium					500	500	100	1	500	100	0
Cadmium					2500	2500	99	0	2500	100	0
Calcium					400000	400000	101	0	400000	100	0
Chromium					15500	15000	100	0	15000	99	1
Cobalt					2500	2500	100	0	2500	101	0
Copper					15000	15000	100	0	15000	99	0
Iron					400000	400000	100	0	400000	100	0
Lead					25000	25000	99	0	25000	99	0
Magnesium					400000	400000	100	0	400000	99	0
Manganese					15000	15000	97	0	14000	96	0
Nickel					2500	2500	100	0	2500	101	0
Potassium					150000	150000	101	0	150000	101	0
Selenium					5000	4700	94	0	4700	95	0
Silver					1250	1200	99	0	1200	98	1
Sodium					400000	400000	99	0	390000	98	0
Thallium					5000	4900	97	0	4900	98	0
Vanadium					2500	2500	101	1	2500	100	0
Zinc					15000	14000	92	1	14000	91	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51879 MA No. : SDG No.: MBHJW4Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB133883 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV095				ID: CCV096		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	99	0	400000	99	0
Antimony					5000	4800	95	0	4700	95	0
Arsenic					5000	4900	98	0	4900	98	0
Barium					10000	9900	99	0	9900	99	0
Beryllium					500	490	98	1	490	98	1
Cadmium					2500	2500	98	0	2500	98	0
Calcium					400000	400000	99	0	400000	99	0
Chromium					15500	15000	98	0	15000	98	0
Cobalt					2500	2500	100	0	2500	100	1
Copper					15000	15000	97	0	14000	97	0
Iron					400000	390000	98	0	390000	98	0
Lead					25000	24000	98	0	24000	98	1
Magnesium					400000	390000	98	0	390000	97	0
Manganese					15000	14000	94	0	14000	93	0
Nickel					2500	2500	99	0	2500	99	1
Potassium					150000	150000	99	1	150000	99	0
Selenium					5000	4600	93	0	4600	93	0
Silver					1250	1200	97	0	1200	97	0
Sodium					400000	390000	98	1	390000	97	0
Thallium					5000	4800	96	0	4800	97	0
Vanadium					2500	2400	97	0	2500	98	0
Zinc					15000	14000	91	0	14000	91	1