

Surrogate Summary

SDG No.: P5018

Client: New York City DEP of Environmental Protection/BWS

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5018-01	14B-1	1,2-Dichloroethane-d4	30	29.9	100	91	110
		Toluene-d8	30	28.8	96	91	112
		4-Bromofluorobenzene	30	24.9	83	63	112
P5018-02	14B-2	1,2-Dichloroethane-d4	30	29.8	99	91	110
		Toluene-d8	30	28.1	94	91	112
		4-Bromofluorobenzene	30	23.9	80	63	112
P5018-03	14B-3	1,2-Dichloroethane-d4	30	30.6	102	91	110
		Toluene-d8	30	28.6	95	91	112
		4-Bromofluorobenzene	30	23.9	80	63	112
P5018-04	14B-4	1,2-Dichloroethane-d4	30	31.1	104	91	110
		Toluene-d8	30	28.3	94	91	112
		4-Bromofluorobenzene	30	23.3	78	63	112
VN1127WBL01	VN1127WBL01	1,2-Dichloroethane-d4	30	31.1	104	91	110
		Toluene-d8	30	28.6	95	91	112
		4-Bromofluorobenzene	30	24.5	82	63	112
VN1127WBS01	VN1127WBS01	1,2-Dichloroethane-d4	30	30.4	101	91	110
		Toluene-d8	30	29.6	99	91	112
		4-Bromofluorobenzene	30	29.4	98	63	112
VN1127WBSD0	VN1127WBSD01	1,2-Dichloroethane-d4	30	30.5	102	91	110
		Toluene-d8	30	29.9	100	91	112
		4-Bromofluorobenzene	30	28.3	94	63	112