

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NEWY17
 Lab Code: CHEM Case No.: P5018 SAS No.: P5018 SDG NO.: P5018
 Lab File ID: VN085052.D Date Analyzed: 11/27/2024
 Instrument ID: MSVOA_N Time Analyzed: 11:44
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	28768	7.81	155364	9.09	140160	11.87
UPPER LIMIT	57536	8.312	310728	9.594	280320	12.365
LOWER LIMIT	14384	7.312	77682	8.594	70080	11.365
EPA SAMPLE NO.						
14B-1	33381	7.81	177303	9.10	151222	11.87
14B-2	21362	7.81	109546	9.09	94649	11.87
14B-3	24374	7.81	124625	9.10	107923	11.87
14B-4	24391	7.81	127521	9.09	110334	11.87
VN1127WBL01	26189	7.81	138613	9.09	117498	11.87
VN1127WBS01	32660	7.81	179004	9.09	161319	11.86
VN1127WBSD01	29451	7.81	158864	9.10	140518	11.87

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.