

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5020

Project: Monthly 2024

Run No.: LB133679

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133679BS							
BOD5	mg/L	198	177		89	1	84.6-115.4	11/28/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5020

Project: Monthly 2024

Run No.: LB133715

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133715BS							
TSS	mg/L	550	536		98	1	90-110	12/03/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5020

Project: Monthly 2024

Run No.: LB133752

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133752BS							
TPH	mg/L	20.0	16.7		84	1	78-114	12/05/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5020

Project: Monthly 2024

Run No.: LB133752

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133752BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	12/05/2024