

DATA PACKAGE

VOLATILE ORGANICS
SEMI-VOLATILE ORGANICS

PROJECT NAME : RFK BRIDGE

LIRO ENGINEERS, INC.

690 Delaware Ave.

Buffalo, NY - 14209

Phone No: 716-882-5476

ORDER ID : P5021

ATTENTION : Martin Wesolowski



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) VOCMS Group1- Case Narrative	4
2.2) SVOCMS Group1- Case Narrative	6
3) Qualifier Page	8
4) QA Checklist	9
5) VOCMS Group1 Data	10
6) SVOCMS Group1 Data	58
7) Shipping Document	101
7.1) CHAIN OF CUSTODY	102
7.2) Lab Certificate	103
7.3) Internal COC	104

1
2
3
4
5
6
7

Cover Page

Order ID : P5021

Project ID : RFK Bridge

Client : LiRo Engineers, Inc.

Lab Sample Number

P5021-01
P5021-02
P5021-03
P5021-04
P5021-05
P5021-06

Client Sample Number

MW-06
MW-08
MW-10
MW-11
MW-13
MW-12

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : N. N. Pandya

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 2:15 pm, Dec 09, 2024

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

LiRo Engineers, Inc.

Project Name: RFK Bridge

Project # N/A

Chemtech Project # P5021

Test Name: VOCMS Group1

A. Number of Samples and Date of Receipt:

6 Water samples were received on 11/27/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOCMS Group1 and VOCMS Group1. This data package contains results for VOCMS Group1.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of VOCMS Group1 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD for met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

The Samples MW-08 was Initially diluted at 10x due to past history of this sample containing high amounts of Benzene, O Xylene, MP Xylenes, etc and rereported as first analyis.

Sample MW-12 was diluted due to high concentration.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount

for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature N. N. Pandya

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 2:20 pm, Dec 09, 2024

CASE NARRATIVE

LiRo Engineers, Inc.

Project Name: RFK Bridge

Project # N/A

Chemtech Project # P5021

Test Name: SVOCMS Group1

A. Number of Samples and Date of Receipt:

6 Water samples were received on 11/27/2024.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOCMS Group1 and VOCMS Group1. This data package contains results for SVOCMS Group1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df The analysis of SVOCMS Group1 was based on method 8270E and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate for {PB165296BSD} with File ID: BF140684.D met requirements for all samples except for Acenaphthylene[110%], Benzo(a)anthracene[112%], Benzo(k)fluoranthene[108%], Chrysene[110%] and Pyrene[105%], marginally high therefore no corrective action was taken.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <15% for the Initial

Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 15% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature N. N. Pandya

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 2:21 pm, Dec 09, 2024

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: P5021

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 12/09/2024

LAB CHRONICLE

OrderID:	P5021	OrderDate:	11/27/2024 10:41:00 AM
Client:	LiRo Engineers, Inc.	Project:	RFK Bridge
Contact:	Martin Wesolowski	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P5021-01	MW-06	Water	VOCMS Group1	8260-Low	11/26/24		12/02/24	11/27/24
P5021-02	MW-08	Water	VOCMS Group1	8260-Low	11/26/24		12/02/24	11/27/24
P5021-03	MW-10	Water	VOCMS Group1	8260-Low	11/26/24		12/02/24	11/27/24
P5021-04	MW-11	Water	VOCMS Group1	8260-Low	11/26/24		12/02/24	11/27/24
P5021-05	MW-13	Water	VOCMS Group1	8260-Low	11/26/24		12/02/24	11/27/24
P5021-06	MW-12	Water	VOCMS Group1	8260-Low	11/26/24		11/27/24	11/27/24
P5021-06DL	MW-12DL	Water	VOCMS Group1	8260-Low	11/26/24		12/02/24	11/27/24

Hit Summary Sheet
SW-846

SDG No.: P5021
Client: LiRo Engineers, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	MW-06							
P5021-01	MW-06	Water	Benzene	8.40		0.16	1.00	ug/L
P5021-01	MW-06	Water	Toluene	1.40		0.18	1.00	ug/L
P5021-01	MW-06	Water	Ethyl Benzene	2.00		0.16	1.00	ug/L
P5021-01	MW-06	Water	m/p-Xylenes	3.50		0.31	2.00	ug/L
P5021-01	MW-06	Water	o-Xylene	1.50		0.14	1.00	ug/L
P5021-01	MW-06	Water	Isopropylbenzene	20.5		0.13	1.00	ug/L
P5021-01	MW-06	Water	n-propylbenzene	48.7		0.14	1.00	ug/L
P5021-01	MW-06	Water	1,3,5-Trimethylbenzene	1.10		0.18	1.00	ug/L
P5021-01	MW-06	Water	tert-Butylbenzene	0.42	J	0.17	1.00	ug/L
P5021-01	MW-06	Water	1,2,4-Trimethylbenzene	2.10		0.18	1.00	ug/L
P5021-01	MW-06	Water	sec-Butylbenzene	3.40		0.17	1.00	ug/L
P5021-01	MW-06	Water	n-Butylbenzene	2.20		0.22	1.00	ug/L
P5021-01	MW-06	Water	Naphthalene	77.5		0.59	1.00	ug/L
			Total Voc :		173			
			Total Concentration:		173			
Client ID:	MW-08							
P5021-02	MW-08	Water	Methyl tert-butyl Ether	25.0		1.60	10.0	ug/L
P5021-02	MW-08	Water	Benzene	180		1.60	10.0	ug/L
P5021-02	MW-08	Water	Toluene	9.70	J	1.80	10.0	ug/L
P5021-02	MW-08	Water	Ethyl Benzene	270		1.60	10.0	ug/L
P5021-02	MW-08	Water	m/p-Xylenes	380		3.10	20.0	ug/L
P5021-02	MW-08	Water	o-Xylene	270		1.40	10.0	ug/L
P5021-02	MW-08	Water	Isopropylbenzene	58.1		1.30	10.0	ug/L
P5021-02	MW-08	Water	n-propylbenzene	130		1.40	10.0	ug/L
P5021-02	MW-08	Water	1,2,4-Trimethylbenzene	610		1.80	10.0	ug/L
P5021-02	MW-08	Water	n-Butylbenzene	13.9		2.20	10.0	ug/L
P5021-02	MW-08	Water	Naphthalene	400		5.90	10.0	ug/L
			Total Voc :		2350			
			Total Concentration:		2350			
Client ID:	MW-10							
P5021-03	MW-10	Water	n-propylbenzene	0.27	J	0.14	1.00	ug/L
P5021-03	MW-10	Water	1,3,5-Trimethylbenzene	0.33	J	0.18	1.00	ug/L
P5021-03	MW-10	Water	tert-Butylbenzene	0.34	J	0.17	1.00	ug/L
P5021-03	MW-10	Water	sec-Butylbenzene	0.33	J	0.17	1.00	ug/L
P5021-03	MW-10	Water	n-Butylbenzene	1.10		0.22	1.00	ug/L
P5021-03	MW-10	Water	Naphthalene	2.10		0.59	1.00	ug/L

Hit Summary Sheet

SW-846

SDG No.: P5021

Client: LiRo Engineers, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Total Voc :				4.47				
Total Concentration:				4.47				
Client ID:	MW-13							
P5021-05	MW-13	Water	Methyl tert-butyl Ether	2.20		0.16	1.00	ug/L
P5021-05	MW-13	Water	Benzene	15.3		0.16	1.00	ug/L
P5021-05	MW-13	Water	Toluene	0.68	J	0.18	1.00	ug/L
P5021-05	MW-13	Water	Ethyl Benzene	19.0		0.16	1.00	ug/L
P5021-05	MW-13	Water	m/p-Xylenes	3.60		0.31	2.00	ug/L
P5021-05	MW-13	Water	o-Xylene	4.40		0.14	1.00	ug/L
P5021-05	MW-13	Water	Isopropylbenzene	1.50		0.13	1.00	ug/L
P5021-05	MW-13	Water	n-propylbenzene	2.20		0.14	1.00	ug/L
P5021-05	MW-13	Water	1,3,5-Trimethylbenzene	0.35	J	0.18	1.00	ug/L
P5021-05	MW-13	Water	1,2,4-Trimethylbenzene	1.60		0.18	1.00	ug/L
P5021-05	MW-13	Water	n-Butylbenzene	0.52	J	0.22	1.00	ug/L
Total Voc :				51.4				
Total Concentration:				51.4				
Client ID:	MW-12							
P5021-06	MW-12	Water	Benzene	330	E	0.16	1.00	ug/L
P5021-06	MW-12	Water	Toluene	38.9		0.18	1.00	ug/L
P5021-06	MW-12	Water	Ethyl Benzene	760	E	0.16	1.00	ug/L
P5021-06	MW-12	Water	m/p-Xylenes	250		0.31	2.00	ug/L
P5021-06	MW-12	Water	o-Xylene	100		0.14	1.00	ug/L
P5021-06	MW-12	Water	Isopropylbenzene	23.1		0.13	1.00	ug/L
P5021-06	MW-12	Water	n-propylbenzene	51.9		0.14	1.00	ug/L
P5021-06	MW-12	Water	1,3,5-Trimethylbenzene	2.70		0.18	1.00	ug/L
P5021-06	MW-12	Water	1,2,4-Trimethylbenzene	450	E	0.18	1.00	ug/L
P5021-06	MW-12	Water	sec-Butylbenzene	2.90		0.17	1.00	ug/L
P5021-06	MW-12	Water	p-Isopropyltoluene	3.50		0.15	1.00	ug/L
P5021-06	MW-12	Water	n-Butylbenzene	4.80		0.22	1.00	ug/L
P5021-06	MW-12	Water	Naphthalene	250	E	0.59	1.00	ug/L
Total Voc :				2270				
Total Concentration:				2270				
Client ID:	MW-12DL							
P5021-06DL	MW-12DL	Water	Benzene	270	D	3.20	20.0	ug/L
P5021-06DL	MW-12DL	Water	Toluene	30.0	D	3.60	20.0	ug/L
P5021-06DL	MW-12DL	Water	Ethyl Benzene	580	D	3.20	20.0	ug/L
P5021-06DL	MW-12DL	Water	m/p-Xylenes	190	D	6.20	40.0	ug/L
P5021-06DL	MW-12DL	Water	o-Xylene	85.2	D	2.80	20.0	ug/L

Hit Summary Sheet SW-846

SDG No.: P5021

Client: LiRo Engineers, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
P5021-06DL	MW-12DL	Water	Isopropylbenzene	18.6	JD	2.60	20.0	ug/L
P5021-06DL	MW-12DL	Water	n-propylbenzene	36.6	D	2.80	20.0	ug/L
P5021-06DL	MW-12DL	Water	1,2,4-Trimethylbenzene	350	D	3.60	20.0	ug/L
P5021-06DL	MW-12DL	Water	Naphthalene	180	D	11.8	20.0	ug/L
Total Voc :				1740				
Total Concentration:				1740				



SAMPLE DATA

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-06	SDG No.:	P5021
Lab Sample ID:	P5021-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044059.D	1		12/02/24 12:35	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	1.00	U	0.16	1.00	ug/L
71-43-2	Benzene	8.40		0.16	1.00	ug/L
108-88-3	Toluene	1.40		0.18	1.00	ug/L
100-41-4	Ethyl Benzene	2.00		0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	3.50		0.31	2.00	ug/L
95-47-6	o-Xylene	1.50		0.14	1.00	ug/L
98-82-8	Isopropylbenzene	20.5		0.13	1.00	ug/L
103-65-1	n-propylbenzene	48.7		0.14	1.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	1.10		0.18	1.00	ug/L
98-06-6	tert-Butylbenzene	0.42	J	0.17	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	2.10		0.18	1.00	ug/L
135-98-8	sec-Butylbenzene	3.40		0.17	1.00	ug/L
99-87-6	p-Isopropyltoluene	1.00	U	0.15	1.00	ug/L
104-51-8	n-Butylbenzene	2.20		0.22	1.00	ug/L
91-20-3	Naphthalene	77.5		0.59	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.0		74 - 125	102%	SPK: 50
1868-53-7	Dibromofluoromethane	47.5		75 - 124	95%	SPK: 50
2037-26-5	Toluene-d8	48.9		86 - 113	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.0		77 - 121	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	138000	5.543			
540-36-3	1,4-Difluorobenzene	259000	6.757			
3114-55-4	Chlorobenzene-d5	223000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	102000	12.024			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-06	SDG No.:	P5021
Lab Sample ID:	P5021-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOCMS Group1

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044059.D	1		12/02/24 12:35	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-08	SDG No.:	P5021
Lab Sample ID:	P5021-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044057.D	10		12/02/24 11:49	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	25.0		1.60	10.0	ug/L
71-43-2	Benzene	180		1.60	10.0	ug/L
108-88-3	Toluene	9.70	J	1.80	10.0	ug/L
100-41-4	Ethyl Benzene	270		1.60	10.0	ug/L
179601-23-1	m/p-Xylenes	380		3.10	20.0	ug/L
95-47-6	o-Xylene	270		1.40	10.0	ug/L
98-82-8	Isopropylbenzene	58.1		1.30	10.0	ug/L
103-65-1	n-propylbenzene	130		1.40	10.0	ug/L
108-67-8	1,3,5-Trimethylbenzene	10.0	U	1.80	10.0	ug/L
98-06-6	tert-Butylbenzene	10.0	U	1.70	10.0	ug/L
95-63-6	1,2,4-Trimethylbenzene	610		1.80	10.0	ug/L
135-98-8	sec-Butylbenzene	10.0	U	1.70	10.0	ug/L
99-87-6	p-Isopropyltoluene	10.0	U	1.50	10.0	ug/L
104-51-8	n-Butylbenzene	13.9		2.20	10.0	ug/L
91-20-3	Naphthalene	400		5.90	10.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.9		74 - 125	98%	SPK: 50
1868-53-7	Dibromofluoromethane	48.0		75 - 124	96%	SPK: 50
2037-26-5	Toluene-d8	48.8		86 - 113	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.6		77 - 121	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	123000	5.544			
540-36-3	1,4-Difluorobenzene	229000	6.757			
3114-55-4	Chlorobenzene-d5	199000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	92200	12.018			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-08	SDG No.:	P5021
Lab Sample ID:	P5021-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOCMS Group1

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044057.D	10		12/02/24 11:49	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-10	SDG No.:	P5021
Lab Sample ID:	P5021-03	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044060.D	1		12/02/24 12:58	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	1.00	U	0.16	1.00	ug/L
71-43-2	Benzene	1.00	U	0.16	1.00	ug/L
108-88-3	Toluene	1.00	U	0.18	1.00	ug/L
100-41-4	Ethyl Benzene	1.00	U	0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	2.00	U	0.31	2.00	ug/L
95-47-6	o-Xylene	1.00	U	0.14	1.00	ug/L
98-82-8	Isopropylbenzene	1.00	U	0.13	1.00	ug/L
103-65-1	n-propylbenzene	0.27	J	0.14	1.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	0.33	J	0.18	1.00	ug/L
98-06-6	tert-Butylbenzene	0.34	J	0.17	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	1.00	U	0.18	1.00	ug/L
135-98-8	sec-Butylbenzene	0.33	J	0.17	1.00	ug/L
99-87-6	p-Isopropyltoluene	1.00	U	0.15	1.00	ug/L
104-51-8	n-Butylbenzene	1.10		0.22	1.00	ug/L
91-20-3	Naphthalene	2.10		0.59	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.8		74 - 125	104%	SPK: 50
1868-53-7	Dibromofluoromethane	46.4		75 - 124	93%	SPK: 50
2037-26-5	Toluene-d8	49.3		86 - 113	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.9		77 - 121	100%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	121000	5.544			
540-36-3	1,4-Difluorobenzene	237000	6.751			
3114-55-4	Chlorobenzene-d5	202000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	89000	12.018			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-10	SDG No.:	P5021
Lab Sample ID:	P5021-03	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOCMS Group1

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044060.D	1		12/02/24 12:58	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-11	SDG No.:	P5021
Lab Sample ID:	P5021-04	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044061.D	1		12/02/24 13:21	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	1.00	U	0.16	1.00	ug/L
71-43-2	Benzene	1.00	U	0.16	1.00	ug/L
108-88-3	Toluene	1.00	U	0.18	1.00	ug/L
100-41-4	Ethyl Benzene	1.00	U	0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	2.00	U	0.31	2.00	ug/L
95-47-6	o-Xylene	1.00	U	0.14	1.00	ug/L
98-82-8	Isopropylbenzene	1.00	U	0.13	1.00	ug/L
103-65-1	n-propylbenzene	1.00	U	0.14	1.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	1.00	U	0.18	1.00	ug/L
98-06-6	tert-Butylbenzene	1.00	U	0.17	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	1.00	U	0.18	1.00	ug/L
135-98-8	sec-Butylbenzene	1.00	U	0.17	1.00	ug/L
99-87-6	p-Isopropyltoluene	1.00	U	0.15	1.00	ug/L
104-51-8	n-Butylbenzene	1.00	U	0.22	1.00	ug/L
91-20-3	Naphthalene	1.00	U	0.59	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.9		74 - 125	104%	SPK: 50
1868-53-7	Dibromofluoromethane	47.0		75 - 124	94%	SPK: 50
2037-26-5	Toluene-d8	51.0		86 - 113	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.6		77 - 121	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	120000	5.544			
540-36-3	1,4-Difluorobenzene	230000	6.757			
3114-55-4	Chlorobenzene-d5	209000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	91500	12.018			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-11	SDG No.:	P5021
Lab Sample ID:	P5021-04	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOCMS Group1

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044061.D	1		12/02/24 13:21	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-13	SDG No.:	P5021
Lab Sample ID:	P5021-05	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044062.D	1		12/02/24 13:44	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	2.20		0.16	1.00	ug/L
71-43-2	Benzene	15.3		0.16	1.00	ug/L
108-88-3	Toluene	0.68	J	0.18	1.00	ug/L
100-41-4	Ethyl Benzene	19.0		0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	3.60		0.31	2.00	ug/L
95-47-6	o-Xylene	4.40		0.14	1.00	ug/L
98-82-8	Isopropylbenzene	1.50		0.13	1.00	ug/L
103-65-1	n-propylbenzene	2.20		0.14	1.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	0.35	J	0.18	1.00	ug/L
98-06-6	tert-Butylbenzene	1.00	U	0.17	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	1.60		0.18	1.00	ug/L
135-98-8	sec-Butylbenzene	1.00	U	0.17	1.00	ug/L
99-87-6	p-Isopropyltoluene	1.00	U	0.15	1.00	ug/L
104-51-8	n-Butylbenzene	0.52	J	0.22	1.00	ug/L
91-20-3	Naphthalene	1.00	U	0.59	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.1		74 - 125	106%	SPK: 50
1868-53-7	Dibromofluoromethane	46.7		75 - 124	93%	SPK: 50
2037-26-5	Toluene-d8	50.2		86 - 113	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.1		77 - 121	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	126000	5.544			
540-36-3	1,4-Difluorobenzene	245000	6.757			
3114-55-4	Chlorobenzene-d5	217000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	94600	12.018			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-13	SDG No.:	P5021
Lab Sample ID:	P5021-05	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOCMS Group1

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044062.D	1		12/02/24 13:44	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-12	SDG No.:	P5021
Lab Sample ID:	P5021-06	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044049.D	1		11/27/24 18:29	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	1.00	U	0.16	1.00	ug/L
71-43-2	Benzene	330	E	0.16	1.00	ug/L
108-88-3	Toluene	38.9		0.18	1.00	ug/L
100-41-4	Ethyl Benzene	760	E	0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	250		0.31	2.00	ug/L
95-47-6	o-Xylene	100		0.14	1.00	ug/L
98-82-8	Isopropylbenzene	23.1		0.13	1.00	ug/L
103-65-1	n-propylbenzene	51.9		0.14	1.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	2.70		0.18	1.00	ug/L
98-06-6	tert-Butylbenzene	1.00	U	0.17	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	450	E	0.18	1.00	ug/L
135-98-8	sec-Butylbenzene	2.90		0.17	1.00	ug/L
99-87-6	p-Isopropyltoluene	3.50		0.15	1.00	ug/L
104-51-8	n-Butylbenzene	4.80		0.22	1.00	ug/L
91-20-3	Naphthalene	250	E	0.59	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.2		74 - 125	106%	SPK: 50
1868-53-7	Dibromofluoromethane	46.4		75 - 124	93%	SPK: 50
2037-26-5	Toluene-d8	50.2		86 - 113	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.3		77 - 121	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	113000	5.55			
540-36-3	1,4-Difluorobenzene	218000	6.757			
3114-55-4	Chlorobenzene-d5	192000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	91200	12.018			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-12	SDG No.:	P5021
Lab Sample ID:	P5021-06	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOCMS Group1

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044049.D	1		11/27/24 18:29	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-12DL	SDG No.:	P5021
Lab Sample ID:	P5021-06DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044058.D	20		12/02/24 12:12	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	20.0	UD	3.20	20.0	ug/L
71-43-2	Benzene	270	D	3.20	20.0	ug/L
108-88-3	Toluene	30.0	D	3.60	20.0	ug/L
100-41-4	Ethyl Benzene	580	D	3.20	20.0	ug/L
179601-23-1	m/p-Xylenes	190	D	6.20	40.0	ug/L
95-47-6	o-Xylene	85.2	D	2.80	20.0	ug/L
98-82-8	Isopropylbenzene	18.6	JD	2.60	20.0	ug/L
103-65-1	n-propylbenzene	36.6	D	2.80	20.0	ug/L
108-67-8	1,3,5-Trimethylbenzene	20.0	UD	3.60	20.0	ug/L
98-06-6	tert-Butylbenzene	20.0	UD	3.40	20.0	ug/L
95-63-6	1,2,4-Trimethylbenzene	350	D	3.60	20.0	ug/L
135-98-8	sec-Butylbenzene	20.0	UD	3.40	20.0	ug/L
99-87-6	p-Isopropyltoluene	20.0	UD	3.00	20.0	ug/L
104-51-8	n-Butylbenzene	20.0	UD	4.40	20.0	ug/L
91-20-3	Naphthalene	180	D	11.8	20.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.2		74 - 125	102%	SPK: 50
1868-53-7	Dibromofluoromethane	46.3		75 - 124	93%	SPK: 50
2037-26-5	Toluene-d8	49.4		86 - 113	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.1		77 - 121	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	126000	5.544			
540-36-3	1,4-Difluorobenzene	245000	6.757			
3114-55-4	Chlorobenzene-d5	215000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	97900	12.018			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-12DL	SDG No.:	P5021
Lab Sample ID:	P5021-06DL	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044058.D	20		12/02/24 12:12	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **P5021**

Client: **LiRo Engineers, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5021-01	MW-06	1,2-Dichloroethane-d4	50	51.0	102	74	125
		Dibromofluoromethane	50	47.5	95	75	124
		Toluene-d8	50	48.9	98	86	113
		4-Bromofluorobenzene	50	52.0	104	77	121
P5021-02	MW-08	1,2-Dichloroethane-d4	50	48.9	98	74	125
		Dibromofluoromethane	50	48.0	96	75	124
		Toluene-d8	50	48.8	98	86	113
		4-Bromofluorobenzene	50	51.6	103	77	121
P5021-03	MW-10	1,2-Dichloroethane-d4	50	51.8	104	74	125
		Dibromofluoromethane	50	46.4	93	75	124
		Toluene-d8	50	49.3	99	86	113
		4-Bromofluorobenzene	50	50.0	100	77	121
P5021-04	MW-11	1,2-Dichloroethane-d4	50	51.9	104	74	125
		Dibromofluoromethane	50	47.0	94	75	124
		Toluene-d8	50	51.0	102	86	113
		4-Bromofluorobenzene	50	51.6	103	77	121
P5021-05	MW-13	1,2-Dichloroethane-d4	50	53.1	106	74	125
		Dibromofluoromethane	50	46.7	93	75	124
		Toluene-d8	50	50.2	100	86	113
		4-Bromofluorobenzene	50	51.1	102	77	121
P5021-06	MW-12	1,2-Dichloroethane-d4	50	53.2	106	74	125
		Dibromofluoromethane	50	46.4	93	75	124
		Toluene-d8	50	50.2	100	86	113
		4-Bromofluorobenzene	50	52.3	105	77	121
P5021-06DL	MW-12DL	1,2-Dichloroethane-d4	50	51.2	102	74	125
		Dibromofluoromethane	50	46.3	93	75	124
		Toluene-d8	50	49.4	99	86	113
		4-Bromofluorobenzene	50	51.1	102	77	121
VX1127WBL01	VX1127WBL01	1,2-Dichloroethane-d4	50	50.8	102	74	125
		Dibromofluoromethane	50	46.6	93	75	124
		Toluene-d8	50	48.9	98	86	113
		4-Bromofluorobenzene	50	47.2	94	77	121
VX1127WBS01	VX1127WBS01	1,2-Dichloroethane-d4	50	54.7	109	74	125
		Dibromofluoromethane	50	53.5	107	75	124
		Toluene-d8	50	53.1	106	86	113
		4-Bromofluorobenzene	50	53.6	107	77	121
VX1202WBL01	VX1202WBL01	1,2-Dichloroethane-d4	50	49.6	99	74	125
		Dibromofluoromethane	50	47.3	95	75	124
		Toluene-d8	50	48.6	97	86	113
		4-Bromofluorobenzene	50	46.5	93	77	121
VX1202WBS01	VX1202WBS01	1,2-Dichloroethane-d4	50	54.9	110	74	125
		Dibromofluoromethane	50	53.7	107	75	124
		Toluene-d8	50	50.8	102	86	113
		4-Bromofluorobenzene	50	52.8	105	77	121
VX1202WBSD0	VX1202WBSD01	1,2-Dichloroethane-d4	50	55.6	111	74	125
		Dibromofluoromethane	50	54.7	109	75	124
		Toluene-d8	50	53.5	107	86	113
		4-Bromofluorobenzene	50	55.0	110	77	121

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5021
Client: LiRo Engineers, Inc.
Analytical Method: SW8260-Low Datafile : VX044028.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1127WBS01	Methyl tert-butyl Ether	20	21.6	ug/L	108			78	114	
	Benzene	20	20.9	ug/L	104			82	109	
	Toluene	20	21.2	ug/L	106			82	110	
	Ethyl Benzene	20	21.1	ug/L	106			83	109	
	m/p-Xylenes	40	42.2	ug/L	106			82	110	
	o-Xylene	20	21.0	ug/L	105			83	109	
	Isopropylbenzene	20	21.5	ug/L	108			83	112	
	N-propylbenzene	20	21.6	ug/L	108			83	112	
	1,3,5-Trimethylbenzene	20	21.9	ug/L	110			85	112	
	tert-Butylbenzene	20	21.5	ug/L	108			83	112	
	1,2,4-Trimethylbenzene	20	21.9	ug/L	110			85	111	
	Sec-butylbenzene	20	21.5	ug/L	108			81	114	
	p-Isopropyltoluene	20	22.4	ug/L	112			78	116	
	n-Butylbenzene	20	21.5	ug/L	108			75	115	
	Naphthalene	20	20.8	ug/L	104			78	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5021
Client: LiRo Engineers, Inc.
Analytical Method: SW8260-Low Datafile : VX044055.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1202WBS01	Methyl tert-butyl Ether	20	20.3	ug/L	102			78	114	
	Benzene	20	19.4	ug/L	97			82	109	
	Toluene	20	19.4	ug/L	97			82	110	
	Ethyl Benzene	20	20.7	ug/L	104			83	109	
	m/p-Xylenes	40	41.5	ug/L	104			82	110	
	o-Xylene	20	21.3	ug/L	106			83	109	
	Isopropylbenzene	20	20.6	ug/L	103			83	112	
	N-propylbenzene	20	20.7	ug/L	104			83	112	
	1,3,5-Trimethylbenzene	20	21.3	ug/L	106			85	112	
	tert-Butylbenzene	20	21.3	ug/L	106			83	112	
	1,2,4-Trimethylbenzene	20	21.2	ug/L	106			85	111	
	Sec-butylbenzene	20	20.8	ug/L	104			81	114	
	p-Isopropyltoluene	20	21.5	ug/L	108			78	116	
	n-Butylbenzene	20	20.1	ug/L	101			75	115	
	Naphthalene	20	20.3	ug/L	102			78	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5021
Client: LiRo Engineers, Inc.
Analytical Method: SW8260-Low Datafile : VX044056.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1202WBSD01	Methyl tert-butyl Ether	20	21.5	ug/L	108	6		78	114	20
	Benzene	20	19.8	ug/L	99	2		82	109	20
	Toluene	20	20.8	ug/L	104	7		82	110	20
	Ethyl Benzene	20	21.0	ug/L	105	1		83	109	20
	m/p-Xylenes	40	40.8	ug/L	102	2		82	110	20
	o-Xylene	20	21.1	ug/L	106	0		83	109	20
	Isopropylbenzene	20	21.1	ug/L	106	3		83	112	20
	N-propylbenzene	20	20.9	ug/L	104	0		83	112	20
	1,3,5-Trimethylbenzene	20	21.3	ug/L	106	0		85	112	20
	tert-Butylbenzene	20	21.4	ug/L	107	1		83	112	20
	1,2,4-Trimethylbenzene	20	21.2	ug/L	106	0		85	111	20
	Sec-butylbenzene	20	21.2	ug/L	106	2		81	114	20
	p-Isopropyltoluene	20	21.8	ug/L	109	1		78	116	20
	n-Butylbenzene	20	21.0	ug/L	105	4		75	115	20
	Naphthalene	20	20.6	ug/L	103	1		78	119	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX1127WBL01

Lab Name: CHEMTECH

Contract: LIRO01

Lab Code: CHEM Case No.: P5021

SAS No.: P5021 SDG NO.: P5021

Lab File ID: VX044026.D

Lab Sample ID: VX1127WBL01

Date Analyzed: 11/27/2024

Time Analyzed: 09:38

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX1127WBS01	VX1127WBS01	VX044028.D	11/27/2024
MW-12	P5021-06	VX044049.D	11/27/2024

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX1202WBL01

Lab Name: CHEMTECH

Contract: LIRO01

Lab Code: CHEM Case No.: P5021

SAS No.: P5021 SDG NO.: P5021

Lab File ID: VX044054.D

Lab Sample ID: VX1202WBL01

Date Analyzed: 12/02/2024

Time Analyzed: 10:37

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX1202WBS01	VX1202WBS01	VX044055.D	12/02/2024
VX1202WBSD01	VX1202WBSD01	VX044056.D	12/02/2024
MW-08	P5021-02	VX044057.D	12/02/2024
MW-12DL	P5021-06DL	VX044058.D	12/02/2024
MW-06	P5021-01	VX044059.D	12/02/2024
MW-10	P5021-03	VX044060.D	12/02/2024
MW-11	P5021-04	VX044061.D	12/02/2024
MW-13	P5021-05	VX044062.D	12/02/2024

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: LIR001
Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021
Lab File ID: VX043924.D BFB Injection Date: 11/21/2024
Instrument ID: MSVOA_X BFB Injection Time: 08:51
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.4
75	30.0 - 60.0% of mass 95	54.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	70.5
175	5.0 - 9.0% of mass 174	4.8 (6.8) 1
176	95.0 - 101.0% of mass 174	67.9 (96.4) 1
177	5.0 - 9.0% of mass 176	4.9 (7.3) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC001	VSTDICC001	VX043925.D	11/21/2024	09:47
VSTDICC005	VSTDICC005	VX043926.D	11/21/2024	10:14
VSTDICC020	VSTDICC020	VX043927.D	11/21/2024	10:37
VSTDICCC050	VSTDICCC050	VX043928.D	11/21/2024	11:17
VSTDICC100	VSTDICC100	VX043929.D	11/21/2024	11:40
VSTDICC150	VSTDICC150	VX043930.D	11/21/2024	12:03

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: LIR001
Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021
Lab File ID: VX044023.D BFB Injection Date: 11/27/2024
Instrument ID: MSVOA_X BFB Injection Time: 07:52
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.5
75	30.0 - 60.0% of mass 95	56.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.8 (1.1) 1
174	50.0 - 100.0% of mass 95	76.1
175	5.0 - 9.0% of mass 174	5.6 (7.3) 1
176	95.0 - 101.0% of mass 174	73 (96) 1
177	5.0 - 9.0% of mass 176	4.8 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX044024.D	11/27/2024	08:46
VX1127WBL01	VX1127WBL01	VX044026.D	11/27/2024	09:38
VX1127WBS01	VX1127WBS01	VX044028.D	11/27/2024	10:27
MW-12	P5021-06	VX044049.D	11/27/2024	18:29

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: LIR001
Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021
Lab File ID: VX044051.D BFB Injection Date: 12/02/2024
Instrument ID: MSVOA_X BFB Injection Time: 08:35
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.2
75	30.0 - 60.0% of mass 95	56.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.5 (0.6) 1
174	50.0 - 100.0% of mass 95	74.9
175	5.0 - 9.0% of mass 174	5.3 (7.1) 1
176	95.0 - 101.0% of mass 174	71.7 (95.7) 1
177	5.0 - 9.0% of mass 176	4.6 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX044052.D	12/02/2024	09:45
VX1202WBL01	VX1202WBL01	VX044054.D	12/02/2024	10:37
VX1202WBS01	VX1202WBS01	VX044055.D	12/02/2024	11:00
VX1202WBSD01	VX1202WBSD01	VX044056.D	12/02/2024	11:27
MW-08	P5021-02	VX044057.D	12/02/2024	11:49
MW-12DL	P5021-06DL	VX044058.D	12/02/2024	12:12
MW-06	P5021-01	VX044059.D	12/02/2024	12:35
MW-10	P5021-03	VX044060.D	12/02/2024	12:58
MW-11	P5021-04	VX044061.D	12/02/2024	13:21
MW-13	P5021-05	VX044062.D	12/02/2024	13:44

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LIRO01
Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021
Lab File ID: VX044024.D Date Analyzed: 11/27/2024
Instrument ID: MSVOA_X Time Analyzed: 08:46
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	130547	5.54	218884	6.75	184783	10.05
UPPER LIMIT	261094	6.037	437768	7.251	369566	10.549
LOWER LIMIT	65273.5	5.037	109442	6.251	92391.5	9.549
EPA SAMPLE NO.						
MW-12	112926	5.55	218400	6.76	192264	10.06
VX1127WBL01	125569	5.54	244455	6.75	212718	10.05
VX1127WBS01	106843	5.54	192894	6.76	162801	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LIRO01
Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021
Lab File ID: VX044024.D Date Analyzed: 11/27/2024
Instrument ID: MSVOA_X Time Analyzed: 08:46
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	86408	12.018				
UPPER LIMIT	172816	12.518				
LOWER LIMIT	43204	11.518				
EPA SAMPLE NO.						
MW-12	91181	12.02				
VX1127WBL01	92049	12.02				
VX1127WBS01	74629	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LIRO01
Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021
Lab File ID: VX044052.D Date Analyzed: 12/02/2024
Instrument ID: MSVOA_X Time Analyzed: 09:45
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	140021	5.54	242098	6.75	204952	10.05
UPPER LIMIT	280042	6.043	484196	7.25	409904	10.549
LOWER LIMIT	70010.5	5.043	121049	6.25	102476	9.549
EPA SAMPLE NO.						
MW-06	137934	5.54	259323	6.76	223099	10.06
MW-08	123344	5.54	229138	6.76	199363	10.05
MW-10	121483	5.54	237218	6.75	202018	10.06
MW-11	120070	5.54	229971	6.76	208817	10.05
MW-13	125796	5.54	244811	6.76	217402	10.05
MW-12DL	125735	5.54	244727	6.76	215420	10.06
VX1202WBL01	137622	5.54	262897	6.76	221008	10.05
VX1202WBS01	133564	5.55	236701	6.76	191057	10.06
VX1202WBSD01	119891	5.54	216106	6.75	186524	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LIRO01
Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021
Lab File ID: VX044052.D Date Analyzed: 12/02/2024
Instrument ID: MSVOA_X Time Analyzed: 09:45
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	98861	12.018				
UPPER LIMIT	197722	12.518				
LOWER LIMIT	49430.5	11.518				
EPA SAMPLE NO.						
MW-06	102016	12.02				
MW-08	92150	12.02				
MW-10	89001	12.02				
MW-11	91525	12.02				
MW-13	94569	12.02				
MW-12DL	97904	12.02				
VX1202WBL01	93273	12.02				
VX1202WBS01	90073	12.02				
VX1202WBSD01	85631	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	VX1127WBL01	SDG No.:	P5021
Lab Sample ID:	VX1127WBL01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044026.D	1		11/27/24 09:38	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	1.00	U	0.16	1.00	ug/L
71-43-2	Benzene	1.00	U	0.16	1.00	ug/L
108-88-3	Toluene	1.00	U	0.18	1.00	ug/L
100-41-4	Ethyl Benzene	1.00	U	0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	2.00	U	0.31	2.00	ug/L
95-47-6	o-Xylene	1.00	U	0.14	1.00	ug/L
98-82-8	Isopropylbenzene	1.00	U	0.13	1.00	ug/L
103-65-1	n-propylbenzene	1.00	U	0.14	1.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	1.00	U	0.18	1.00	ug/L
98-06-6	tert-Butylbenzene	1.00	U	0.17	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	1.00	U	0.18	1.00	ug/L
135-98-8	sec-Butylbenzene	1.00	U	0.17	1.00	ug/L
99-87-6	p-Isopropyltoluene	1.00	U	0.15	1.00	ug/L
104-51-8	n-Butylbenzene	1.00	U	0.22	1.00	ug/L
91-20-3	Naphthalene	1.00	U	0.59	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.8		74 - 125	102%	SPK: 50
1868-53-7	Dibromofluoromethane	46.6		75 - 124	93%	SPK: 50
2037-26-5	Toluene-d8	48.9		86 - 113	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.2		77 - 121	94%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	126000	5.538			
540-36-3	1,4-Difluorobenzene	244000	6.751			
3114-55-4	Chlorobenzene-d5	213000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	92000	12.018			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	VX1127WBL01	SDG No.:	P5021
Lab Sample ID:	VX1127WBL01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044026.D	1		11/27/24 09:38	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	VX1202WBL01	SDG No.:	P5021
Lab Sample ID:	VX1202WBL01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044054.D	1		12/02/24 10:37	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	1.00	U	0.16	1.00	ug/L
71-43-2	Benzene	1.00	U	0.16	1.00	ug/L
108-88-3	Toluene	1.00	U	0.18	1.00	ug/L
100-41-4	Ethyl Benzene	1.00	U	0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	2.00	U	0.31	2.00	ug/L
95-47-6	o-Xylene	1.00	U	0.14	1.00	ug/L
98-82-8	Isopropylbenzene	1.00	U	0.13	1.00	ug/L
103-65-1	n-propylbenzene	1.00	U	0.14	1.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	1.00	U	0.18	1.00	ug/L
98-06-6	tert-Butylbenzene	1.00	U	0.17	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	1.00	U	0.18	1.00	ug/L
135-98-8	sec-Butylbenzene	1.00	U	0.17	1.00	ug/L
99-87-6	p-Isopropyltoluene	1.00	U	0.15	1.00	ug/L
104-51-8	n-Butylbenzene	1.00	U	0.22	1.00	ug/L
91-20-3	Naphthalene	1.00	U	0.59	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.6		74 - 125	99%	SPK: 50
1868-53-7	Dibromofluoromethane	47.3		75 - 124	95%	SPK: 50
2037-26-5	Toluene-d8	48.6		86 - 113	97%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.5		77 - 121	93%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	138000	5.544			
540-36-3	1,4-Difluorobenzene	263000	6.757			
3114-55-4	Chlorobenzene-d5	221000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	93300	12.018			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	VX1202WBL01	SDG No.:	P5021
Lab Sample ID:	VX1202WBL01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044054.D	1		12/02/24 10:37	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	VX1127WBS01	SDG No.:	P5021
Lab Sample ID:	VX1127WBS01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044028.D	1		11/27/24 10:27	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	21.6		0.16	1.00	ug/L
71-43-2	Benzene	20.9		0.16	1.00	ug/L
108-88-3	Toluene	21.2		0.18	1.00	ug/L
100-41-4	Ethyl Benzene	21.1		0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	42.2		0.31	2.00	ug/L
95-47-6	o-Xylene	21.0		0.14	1.00	ug/L
98-82-8	Isopropylbenzene	21.5		0.13	1.00	ug/L
103-65-1	n-propylbenzene	21.6		0.14	1.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	21.9		0.18	1.00	ug/L
98-06-6	tert-Butylbenzene	21.5		0.17	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	21.9		0.18	1.00	ug/L
135-98-8	sec-Butylbenzene	21.5		0.17	1.00	ug/L
99-87-6	p-Isopropyltoluene	22.4		0.15	1.00	ug/L
104-51-8	n-Butylbenzene	21.5		0.22	1.00	ug/L
91-20-3	Naphthalene	20.8		0.59	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.7		74 - 125	109%	SPK: 50
1868-53-7	Dibromofluoromethane	53.5		75 - 124	107%	SPK: 50
2037-26-5	Toluene-d8	53.1		86 - 113	106%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.6		77 - 121	107%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	107000	5.544			
540-36-3	1,4-Difluorobenzene	193000	6.757			
3114-55-4	Chlorobenzene-d5	163000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	74600	12.018			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	VX1127WBS01	SDG No.:	P5021
Lab Sample ID:	VX1127WBS01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044028.D	1		11/27/24 10:27	VX112724

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	VX1202WBS01	SDG No.:	P5021
Lab Sample ID:	VX1202WBS01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044055.D	1		12/02/24 11:00	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	20.3		0.16	1.00	ug/L
71-43-2	Benzene	19.4		0.16	1.00	ug/L
108-88-3	Toluene	19.4		0.18	1.00	ug/L
100-41-4	Ethyl Benzene	20.7		0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	41.5		0.31	2.00	ug/L
95-47-6	o-Xylene	21.3		0.14	1.00	ug/L
98-82-8	Isopropylbenzene	20.6		0.13	1.00	ug/L
103-65-1	n-propylbenzene	20.7		0.14	1.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	21.3		0.18	1.00	ug/L
98-06-6	tert-Butylbenzene	21.3		0.17	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	21.2		0.18	1.00	ug/L
135-98-8	sec-Butylbenzene	20.8		0.17	1.00	ug/L
99-87-6	p-Isopropyltoluene	21.5		0.15	1.00	ug/L
104-51-8	n-Butylbenzene	20.1		0.22	1.00	ug/L
91-20-3	Naphthalene	20.3		0.59	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.9		74 - 125	110%	SPK: 50
1868-53-7	Dibromofluoromethane	53.7		75 - 124	107%	SPK: 50
2037-26-5	Toluene-d8	50.8		86 - 113	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		77 - 121	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	134000	5.55			
540-36-3	1,4-Difluorobenzene	237000	6.757			
3114-55-4	Chlorobenzene-d5	191000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	90100	12.018			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	VX1202WBS01	SDG No.:	P5021
Lab Sample ID:	VX1202WBS01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044055.D	1		12/02/24 11:00	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	VX1202WBSD01	SDG No.:	P5021
Lab Sample ID:	VX1202WBSD01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044056.D	1		12/02/24 11:27	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	21.5		0.16	1.00	ug/L
71-43-2	Benzene	19.8		0.16	1.00	ug/L
108-88-3	Toluene	20.8		0.18	1.00	ug/L
100-41-4	Ethyl Benzene	21.0		0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	40.8		0.31	2.00	ug/L
95-47-6	o-Xylene	21.1		0.14	1.00	ug/L
98-82-8	Isopropylbenzene	21.1		0.13	1.00	ug/L
103-65-1	n-propylbenzene	20.9		0.14	1.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	21.3		0.18	1.00	ug/L
98-06-6	tert-Butylbenzene	21.4		0.17	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	21.2		0.18	1.00	ug/L
135-98-8	sec-Butylbenzene	21.2		0.17	1.00	ug/L
99-87-6	p-Isopropyltoluene	21.8		0.15	1.00	ug/L
104-51-8	n-Butylbenzene	21.0		0.22	1.00	ug/L
91-20-3	Naphthalene	20.6		0.59	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.6		74 - 125	111%	SPK: 50
1868-53-7	Dibromofluoromethane	54.7		75 - 124	109%	SPK: 50
2037-26-5	Toluene-d8	53.5		86 - 113	107%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.0		77 - 121	110%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	120000	5.544			
540-36-3	1,4-Difluorobenzene	216000	6.751			
3114-55-4	Chlorobenzene-d5	187000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	85600	12.018			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	VX1202WBSD01	SDG No.:	P5021
Lab Sample ID:	VX1202WBSD01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	DB-624UI	ID :	0.18
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044056.D	1		12/02/24 11:27	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: LIR001

Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG No.: P5021

Instrument ID: MSVOA_X Calibration Date(s): 11/21/2024 11/21/2024

Heated Purge: (Y/N) N Calibration Time(s): 09:47 12:03

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX043925.D		RRF005 = VX043926.D		RRF020 = VX043927.D			
		RRF050 = VX043928.D		RRF100 = VX043929.D		RRF150 = VX043930.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Methyl tert-butyl Ether	1.732	1.991	2.212	2.264	2.160	2.192	2.092	9.5	
Benzene	1.211	1.369	1.457	1.494	1.388	1.402	1.387	7	
Toluene	0.636	0.839	0.883	0.921	0.850	0.852	0.830	12	
Ethyl Benzene	1.579	1.831	1.910	2.024	1.884	1.938	1.861	8.2	
m/p-Xylenes	0.610	0.663	0.704	0.762	0.700	0.725	0.694	7.6	
o-Xylene	0.543	0.665	0.691	0.757	0.693	0.716	0.678	10.7	
Isopropylbenzene	3.435	3.927	3.996	4.097	3.679	3.927	3.843	6.3	
n-propylbenzene	3.667	4.115	4.469	4.733	4.286	4.486	4.293	8.6	
1,3,5-Trimethylbenzene	2.704	3.115	3.352	3.444	3.118	3.238	3.162	8.2	
tert-Butylbenzene	2.581	3.097	3.266	3.440	3.152	3.269	3.134	9.4	
1,2,4-Trimethylbenzene	2.557	3.173	3.366	3.440	3.124	3.228	3.148	9.9	
sec-Butylbenzene	3.116	3.921	3.968	4.249	3.812	3.974	3.840	10	
p-Isopropyltoluene	2.255	3.043	3.316	3.521	3.219	3.362	3.119	14.5	
n-Butylbenzene	1.897	2.580	2.805	3.184	2.909	3.126	2.750	17.2	
Naphthalene	2.605	3.361	3.698	4.005	3.733	4.058	3.576	15	
1,2-Dichloroethane-d4		0.926	0.876	0.751	0.855	0.880	0.858	7.6	
Dibromofluoromethane		0.371	0.353	0.327	0.359	0.376	0.357	5.4	
Toluene-d8		1.327	1.237	1.104	1.232	1.257	1.232	6.6	
4-Bromofluorobenzene		0.401	0.414	0.387	0.438	0.455	0.419	6.6	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: LIRO01

Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG No.: P5021

Instrument ID: MSVOA_X Calibration Date/Time: 11/27/2024 08:46

Lab File ID: VX044024.D Init. Calib. Date(s): 11/21/2024 11/21/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:47 12:03

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Methyl tert-butyl Ether	2.092	1.967		-5.97	20
Benzene	1.387	1.345		-3.03	20
Toluene	0.830	0.830		0	20
Ethyl Benzene	1.861	1.877		0.86	20
m/p-Xylenes	0.694	0.703		1.3	20
o-Xylene	0.678	0.683		0.74	20
Isopropylbenzene	3.843	3.898		1.43	20
n-propylbenzene	4.293	4.376		1.93	20
1,3,5-Trimethylbenzene	3.162	3.201		1.23	20
tert-Butylbenzene	3.134	3.187		1.69	20
1,2,4-Trimethylbenzene	3.148	3.189		1.3	20
sec-Butylbenzene	3.840	3.885		1.17	20
p-Isopropyltoluene	3.119	3.286		5.35	20
n-Butylbenzene	2.750	2.884		4.87	20
Naphthalene	3.576	3.607		0.87	20
1,2-Dichloroethane-d4	0.858	0.810		-5.59	20
Dibromofluoromethane	0.357	0.370		3.64	20
Toluene-d8	1.232	1.254		1.79	20
4-Bromofluorobenzene	0.419	0.419		0	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: LIRO01

Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG No.: P5021

Instrument ID: MSVOA_X Calibration Date/Time: 12/02/2024 09:45

Lab File ID: VX044052.D Init. Calib. Date(s): 11/21/2024 11/21/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:47 12:03

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Methyl tert-butyl Ether	2.092	2.203		5.31	20
Benzene	1.387	1.412		1.8	20
Toluene	0.830	0.852		2.65	20
Ethyl Benzene	1.861	1.956		5.11	20
m/p-Xylenes	0.694	0.726		4.61	20
o-Xylene	0.678	0.704		3.84	20
Isopropylbenzene	3.843	3.944		2.63	20
n-propylbenzene	4.293	4.440		3.42	20
1,3,5-Trimethylbenzene	3.162	3.291		4.08	20
tert-Butylbenzene	3.134	3.312		5.68	20
1,2,4-Trimethylbenzene	3.148	3.302		4.89	20
sec-Butylbenzene	3.840	4.041		5.23	20
p-Isopropyltoluene	3.119	3.375		8.21	20
n-Butylbenzene	2.750	3.049		10.87	20
Naphthalene	3.576	3.761		5.17	20
1,2-Dichloroethane-d4	0.858	0.890		3.73	20
Dibromofluoromethane	0.357	0.379		6.16	20
Toluene-d8	1.232	1.268		2.92	20
4-Bromofluorobenzene	0.419	0.442		5.49	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

LAB CHRONICLE

OrderID:	P5021	OrderDate:	11/27/2024 10:41:00 AM
Client:	LiRo Engineers, Inc.	Project:	RFK Bridge
Contact:	Martin Wesolowski	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P5021-01	MW-06	Water			11/26/24			11/27/24
			SVOCMS Group1	8270E		11/27/24	12/02/24	
P5021-02	MW-08	Water			11/26/24			11/27/24
			SVOCMS Group1	8270E		11/27/24	12/03/24	
P5021-03	MW-10	Water			11/26/24			11/27/24
			SVOCMS Group1	8270E		11/27/24	12/03/24	
P5021-04	MW-11	Water			11/26/24			11/27/24
			SVOCMS Group1	8270E		11/27/24	12/02/24	
P5021-05	MW-13	Water			11/26/24			11/27/24
			SVOCMS Group1	8270E		11/27/24	12/02/24	
P5021-06	MW-12	Water			11/26/24			11/27/24
			SVOCMS Group1	8270E		11/27/24	12/03/24	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: P5021
Client: LiRo Engineers, Inc.

Sample ID	Client ID		Parameter	Concentration	C	MDL	RDL	Units
Client ID : MW-06								
P5021-01	MW-06	WATER	Acenaphthene	11.800		0.83	5.1	ug/L
P5021-01	MW-06	WATER	Fluorene	5.200		0.98	5.1	ug/L
P5021-01	MW-06	WATER	Anthracene	2.700	J	1.1	5.1	ug/L
P5021-01	MW-06	WATER	Fluoranthene	8.900		1.3	5.1	ug/L
P5021-01	MW-06	WATER	Pyrene	6.800	Q	1.1	5.1	ug/L
Total Svoc :						35.40		
Total Concentration:						35.40		



SAMPLE DATA

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-06	SDG No.:	P5021
Lab Sample ID:	P5021-01	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	980 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140691.D	1	11/27/24 08:40	12/02/24 16:21	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
208-96-8	Acenaphthylene	5.10	UQ	1.10	5.10	ug/L
83-32-9	Acenaphthene	11.8		0.83	5.10	ug/L
86-73-7	Fluorene	5.20		0.98	5.10	ug/L
85-01-8	Phenanthrene	5.10	U	0.91	5.10	ug/L
120-12-7	Anthracene	2.70	J	1.10	5.10	ug/L
206-44-0	Fluoranthene	8.90		1.30	5.10	ug/L
129-00-0	Pyrene	6.80	Q	1.10	5.10	ug/L
56-55-3	Benzo(a)anthracene	5.10	UQ	0.96	5.10	ug/L
218-01-9	Chrysene	5.10	UQ	0.88	5.10	ug/L
205-99-2	Benzo(b)fluoranthene	5.10	U	1.20	5.10	ug/L
207-08-9	Benzo(k)fluoranthene	5.10	UQ	1.20	5.10	ug/L
50-32-8	Benzo(a)pyrene	5.10	U	1.70	5.10	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.10	U	1.00	5.10	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.10	U	1.20	5.10	ug/L
191-24-2	Benzo(g,h,i)perylene	5.10	U	1.20	5.10	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	89.8		49 - 133	90%	SPK: 100
321-60-8	2-Fluorobiphenyl	83.0		52 - 132	83%	SPK: 100
1718-51-0	Terphenyl-d14	80.1		48 - 125	80%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	83000		6.869		
1146-65-2	Naphthalene-d8	310000		8.151		
15067-26-2	Acenaphthene-d10	187000		9.904		
1517-22-2	Phenanthrene-d10	344000		11.398		
1719-03-5	Chrysene-d12	220000		14.045		
1520-96-3	Perylene-d12	187000		15.551		

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-06	SDG No.:	P5021
Lab Sample ID:	P5021-01	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	980	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3510C	Level :	LOW
		Final Vol:	1000
		Test:	SVOCMS Group1
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140691.D	1	11/27/24 08:40	12/02/24 16:21	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-08	SDG No.:	P5021
Lab Sample ID:	P5021-02	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140710.D	1	11/27/24 08:40	12/03/24 14:59	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
208-96-8	Acenaphthylene	5.10	UQ	1.10	5.10	ug/L
83-32-9	Acenaphthene	5.10	U	0.82	5.10	ug/L
86-73-7	Fluorene	5.10	U	0.97	5.10	ug/L
85-01-8	Phenanthrene	5.10	U	0.90	5.10	ug/L
120-12-7	Anthracene	5.10	U	1.10	5.10	ug/L
206-44-0	Fluoranthene	5.10	U	1.30	5.10	ug/L
129-00-0	Pyrene	5.10	UQ	1.10	5.10	ug/L
56-55-3	Benzo(a)anthracene	5.10	UQ	0.95	5.10	ug/L
218-01-9	Chrysene	5.10	UQ	0.87	5.10	ug/L
205-99-2	Benzo(b)fluoranthene	5.10	U	1.20	5.10	ug/L
207-08-9	Benzo(k)fluoranthene	5.10	UQ	1.20	5.10	ug/L
50-32-8	Benzo(a)pyrene	5.10	U	1.70	5.10	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.10	U	1.00	5.10	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.10	U	1.20	5.10	ug/L
191-24-2	Benzo(g,h,i)perylene	5.10	U	1.20	5.10	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	75.9		49 - 133	76%	SPK: 100
321-60-8	2-Fluorobiphenyl	76.6		52 - 132	77%	SPK: 100
1718-51-0	Terphenyl-d14	80.7		48 - 125	81%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	83400	6.869			
1146-65-2	Naphthalene-d8	299000	8.151			
15067-26-2	Acenaphthene-d10	175000	9.904			
1517-22-2	Phenanthrene-d10	305000	11.398			
1719-03-5	Chrysene-d12	169000	14.051			
1520-96-3	Perylene-d12	158000	15.563			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-08	SDG No.:	P5021
Lab Sample ID:	P5021-02	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140710.D	1	11/27/24 08:40	12/03/24 14:59	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-10	SDG No.:	P5021
Lab Sample ID:	P5021-03	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	980 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140711.D	1	11/27/24 08:40	12/03/24 15:24	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
208-96-8	Acenaphthylene	5.10	UQ	1.10	5.10	ug/L
83-32-9	Acenaphthene	5.10	U	0.83	5.10	ug/L
86-73-7	Fluorene	5.10	U	0.98	5.10	ug/L
85-01-8	Phenanthrene	5.10	U	0.91	5.10	ug/L
120-12-7	Anthracene	5.10	U	1.10	5.10	ug/L
206-44-0	Fluoranthene	5.10	U	1.30	5.10	ug/L
129-00-0	Pyrene	5.10	UQ	1.10	5.10	ug/L
56-55-3	Benzo(a)anthracene	5.10	UQ	0.96	5.10	ug/L
218-01-9	Chrysene	5.10	UQ	0.88	5.10	ug/L
205-99-2	Benzo(b)fluoranthene	5.10	U	1.20	5.10	ug/L
207-08-9	Benzo(k)fluoranthene	5.10	UQ	1.20	5.10	ug/L
50-32-8	Benzo(a)pyrene	5.10	U	1.70	5.10	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.10	U	1.00	5.10	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.10	U	1.20	5.10	ug/L
191-24-2	Benzo(g,h,i)perylene	5.10	U	1.20	5.10	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	91.8		49 - 133	92%	SPK: 100
321-60-8	2-Fluorobiphenyl	94.2		52 - 132	94%	SPK: 100
1718-51-0	Terphenyl-d14	93.7		48 - 125	94%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	71000	6.869			
1146-65-2	Naphthalene-d8	270000	8.151			
15067-26-2	Acenaphthene-d10	150000	9.904			
1517-22-2	Phenanthrene-d10	272000	11.398			
1719-03-5	Chrysene-d12	165000	14.051			
1520-96-3	Perylene-d12	152000	15.574			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-10	SDG No.:	P5021
Lab Sample ID:	P5021-03	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	980 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140711.D	1	11/27/24 08:40	12/03/24 15:24	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-11	SDG No.:	P5021
Lab Sample ID:	P5021-04	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140689.D	1	11/27/24 08:40	12/02/24 15:29	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
208-96-8	Acenaphthylene	5.10	UQ	1.10	5.10	ug/L
83-32-9	Acenaphthene	5.10	U	0.82	5.10	ug/L
86-73-7	Fluorene	5.10	U	0.97	5.10	ug/L
85-01-8	Phenanthrene	5.10	U	0.90	5.10	ug/L
120-12-7	Anthracene	5.10	U	1.10	5.10	ug/L
206-44-0	Fluoranthene	5.10	U	1.30	5.10	ug/L
129-00-0	Pyrene	5.10	UQ	1.10	5.10	ug/L
56-55-3	Benzo(a)anthracene	5.10	UQ	0.95	5.10	ug/L
218-01-9	Chrysene	5.10	UQ	0.87	5.10	ug/L
205-99-2	Benzo(b)fluoranthene	5.10	U	1.20	5.10	ug/L
207-08-9	Benzo(k)fluoranthene	5.10	UQ	1.20	5.10	ug/L
50-32-8	Benzo(a)pyrene	5.10	U	1.70	5.10	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.10	U	1.00	5.10	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.10	U	1.20	5.10	ug/L
191-24-2	Benzo(g,h,i)perylene	5.10	U	1.20	5.10	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	95.6		49 - 133	96%	SPK: 100
321-60-8	2-Fluorobiphenyl	94.4		52 - 132	94%	SPK: 100
1718-51-0	Terphenyl-d14	91.0		48 - 125	91%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	79000	6.869			
1146-65-2	Naphthalene-d8	302000	8.151			
15067-26-2	Acenaphthene-d10	171000	9.904			
1517-22-2	Phenanthrene-d10	322000	11.398			
1719-03-5	Chrysene-d12	192000	14.051			
1520-96-3	Perylene-d12	157000	15.574			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-11	SDG No.:	P5021
Lab Sample ID:	P5021-04	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140689.D	1	11/27/24 08:40	12/02/24 15:29	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-13	SDG No.:	P5021
Lab Sample ID:	P5021-05	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140688.D	1	11/27/24 08:40	12/02/24 15:03	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
208-96-8	Acenaphthylene	5.10	UQ	1.10	5.10	ug/L
83-32-9	Acenaphthene	5.10	U	0.82	5.10	ug/L
86-73-7	Fluorene	5.10	U	0.97	5.10	ug/L
85-01-8	Phenanthrene	5.10	U	0.90	5.10	ug/L
120-12-7	Anthracene	5.10	U	1.10	5.10	ug/L
206-44-0	Fluoranthene	5.10	U	1.30	5.10	ug/L
129-00-0	Pyrene	5.10	UQ	1.10	5.10	ug/L
56-55-3	Benzo(a)anthracene	5.10	UQ	0.95	5.10	ug/L
218-01-9	Chrysene	5.10	UQ	0.87	5.10	ug/L
205-99-2	Benzo(b)fluoranthene	5.10	U	1.20	5.10	ug/L
207-08-9	Benzo(k)fluoranthene	5.10	UQ	1.20	5.10	ug/L
50-32-8	Benzo(a)pyrene	5.10	U	1.70	5.10	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.10	U	1.00	5.10	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.10	U	1.20	5.10	ug/L
191-24-2	Benzo(g,h,i)perylene	5.10	U	1.20	5.10	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	96.1		49 - 133	96%	SPK: 100
321-60-8	2-Fluorobiphenyl	94.0		52 - 132	94%	SPK: 100
1718-51-0	Terphenyl-d14	90.8		48 - 125	91%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	82100	6.869			
1146-65-2	Naphthalene-d8	307000	8.151			
15067-26-2	Acenaphthene-d10	175000	9.904			
1517-22-2	Phenanthrene-d10	319000	11.398			
1719-03-5	Chrysene-d12	197000	14.051			
1520-96-3	Perylene-d12	158000	15.568			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-13	SDG No.:	P5021
Lab Sample ID:	P5021-05	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140688.D	1	11/27/24 08:40	12/02/24 15:03	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-12	SDG No.:	P5021
Lab Sample ID:	P5021-06	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	980 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140712.D	1	11/27/24 08:40	12/03/24 15:50	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
208-96-8	Acenaphthylene	5.10	UQ	1.10	5.10	ug/L
83-32-9	Acenaphthene	5.10	U	0.83	5.10	ug/L
86-73-7	Fluorene	5.10	U	0.98	5.10	ug/L
85-01-8	Phenanthrene	5.10	U	0.91	5.10	ug/L
120-12-7	Anthracene	5.10	U	1.10	5.10	ug/L
206-44-0	Fluoranthene	5.10	U	1.30	5.10	ug/L
129-00-0	Pyrene	5.10	UQ	1.10	5.10	ug/L
56-55-3	Benzo(a)anthracene	5.10	UQ	0.96	5.10	ug/L
218-01-9	Chrysene	5.10	UQ	0.88	5.10	ug/L
205-99-2	Benzo(b)fluoranthene	5.10	U	1.20	5.10	ug/L
207-08-9	Benzo(k)fluoranthene	5.10	UQ	1.20	5.10	ug/L
50-32-8	Benzo(a)pyrene	5.10	U	1.70	5.10	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.10	U	1.00	5.10	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.10	U	1.20	5.10	ug/L
191-24-2	Benzo(g,h,i)perylene	5.10	U	1.20	5.10	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	93.5		49 - 133	93%	SPK: 100
321-60-8	2-Fluorobiphenyl	88.4		52 - 132	88%	SPK: 100
1718-51-0	Terphenyl-d14	97.2		48 - 125	97%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	67800	6.869			
1146-65-2	Naphthalene-d8	251000	8.151			
15067-26-2	Acenaphthene-d10	156000	9.904			
1517-22-2	Phenanthrene-d10	260000	11.398			
1719-03-5	Chrysene-d12	154000	14.051			
1520-96-3	Perylene-d12	131000	15.563			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	11/26/24
Project:	RFK Bridge	Date Received:	11/27/24
Client Sample ID:	MW-12	SDG No.:	P5021
Lab Sample ID:	P5021-06	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	980 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140712.D	1	11/27/24 08:40	12/03/24 15:50	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: P5021

Client: LiRo Engineers, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P5021-01	MW-06	Nitrobenzene-d5	100	89.8	90		49	133
		2-Fluorobiphenyl	100	83.0	83		52	132
		Terphenyl-d14	100	80.1	80		48	125
P5021-02	MW-08	Nitrobenzene-d5	100	75.9	76		49	133
		2-Fluorobiphenyl	100	76.6	77		52	132
		Terphenyl-d14	100	80.7	81		48	125
P5021-03	MW-10	Nitrobenzene-d5	100	91.8	92		49	133
		2-Fluorobiphenyl	100	94.2	94		52	132
		Terphenyl-d14	100	93.7	94		48	125
P5021-04	MW-11	Nitrobenzene-d5	100	95.6	96		49	133
		2-Fluorobiphenyl	100	94.4	94		52	132
		Terphenyl-d14	100	91.0	91		48	125
P5021-05	MW-13	Nitrobenzene-d5	100	96.1	96		49	133
		2-Fluorobiphenyl	100	94.0	94		52	132
		Terphenyl-d14	100	90.8	91		48	125
P5021-06	MW-12	Nitrobenzene-d5	100	93.5	93		49	133
		2-Fluorobiphenyl	100	88.4	88		52	132
		Terphenyl-d14	100	97.2	97		48	125
PB165296BL	PB165296BL	Nitrobenzene-d5	100	94.5	94		49	133
		2-Fluorobiphenyl	100	96.0	96		52	132
		Terphenyl-d14	100	94.7	95		48	125
PB165296BS	PB165296BS	Nitrobenzene-d5	100	86.7	87		49	133
		2-Fluorobiphenyl	100	87.6	88		52	132
		Terphenyl-d14	100	87.5	87		48	125
PB165296BSD	PB165296BSD	Nitrobenzene-d5	100	96.7	97		49	133
		2-Fluorobiphenyl	100	95.4	95		52	132
		Terphenyl-d14	100	101	101		48	125

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5021

Client: LiRo Engineers, Inc.

Analytical Method: 8270E DataFile: BF140683.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB165296BS	Acenaphthylene	50	49.7	ug/L	99				79	103	
	Acenaphthene	50	47.6	ug/L	95				59	113	
	Fluorene	50	47.3	ug/L	95				64	107	
	Phenanthrene	50	47.6	ug/L	95				62	109	
	Anthracene	50	49.7	ug/L	99				65	110	
	Fluoranthene	50	50.5	ug/L	101				64	110	
	Pyrene	50	46.0	ug/L	92				71	103	
	Benzo(a)anthracene	50	48.2	ug/L	96				62	107	
	Chrysene	50	48.5	ug/L	97				61	108	
	Benzo(b)fluoranthene	50	47.2	ug/L	94				77	113	
	Benzo(k)fluoranthene	50	48.9	ug/L	98				77	105	
	Benzo(a)pyrene	50	52.4	ug/L	105				72	131	
	Indeno(1,2,3-cd)pyrene	50	47.6	ug/L	95				72	105	
	Dibenz(a,h)anthracene	50	47.4	ug/L	95				78	115	
	Benzo(g,h,i)perylene	50	43.9	ug/L	88				75	118	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5021

Client: LiRo Engineers, Inc.

Analytical Method: 8270E DataFile: BF140684.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB165296BSD	Acenaphthylene	50	54.9	ug/L	110	10	*		79	103	20
	Acenaphthene	50	52.4	ug/L	105	10			59	113	20
	Fluorene	50	52.2	ug/L	104	10			64	107	20
	Phenanthrene	50	53.2	ug/L	106	11			62	109	20
	Anthracene	50	54.6	ug/L	109	9			65	110	20
	Fluoranthene	50	46.8	ug/L	94	8			64	110	20
	Pyrene	50	52.5	ug/L	105	13	*		71	103	20
	Benzo(a)anthracene	50	56.0	ug/L	112	15	*		62	107	20
	Chrysene	50	54.9	ug/L	110	12	*		61	108	20
	Benzo(b)fluoranthene	50	55.5	ug/L	111	16			77	113	20
	Benzo(k)fluoranthene	50	54.0	ug/L	108	10	*		77	105	20
	Benzo(a)pyrene	50	62.5	ug/L	125	18			72	131	20
	Indeno(1,2,3-cd)pyrene	50	49.9	ug/L	100	5			72	105	20
	Dibenz(a,h)anthracene	50	49.7	ug/L	99	5			78	115	20
	Benzo(g,h,i)perylene	50	49.2	ug/L	98	11			75	118	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB165296BL

Lab Name: CHEMTECH

Contract: LIRO01

Lab Code: CHEM Case No.: P5021

SAS No.: P5021 SDG NO.: P5021

Lab File ID: BF140682.D

Lab Sample ID: PB165296BL

Instrument ID: BNA_F

Date Extracted: 11/27/2024

Matrix: (soil/water) Water

Date Analyzed: 12/02/2024

Level: (low/med) LOW

Time Analyzed: 10:57

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB165296BS	PB165296BS	BF140683.D	12/02/2024
PB165296BSD	PB165296BSD	BF140684.D	12/02/2024
MW-13	P5021-05	BF140688.D	12/02/2024
MW-08	P5021-02	BF140710.D	12/03/2024
MW-10	P5021-03	BF140711.D	12/03/2024
MW-12	P5021-06	BF140712.D	12/03/2024
MW-06	P5021-01	BF140691.D	12/02/2024
MW-11	P5021-04	BF140689.D	12/02/2024

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: LIRO01

Lab Code: CHEM

SAS No.: P5021

SDG NO.: P5021

Lab File ID: BF140526.D

DFTPP Injection Date: 11/21/2024

Instrument ID: BNA_F

DFTPP Injection Time: 10:17

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	33.3
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	35.5
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	48
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	29.2
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	15.7
442	Greater than 50% of mass 198	99.8
443	15.0 - 24.0% of mass 442	18.1 (18.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF140528.D	11/21/2024	11:13
SSTDICC005	SSTDICC005	BF140529.D	11/21/2024	11:39
SSTDICC010	SSTDICC010	BF140530.D	11/21/2024	12:05
SSTDICC020	SSTDICC020	BF140531.D	11/21/2024	12:32
SSTDICCC040	SSTDICCC040	BF140532.D	11/21/2024	12:58
SSTDICC050	SSTDICC050	BF140533.D	11/21/2024	13:25
SSTDICC060	SSTDICC060	BF140534.D	11/21/2024	13:51
SSTDICC080	SSTDICC080	BF140535.D	11/21/2024	14:18

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: LIRO01

Lab Code: CHEM

SAS No.: P5021 SDG NO.: P5021

Lab File ID: BF140680.D

DFTPP Injection Date: 12/02/2024

Instrument ID: BNA_F

DFTPP Injection Time: 09:05

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.7
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	35
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	45.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
275	10.0 - 60.0% of mass 198	29.2
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	15.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.4 (18.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140681.D	12/02/2024	10:22
PB165296BL	PB165296BL	BF140682.D	12/02/2024	10:57
PB165296BS	PB165296BS	BF140683.D	12/02/2024	11:23
PB165296BSD	PB165296BSD	BF140684.D	12/02/2024	11:49
MW-13	P5021-05	BF140688.D	12/02/2024	15:03
MW-11	P5021-04	BF140689.D	12/02/2024	15:29
MW-06	P5021-01	BF140691.D	12/02/2024	16:21

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: LIRO01

Lab Code: CHEM

SAS No.: P5021

SDG NO.: P5021

Lab File ID: BF140701.D

DFTPP Injection Date: 12/03/2024

Instrument ID: BNA_F

DFTPP Injection Time: 10:24

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.9
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	35
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	46.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	28.2
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.7 (18.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140702.D	12/03/2024	11:24
MW-08	P5021-02	BF140710.D	12/03/2024	14:59
MW-10	P5021-03	BF140711.D	12/03/2024	15:24
MW-12	P5021-06	BF140712.D	12/03/2024	15:50

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021

EPA Sample No.: SSTDCCC040 Date Analyzed: 12/02/2024

Lab File ID: BF140681.D Time Analyzed: 10:22

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	83338	6.869	303615	8.15	170338	9.91
UPPER LIMIT	166676	7.369	607230	8.651	340676	10.41
LOWER LIMIT	41669	6.369	151808	7.651	85169	9.41
EPA SAMPLE NO.						
01 PB165296BL	73230	6.87	279727	8.15	158263	9.90
02 PB165296BS	84060	6.87	323501	8.15	181370	9.91
03 PB165296BSD	80604	6.87	308077	8.15	176945	9.91
04 MW-06	82990	6.87	310195	8.15	187284	9.90
05 MW-11	78955	6.87	302257	8.15	171220	9.90
06 MW-13	82130	6.87	307028	8.15	174760	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021

EPA Sample No.: SSTDCCC040 Date Analyzed: 12/02/2024

Lab File ID: BF140681.D Time Analyzed: 10:22

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	321182	11.398	183379	14.063	159952	15.58
UPPER LIMIT	642364	11.898	366758	14.563	319904	16.08
LOWER LIMIT	160591	10.898	91689.5	13.563	79976	15.08
EPA SAMPLE NO.						
01 PB165296BL	307927	11.40	198812	14.07	157266	15.60
02 PB165296BS	351850	11.40	225717	14.06	182835	15.58
03 PB165296BSD	364189	11.40	185704	14.06	166131	15.57
04 MW-06	343851	11.40	220062	14.05	186765	15.55
05 MW-11	321567	11.40	192319	14.05	156519	15.57
06 MW-13	319066	11.40	197195	14.05	158239	15.57

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021

EPA Sample No.: SSTDCCC040 Date Analyzed: 12/03/2024

Lab File ID: BF140702.D Time Analyzed: 11:24

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	78577	6.869	287976	8.15	161859	9.91
UPPER LIMIT	157154	7.369	575952	8.651	323718	10.41
LOWER LIMIT	39288.5	6.369	143988	7.651	80929.5	9.41
EPA SAMPLE NO.						
01 MW-08	83408	6.87	298688	8.15	174661	9.90
02 MW-10	70972	6.87	269570	8.15	149746	9.90
03 MW-12	67828	6.87	251382	8.15	155575	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021

EPA Sample No.: SSTDCCC040 Date Analyzed: 12/03/2024

Lab File ID: BF140702.D Time Analyzed: 11:24

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	309121	11.398	178961	14.057	151018	15.568
UPPER LIMIT	618242	11.898	357922	14.557	302036	16.068
LOWER LIMIT	154561	10.898	89480.5	13.557	75509	15.068
EPA SAMPLE NO.						
01 MW-08	305249	11.40	169161	14.05	157526	15.56
02 MW-10	271510	11.40	165452	14.05	151648	15.57
03 MW-12	260123	11.40	154428	14.05	130761	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	PB165296BL	SDG No.:	P5021
Lab Sample ID:	PB165296BL	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140682.D	1	11/27/24 08:40	12/02/24 10:57	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
208-96-8	Acenaphthylene	5.00	U	1.00	5.00	ug/L
83-32-9	Acenaphthene	5.00	U	0.81	5.00	ug/L
86-73-7	Fluorene	5.00	U	0.96	5.00	ug/L
85-01-8	Phenanthrene	5.00	U	0.89	5.00	ug/L
120-12-7	Anthracene	5.00	U	1.10	5.00	ug/L
206-44-0	Fluoranthene	5.00	U	1.30	5.00	ug/L
129-00-0	Pyrene	5.00	U	1.10	5.00	ug/L
56-55-3	Benzo(a)anthracene	5.00	U	0.94	5.00	ug/L
218-01-9	Chrysene	5.00	U	0.86	5.00	ug/L
205-99-2	Benzo(b)fluoranthene	5.00	U	1.10	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	5.00	U	1.20	5.00	ug/L
50-32-8	Benzo(a)pyrene	5.00	U	1.70	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.00	U	1.00	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.00	U	1.20	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	5.00	U	1.20	5.00	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	94.5		49 - 133	94%	SPK: 100
321-60-8	2-Fluorobiphenyl	96.0		52 - 132	96%	SPK: 100
1718-51-0	Terphenyl-d14	94.7		48 - 125	95%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	73200	6.869			
1146-65-2	Naphthalene-d8	280000	8.151			
15067-26-2	Acenaphthene-d10	158000	9.904			
1517-22-2	Phenanthrene-d10	308000	11.398			
1719-03-5	Chrysene-d12	199000	14.068			
1520-96-3	Perylene-d12	157000	15.598			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	PB165296BL	SDG No.:	P5021
Lab Sample ID:	PB165296BL	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140682.D	1	11/27/24 08:40	12/02/24 10:57	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	PB165296BS	SDG No.:	P5021
Lab Sample ID:	PB165296BS	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140683.D	1	11/27/24 08:40	12/02/24 11:23	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
208-96-8	Acenaphthylene	49.7		1.00	5.00	ug/L
83-32-9	Acenaphthene	47.6		0.81	5.00	ug/L
86-73-7	Fluorene	47.3		0.96	5.00	ug/L
85-01-8	Phenanthrene	47.6		0.89	5.00	ug/L
120-12-7	Anthracene	49.7		1.10	5.00	ug/L
206-44-0	Fluoranthene	50.5		1.30	5.00	ug/L
129-00-0	Pyrene	46.0		1.10	5.00	ug/L
56-55-3	Benzo(a)anthracene	48.2		0.94	5.00	ug/L
218-01-9	Chrysene	48.5		0.86	5.00	ug/L
205-99-2	Benzo(b)fluoranthene	47.2		1.10	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	48.9		1.20	5.00	ug/L
50-32-8	Benzo(a)pyrene	52.4		1.70	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	47.6		1.00	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	47.4		1.20	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	43.9		1.20	5.00	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	86.7		49 - 133	87%	SPK: 100
321-60-8	2-Fluorobiphenyl	87.6		52 - 132	88%	SPK: 100
1718-51-0	Terphenyl-d14	87.5		48 - 125	87%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	84100	6.869			
1146-65-2	Naphthalene-d8	324000	8.151			
15067-26-2	Acenaphthene-d10	181000	9.91			
1517-22-2	Phenanthrene-d10	352000	11.398			
1719-03-5	Chrysene-d12	226000	14.063			
1520-96-3	Perylene-d12	183000	15.58			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	PB165296BS	SDG No.:	P5021
Lab Sample ID:	PB165296BS	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140683.D	1	11/27/24 08:40	12/02/24 11:23	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	PB165296BSD	SDG No.:	P5021
Lab Sample ID:	PB165296BSD	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140684.D	1	11/27/24 08:40	12/02/24 11:49	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
208-96-8	Acenaphthylene	54.9		1.00	5.00	ug/L
83-32-9	Acenaphthene	52.4		0.81	5.00	ug/L
86-73-7	Fluorene	52.2		0.96	5.00	ug/L
85-01-8	Phenanthrene	53.2		0.89	5.00	ug/L
120-12-7	Anthracene	54.6		1.10	5.00	ug/L
206-44-0	Fluoranthene	46.8		1.30	5.00	ug/L
129-00-0	Pyrene	52.5		1.10	5.00	ug/L
56-55-3	Benzo(a)anthracene	56.0		0.94	5.00	ug/L
218-01-9	Chrysene	54.9		0.86	5.00	ug/L
205-99-2	Benzo(b)fluoranthene	55.5		1.10	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	54.0		1.20	5.00	ug/L
50-32-8	Benzo(a)pyrene	62.5		1.70	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	49.9		1.00	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	49.7		1.20	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	49.2		1.20	5.00	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	96.7		49 - 133	97%	SPK: 100
321-60-8	2-Fluorobiphenyl	95.4		52 - 132	95%	SPK: 100
1718-51-0	Terphenyl-d14	101		48 - 125	101%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	80600	6.869			
1146-65-2	Naphthalene-d8	308000	8.151			
15067-26-2	Acenaphthene-d10	177000	9.91			
1517-22-2	Phenanthrene-d10	364000	11.398			
1719-03-5	Chrysene-d12	186000	14.057			
1520-96-3	Perylene-d12	166000	15.574			

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	PB165296BSD	SDG No.:	P5021
Lab Sample ID:	PB165296BSD	Matrix:	Water
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF140684.D	1	11/27/24 08:40	12/02/24 11:49	PB165296

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF112124.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Nov 21 15:23:48 2024
 Response Via : Initial Calibration

Calibration Files

2.5 =BF140528.D 5 =BF140529.D 10 =BF140530.D 20 =BF140531.D 40 =BF140532.D 50 =BF140533.D 60 =BF140534.D 80 =BF140535.D

Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I 1,4-Dichlorobenzen...	-----ISTD-----									
2) 1,4-Dioxane		0.501	0.502	0.576	0.489	0.452	0.454	0.477	0.493	8.46
3) Pyridine		0.981	1.019	1.121	1.192	1.062	1.087	1.128	1.084	6.54
4) n-Nitrosodimet...		0.616	0.611	0.661	0.648	0.632	0.632	0.660	0.637	3.15
5) S 2-Fluorophenol		1.261	1.233	1.202	1.174	1.117	1.127	1.091	1.172	5.40
6) Aniline		1.159	1.195	1.133	1.251	1.042	1.020	0.835	1.091	12.73
7) S Phenol-d6		1.729	1.617	1.559	1.602	1.465	1.470	1.407	1.550	7.14
8) 2-Chlorophenol		1.385	1.328	1.278	1.283	1.201	1.208	1.145	1.261	6.51
9) Benzaldehyde			1.007	0.970	0.738	0.752	0.635		0.820	19.55
10) C Phenol		1.723	1.648	1.620	1.667	1.514	1.490	1.417	1.583	6.98
11) bis(2-Chloroet...		1.292	1.242	1.214	1.246	1.146	1.200	1.138	1.211	4.56
12) 1,3-Dichlorobe...		1.600	1.493	1.433	1.399	1.345	1.361	1.288	1.417	7.31
13) C 1,4-Dichlorobe...		1.613	1.501	1.456	1.428	1.358	1.372	1.314	1.435	7.04
14) 1,2-Dichlorobe...		1.503	1.434	1.375	1.355	1.268	1.266	1.210	1.344	7.71
15) Benzyl Alcohol		1.234	1.174	1.161	1.224	1.113	1.090	1.048	1.149	5.98
16) 2,2'-oxybis(1-...		1.650	1.480	1.434	1.463	1.335	1.377	1.276	1.431	8.43
17) 2-Methylphenol		1.121	1.034	1.033	1.042	0.956	0.959	0.922	1.009	6.74
18) Hexachloroethane		0.598	0.545	0.549	0.537	0.514	0.510	0.499	0.536	6.17
19) P n-Nitroso-di-n...	0.972	1.002	0.949	0.916	0.946	0.856	0.857	0.829	0.916	6.82
20) 3+4-Methylphenols		1.495	1.362	1.305	1.347	1.211	1.209	1.154	1.298	8.98

21) I Naphthalene-d8	-----ISTD-----									
22) Acetophenone		0.536	0.511	0.494	0.481	0.463	0.460	0.468	0.488	5.75
23) S Nitrobenzene-d5		0.409	0.403	0.395	0.392	0.377	0.377	0.383	0.391	3.21
24) Nitrobenzene		0.439	0.409	0.409	0.401	0.388	0.391	0.392	0.404	4.35
25) Isophorone		0.694	0.654	0.657	0.662	0.629	0.634	0.635	0.652	3.44
26) C 2-Nitrophenol		0.178	0.172	0.185	0.180	0.178	0.180	0.180	0.179	2.17
27) 2,4-Dimethylph...		0.221	0.214	0.213	0.224	0.204	0.207	0.218	0.214	3.40
28) bis(2-Chloroet...		0.428	0.408	0.403	0.397	0.378	0.384	0.383	0.397	4.37
29) C 2,4-Dichloroph...		0.301	0.291	0.290	0.282	0.277	0.274	0.271	0.284	3.82
30) 1,2,4-Trichlor...		0.342	0.338	0.332	0.317	0.317	0.312	0.312	0.324	3.91
31) Naphthalene		1.116	1.075	1.062	1.013	0.993	0.983	0.970	1.030	5.32
32) Benzoic acid			0.101	0.126	0.177	0.185	0.192	0.202	0.164	24.72
33) 4-Chloroaniline		0.308	0.318	0.309	0.325	0.303	0.308	0.289	0.308	3.71
34) C Hexachlorobuta...		0.227	0.225	0.219	0.212	0.209	0.207	0.204	0.215	4.15
35) Caprolactam		0.091	0.091	0.091	0.090	0.085	0.085	0.083	0.088	3.99
36) C 4-Chloro-3-met...		0.348	0.318	0.317	0.327	0.307	0.307	0.301	0.318	4.95
37) 2-Methylnaphth...		0.724	0.680	0.669	0.650	0.623	0.620	0.615	0.654	6.09
38) 1-Methylnaphth...		0.707	0.665	0.659	0.638	0.611	0.608	0.601	0.641	5.98

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF112124.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.635	0.604	0.606	0.563	0.561	0.565	0.565	0.586	5.02	
41) P	Hexachlorocycl...		0.055	0.090	0.114	0.121	0.129	0.133	0.107	27.80	
42) S	2,4,6-Tribromo...	0.222	0.209	0.220	0.216	0.210	0.210	0.211	0.214	2.52	
43) C	2,4,6-Trichlor...	0.384	0.358	0.375	0.366	0.364	0.360	0.365	0.367	2.45	
44)	2,4,5-Trichlor...	0.402	0.396	0.410	0.404	0.390	0.397	0.392	0.399	1.80	
45) S	2-Fluorobiphenyl	1.550	1.402	1.423	1.291	1.255	1.240	1.235	1.342	8.90	
46)	1,1'-Biphenyl	1.666	1.530	1.563	1.447	1.427	1.418	1.403	1.493	6.50	
47)	2-Chloronaphth...	1.251	1.145	1.162	1.106	1.082	1.084	1.091	1.131	5.39	
48)	2-Nitroaniline	0.367	0.351	0.379	0.367	0.363	0.357	0.359	0.363	2.50	
49)	Acenaphthylene	1.890	1.765	1.808	1.662	1.628	1.623	1.590	1.710	6.58	
50)	Dimethylphthalate	1.455	1.329	1.346	1.299	1.267	1.270	1.254	1.317	5.28	
51)	2,6-Dinitrotol...	0.314	0.299	0.307	0.301	0.291	0.293	0.286	0.299	3.24	
52) C	Acenaphthene	1.182	1.110	1.134	1.063	1.052	1.037	1.026	1.086	5.29	
53)	3-Nitroaniline	0.307	0.297	0.309	0.300	0.289	0.281	0.262	0.292	5.71	
54) P	2,4-Dinitrophenol		0.057	0.089	0.140	0.145	0.150	0.154	0.122	32.70	
55)	Dibenzofuran	1.898	1.739	1.739	1.622	1.559	1.531	1.509	1.657	8.53	
56) P	4-Nitrophenol		0.160	0.195	0.207	0.212	0.214	0.208	0.199	10.31	
57)	2,4-Dinitrotol...	0.403	0.403	0.416	0.404	0.386	0.389	0.379	0.397	3.24	
58)	Fluorene	1.509	1.409	1.399	1.295	1.263	1.224	1.210	1.330	8.37	
59)	2,3,4,6-Tetrac...	0.307	0.296	0.308	0.312	0.305	0.305	0.312	0.306	1.72	
60)	Diethylphthalate	1.495	1.375	1.393	1.311	1.292	1.268	1.234	1.338	6.66	
61)	4-Chlorophenyl...	0.739	0.682	0.685	0.639	0.619	0.605	0.599	0.653	7.90	
62)	4-Nitroaniline	0.307	0.301	0.315	0.312	0.310	0.309	0.291	0.306	2.67	
63)	Azobenzene	1.406	1.320	1.320	1.235	1.205	1.206	1.179	1.267	6.55	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...		0.073	0.090	0.110	0.110	0.117	0.113	0.102	16.74	
66) c	n-Nitrosodiphe...	0.632	0.618	0.597	0.578	0.577	0.580	0.558	0.591	4.39	
67)	4-Bromophenyl-...	0.218	0.215	0.206	0.200	0.200	0.205	0.198	0.206	3.79	
68)	Hexachlorobenzene	0.257	0.240	0.239	0.233	0.232	0.236	0.232	0.238	3.65	
69)	Atrazine	0.181	0.171	0.131	0.140	0.147	0.196	0.198	0.166	16.37	
70) C	Pentachlorophenol		0.071	0.090	0.116	0.115	0.121	0.118	0.105	19.24	
71)	Phenanthrene	1.074	1.020	0.970	0.944	0.925	0.916	0.881	0.961	6.87	
72)	Anthracene	1.038	0.994	0.958	0.925	0.905	0.904	0.859	0.940	6.47	
73)	Carbazole	1.004	0.949	0.930	0.889	0.876	0.862	0.821	0.905	6.75	
74)	Di-n-butylphth...	1.144	1.075	1.074	1.023	1.021	1.007	0.967	1.044	5.56	
75) C	Fluoranthene	1.186	1.111	1.114	1.014	0.999	0.962	0.921	1.044	9.11	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine	0.268	0.422	0.296	0.575	0.734	1.025	0.751	0.581	47.31	
78)	Pyrene	1.905	1.801	1.897	1.853	1.791	1.898	1.799	1.849	2.79	
79) S	Terphenyl-d14	1.351	1.254	1.308	1.283	1.228	1.313	1.254	1.284	3.31	
80)	Butylbenzylphth...	0.676	0.644	0.694	0.679	0.655	0.674	0.641	0.666	2.96	
81)	Benzo(a)anthra...	1.409	1.319	1.379	1.286	1.295	1.338	1.242	1.324	4.30	
82)	3,3'-Dichlorob...	0.375	0.383	0.393	0.413	0.406	0.419	0.394	0.397	4.03	
83)	Chrysene	1.354	1.263	1.218	1.201	1.137	1.173	1.128	1.211	6.50	
84)	Bis(2-ethylhex...	0.904	0.828	0.862	0.833	0.824	0.841	0.797	0.841	4.00	
85) c	Di-n-octyl pht...	1.198	1.133	1.147	1.132	1.147	1.169	1.121	1.150	2.29	


Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF112124.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.209	1.283	1.299	1.304	1.306	1.409	1.313	1.303	4.51
88)	Benzo(b)fluora...	1.347	1.256	1.380	1.186	1.248	1.207	1.168	1.256	6.41
89)	Benzo(k)fluora...	1.243	1.236	1.059	1.101	1.022	1.055	0.980	1.099	9.34
90) C	Benzo(a)pyrene	1.062	1.050	1.063	1.007	0.998	1.014	0.957	1.021	3.81
91)	Dibenzo(a,h)an...	1.015	1.038	1.063	1.068	1.075	1.149	1.064	1.067	3.90
92)	Benzo(g,h,i)pe...	1.033	1.090	1.078	1.085	1.094	1.161	1.083	1.089	3.48

(#) = Out of Range

Benzoic acid


Response Ratio



Response = 2.193e-001 * Amt - 7.887e-002
 Coef of Det (r^2) = 0.997922 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\86 of 105
 Calibration Table Last Updated: Thu Nov 21 15:23:48 2024

Hexachlorocyclopentadiene


Response Ratio



Response = 1.448e-001 * Amt - 5.280e-002
Coef of Det (r^2) = 0.998763 Curve Fit: Linear
MetP5021Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\897 of 10512124.M
Calibration Table Last Updated: Thu Nov 21 15:23:48 2024

2,4-Dinitrophenol

Response Ratio



Response = 1.715e-001 * Amt - 6.708e-002
Coef of Det (r^2) = 0.998475 Curve Fit: Linear
MetP5021Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8970112124.M
Calibration Table Last Updated: Thu Nov 21 15:23:48 2024

6

A

B

C

D

E

F

G

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: LIRO01

Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG No.: P5021

Instrument ID: BNA_F Calibration Date/Time: 12/02/2024 10:22

Lab File ID: BF140681.D Init. Calib. Date(s): 11/21/2024 11/21/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:13 14:18

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.172	1.148		-2.0	
Phenol-d6	1.550	1.511		-2.5	
Nitrobenzene-d5	0.391	0.381		-2.6	
2-Fluorobiphenyl	1.342	1.303		-2.9	
Acenaphthylene	1.710	1.674		-2.1	
Acenaphthene	1.086	1.073		-1.2	20.0
Fluorene	1.330	1.312		-1.4	
2,4,6-Tribromophenol	0.214	0.211		-1.4	
Phenanthrene	0.961	0.947		-1.5	
Anthracene	0.940	0.935		-0.5	
Fluoranthene	1.044	1.053		0.9	20.0
Pyrene	1.849	1.875		1.4	
Terphenyl-d14	1.284	1.262		-1.7	
Benzo (a) anthracene	1.324	1.320		-0.3	
Chrysene	1.211	1.200		-0.9	
Benzo (b) fluoranthene	1.256	1.244		-1.0	
Benzo (k) fluoranthene	1.099	1.048		-4.6	
Benzo (a) pyrene	1.021	1.000		-2.1	20.0
Indeno (1,2,3-cd) pyrene	1.303	1.265		-2.9	
Dibenzo (a,h) anthracene	1.067	1.025		-3.9	
Benzo (g,h,i) perylene	1.089	1.064		-2.3	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: LIRO01

Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG No.: P5021

Instrument ID: BNA_F Calibration Date/Time: 12/03/2024 11:24

Lab File ID: BF140702.D Init. Calib. Date(s): 11/21/2024 11/21/2024

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:13 14:18

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.172	1.135		-3.2	
Phenol-d6	1.550	1.481		-4.5	
Nitrobenzene-d5	0.391	0.372		-4.9	
2-Fluorobiphenyl	1.342	1.315		-2.0	
Acenaphthylene	1.710	1.655		-3.2	
Acenaphthene	1.086	1.031		-5.1	20.0
Fluorene	1.330	1.306		-1.8	
2,4,6-Tribromophenol	0.214	0.209		-2.3	
Phenanthrene	0.961	0.921		-4.2	
Anthracene	0.940	0.906		-3.6	
Fluoranthene	1.044	1.025		-1.8	20.0
Pyrene	1.849	1.786		-3.4	
Terphenyl-d14	1.284	1.295		0.9	
Benzo (a) anthracene	1.324	1.330		0.5	
Chrysene	1.211	1.133		-6.4	
Benzo (b) fluoranthene	1.256	1.178		-6.2	
Benzo (k) fluoranthene	1.099	1.085		-1.3	
Benzo (a) pyrene	1.021	0.989		-3.1	20.0
Indeno (1,2,3-cd) pyrene	1.303	1.389		6.6	
Dibenzo (a,h) anthracene	1.067	1.149		7.7	
Benzo (g,h,i) perylene	1.089	1.195		9.7	

All other compounds must meet a minimum RRF of 0.010.



SHIPPING DOCUMENTS

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 78-8922

www.chemtech.net

Chemtech Project Number:

P5021

COC Number:

CLIENT INFORMATION

COMPANY: Liro Environmental Inc.
ADDRESS: 690 Delaware Ave
CITY: Buffalo STATE: NY ZIP: 14209
ATTENTION: Martin Wesolowski
PHONE: 716.970.4273 FAX: 716.882.9640

PROJECT INFORMATION

PROJECT NAME: RMB, RFK Bridge Randall's Island
PROJECT #: RMB-2023 LOCATION: NY
PROJECT MANAGER: Martin Wesolowski
E-MAIL: Wesolowskim@liro.com
PHONE: 716-970-9640 FAX: 716-882-9640

BILLING INFORMATION

BILL TO: Same PO#
ADDRESS: Same
CITY: Same STATE: ZIP:
ATTENTION: Same PHONE:

DATA TURNAROUND INFORMATION

FAX: _____ DAYS*
HARD COPY: standard DAYS*
EDD _____ DAYS*
* TO BE APPROVED BY CHEMTECH
STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ RESULTS ONLY ☐ USEPA CLP
☐ RESULTS + QC ☐ New York State ASP "B"
☐ New Jersey REDUCED ☐ New York State ASP "A"
☐ New Jersey CLP ☐ Other _____
☐ EDD Format

ANALYSIS

CP-51 VOCs	CP-51 SVOCs	Tables 2 & 3							
1	2	3	4	5	6	7	8	9	

PRESERVATIVES

COMMENTS

<- Specify Preservatives
A-HCl B-HNO3
C-H2SO4 D-NaOH
E-ICE F-Other

CHEMTECH
SAMPLE
IDPROJECT
SAMPLE IDENTIFICATIONSAMPLE
MATRIXSAMPLE
TYPESAMPLE
COLLECTION

of Bottles

A

Ice

1

2

3

4

5

6

7

8

9

1.

MW-06

GW

X

11/26/24

11:05

5

X

X

X

2.

MW-08

GW

X

11/26/24

11:55

5

X

X

X

3.

MW-10

GW

X

11/26/24

12:40

5

X

X

X

4.

MW-11

GW

X

11/26/24

13:20

5

X

X

X

5.

MW-13

GW

X

11/26/24

13:35

4

X

X

X

6.

MW-12

GW

X

11/26/24

13:50

5

X

X

X

7.

8.

9.

10.

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY: SAMPLER

DATE/TIME

RECEIVED BY

1.

11/26/24 1600

1.

RELINQUISHED BY

DATE/TIME

RECEIVED BY

2.

1005

2.

RELINQUISHED BY

DATE/TIME

RECEIVED FOR LAB BY

3.

DATE/TIME

RECEIVED BY

Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp 24°C

MeOH extraction requires an additional 4oz. Jar for percent solid

☐ Ice in Cooler? yes

Comments:

IR Count #1

Page _____ of _____

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight
CHEMTECH: ☐ Picked Up ☐ Overnight

Shipment Complete

☐ YES ☐ NO

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT

YELLOW - CHEMTECH COPY

PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : P5021 **LIRO01**

Order Date : 11/27/2024 10:41:00 AM

Project Mgr :

Client Name : LiRo Engineers, Inc.

Project Name : RFK Bridge

Report Type : NYS ASP A

Client Contact : Martin Wesolowski

Receive DateTime : 11/27/2024 10:05:00 AM

EDD Type : Excel NY

Invoice Name : LiRo Engineers, Inc.

Purchase Order :

Hard Copy Date :

Invoice Contact : Martin Wesolowski

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
P5021-01	MW-06	Water	11/26/2024	11:05					
					VOCMS Group1		8260-Low	10 Bus. Days	
P5021-02	MW-08	Water	11/26/2024	11:55					
					VOCMS Group1		8260-Low	10 Bus. Days	
P5021-03	MW-10	Water	11/26/2024	12:40					
					VOCMS Group1		8260-Low	10 Bus. Days	
P5021-04	MW-11	Water	11/26/2024	13:20					
					VOCMS Group1		8260-Low	10 Bus. Days	
P5021-05	MW-13	Water	11/26/2024	13:35					
					VOCMS Group1		8260-Low	10 Bus. Days	
P5021-06	MW-12	Water	11/26/2024	13:50					
					VOCMS Group1		8260-Low	10 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : P5021	LIRO01	Order Date : 11/27/2024 10:41:00 AM	Project Mgr :
Client Name : LiRo Engineers, Inc.		Project Name : RFK Bridge	Report Type : NYS ASPA
Client Contact : Martin Wesolowski		Receive DateTime : 11/27/2024 10:05:00 AM	EDD Type : Excel NY
Invoice Name : LiRo Engineers, Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Martin Wesolowski			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
--------	-----------	--------	----------------	----------------	------	------------	--------	----------	--------------

Relinquished By :

Date / Time : 11-27-24 1200

Received By :

Date / Time :

Storage Area : VOA Refridgerator Room