

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5021

Client: LiRo Engineers, Inc.

Analytical Method: 8270E DataFile: BF140684.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165296BSD	Acenaphthylene	50	54.9	ug/L	110	10	*		79	103	20
	Acenaphthene	50	52.4	ug/L	105	10			59	113	20
	Fluorene	50	52.2	ug/L	104	10			64	107	20
	Phenanthrene	50	53.2	ug/L	106	11			62	109	20
	Anthracene	50	54.6	ug/L	109	9			65	110	20
	Fluoranthene	50	46.8	ug/L	94	8			64	110	20
	Pyrene	50	52.5	ug/L	105	13	*		71	103	20
	Benzo(a)anthracene	50	56.0	ug/L	112	15	*		62	107	20
	Chrysene	50	54.9	ug/L	110	12	*		61	108	20
	Benzo(b)fluoranthene	50	55.5	ug/L	111	16			77	113	20
	Benzo(k)fluoranthene	50	54.0	ug/L	108	10	*		77	105	20
	Benzo(a)pyrene	50	62.5	ug/L	125	18			72	131	20
	Indeno(1,2,3-cd)pyrene	50	49.9	ug/L	100	5			72	105	20
	Dibenz(a,h)anthracene	50	49.7	ug/L	99	5			78	115	20
	Benzo(g,h,i)perylene	50	49.2	ug/L	98	11			75	118	20