

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021

EPA Sample No.: SSTDCCC040 Date Analyzed: 12/03/2024

Lab File ID: BF140702.D Time Analyzed: 11:24

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	78577	6.869	287976	8.15	161859	9.91
UPPER LIMIT	157154	7.369	575952	8.651	323718	10.41
LOWER LIMIT	39288.5	6.369	143988	7.651	80929.5	9.41
EPA SAMPLE NO.						
01 MW-08	83408	6.87	298688	8.15	174661	9.90
02 MW-10	70972	6.87	269570	8.15	149746	9.90
03 MW-12	67828	6.87	251382	8.15	155575	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.