

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: LIRO01

Lab Code: CHEM

SAS No.: P5021

SDG NO.: P5021

Lab File ID: BF140680.D

DFTPP Injection Date: 12/02/2024

Instrument ID: BNA_F

DFTPP Injection Time: 09:05

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.7
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	35
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	45.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
275	10.0 - 60.0% of mass 198	29.2
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	15.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.4 (18.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140681.D	12/02/2024	10:22
PB165296BL	PB165296BL	BF140682.D	12/02/2024	10:57
PB165296BS	PB165296BS	BF140683.D	12/02/2024	11:23
PB165296BSD	PB165296BSD	BF140684.D	12/02/2024	11:49
MW-13	P5021-05	BF140688.D	12/02/2024	15:03
MW-11	P5021-04	BF140689.D	12/02/2024	15:29
MW-06	P5021-01	BF140691.D	12/02/2024	16:21