

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: LIRO01

Lab Code: CHEM

SAS No.: P5021

SDG NO.: P5021

Lab File ID: BF140701.D

DFTPP Injection Date: 12/03/2024

Instrument ID: BNA_F

DFTPP Injection Time: 10:24

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.9
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	35
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	46.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	28.2
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.7 (18.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140702.D	12/03/2024	11:24
MW-08	P5021-02	BF140710.D	12/03/2024	14:59
MW-10	P5021-03	BF140711.D	12/03/2024	15:24
MW-12	P5021-06	BF140712.D	12/03/2024	15:50