

Report of Analysis

Client:	LiRo Engineers, Inc.	Date Collected:	
Project:	RFK Bridge	Date Received:	
Client Sample ID:	VX1202WBS01	SDG No.:	P5021
Lab Sample ID:	VX1202WBS01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044055.D	1		12/02/24 11:00	VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	20.3		0.16	1.00	ug/L
71-43-2	Benzene	19.4		0.16	1.00	ug/L
108-88-3	Toluene	19.4		0.18	1.00	ug/L
100-41-4	Ethyl Benzene	20.7		0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	41.5		0.31	2.00	ug/L
95-47-6	o-Xylene	21.3		0.14	1.00	ug/L
98-82-8	Isopropylbenzene	20.6		0.13	1.00	ug/L
103-65-1	n-propylbenzene	20.7		0.14	1.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	21.3		0.18	1.00	ug/L
98-06-6	tert-Butylbenzene	21.3		0.17	1.00	ug/L
95-63-6	1,2,4-Trimethylbenzene	21.2		0.18	1.00	ug/L
135-98-8	sec-Butylbenzene	20.8		0.17	1.00	ug/L
99-87-6	p-Isopropyltoluene	21.5		0.15	1.00	ug/L
104-51-8	n-Butylbenzene	20.1		0.22	1.00	ug/L
91-20-3	Naphthalene	20.3		0.59	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.9		74 - 125	110%	SPK: 50
1868-53-7	Dibromofluoromethane	53.7		75 - 124	107%	SPK: 50
2037-26-5	Toluene-d8	50.8		86 - 113	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		77 - 121	105%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	134000	5.55			
540-36-3	1,4-Difluorobenzene	237000	6.757			
3114-55-4	Chlorobenzene-d5	191000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	90100	12.018			

Report of Analysis

Client: LiRo Engineers, Inc.

Date Collected:

Project: RFK Bridge

Date Received:

Client Sample ID: VX1202WBS01

SDG No.: P5021

Lab Sample ID: VX1202WBS01

Matrix: Water

Analytical Method: SW8260

% Solid: 0

Sample Wt/Vol: 5 Units: mL

Final Vol: 5000 uL

Soil Aliquot Vol: _____ uL

Test: VOCMS Group1

GC Column: DB-624UI ID : 0.18

Level : LOW

Prep Method :

File ID/Qc Batch:

Dilution:

Prep Date

Date Analyzed

Prep Batch ID

VX044055.D

1

12/02/24 11:00

VX120224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products