

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX112124

Review By	John Carlone	Review On	11/21/2024 9:44:50 AM
Supervise By	Maresh Dadoda	Supervise On	11/22/2024 3:34:00 PM
SubDirectory	VX112124	HP Acquire Method	HP Processing Method 82X112124W.M

STD. NAME	STD REF.#
Tune/Reschk	VP131686,VP131690
Initial Calibration Stds	VP131716,VP131717,VP131718,VP131719,VP131720,VP131721
CCC	VP131691
Internal Standard/PEM	
ICV/I.BLK	VP131722
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX043915.D	20 Nov 2024 15:43		JC/MD	Not Ok
2	VSTDIC001	VSTDIC001	VX043916.D	20 Nov 2024 16:11	%D failed for Comp. #71,74,90,92 in 01 PPB	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX043917.D	20 Nov 2024 16:57	%D failed for Comp. #81 in 05 PPB	JC/MD	Not Ok
4	VSTDIC020	VSTDIC020	VX043918.D	20 Nov 2024 17:20		JC/MD	Not Ok
5	VSTDIC050	VSTDIC050	VX043919.D	20 Nov 2024 17:43		JC/MD	Not Ok
6	VSTDIC100	VSTDIC100	VX043920.D	20 Nov 2024 18:06		JC/MD	Not Ok
7	VSTDIC150	VSTDIC150	VX043921.D	20 Nov 2024 18:30		JC/MD	Not Ok
8	IBLK	IBLK	VX043922.D	20 Nov 2024 18:53		JC/MD	Not Ok
9	VSTDICV050	ICVVX112124	VX043923.D	20 Nov 2024 19:16	Failed	JC/MD	Not Ok
10	BFB	BFB	VX043924.D	21 Nov 2024 08:51		JC/MD	Ok
11	VSTDIC001	VSTDIC001	VX043925.D	21 Nov 2024 09:47	%D failed for Comp. #71 in 01 PPB	JC/MD	Ok,M
12	VSTDIC005	VSTDIC005	VX043926.D	21 Nov 2024 10:14	QR- 58,71	JC/MD	Ok,M
13	VSTDIC020	VSTDIC020	VX043927.D	21 Nov 2024 10:37		JC/MD	Ok,M
14	VSTDIC050	VSTDIC050	VX043928.D	21 Nov 2024 11:17		JC/MD	Ok,M
15	VSTDIC100	VSTDIC100	VX043929.D	21 Nov 2024 11:40		JC/MD	Ok,M
16	VSTDIC150	VSTDIC150	VX043930.D	21 Nov 2024 12:03		JC/MD	Ok,M
17	IBLK	IBLK	VX043931.D	21 Nov 2024 12:27		JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX112124

Review By	John Carlone	Review On	11/21/2024 9:44:50 AM
Supervise By	Maresh Dadoda	Supervise On	11/22/2024 3:34:00 PM
SubDirectory	VX112124	HP Acquire Method	HP Processing Method 82X112124W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP131686,VP131690		
Initial Calibration Stds	VP131716,VP131717,VP131718,VP131719,VP131720,VP131721		
CCC	VP131691		
Internal Standard/PEM			
ICV/I.BLK	VP131722		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	VSTDICV050	ICVVX112124	VX043932.D	21 Nov 2024 13:57		JC/MD	Ok,M
19	VX1121MBL01	VX1121MBL01	VX043933.D	21 Nov 2024 14:27		JC/MD	Ok
20	VX1121WBL01	VX1121WBL01	VX043934.D	21 Nov 2024 14:50		JC/MD	Ok
21	VX1121WBS01	VX1121WBS01	VX043935.D	21 Nov 2024 15:13		JC/MD	Ok,M
22	VX1121WBSD01	VX1121WBSD01	VX043936.D	21 Nov 2024 15:40		JC/MD	Ok,M
23	P4921-01	WC-11-A-202411	VX043937.D	21 Nov 2024 16:03	vial B pH#5.0	JC/MD	Ok
24	P4934-01	TB-01-20241118	VX043938.D	21 Nov 2024 16:27	vial A pH<2 TB;	JC/MD	Ok
25	P4934-09	FB-01-20241119	VX043939.D	21 Nov 2024 16:50	vial A pH<2 FB;	JC/MD	Ok
26	P4934-11	RB-01-20241119	VX043940.D	21 Nov 2024 17:13	vial A pH<2	JC/MD	Ok
27	P4934-02	BPOW6-11-20241118	VX043941.D	21 Nov 2024 17:36	vial A pH<2	JC/MD	Ok
28	P4934-03	BPOW6-7-20241118	VX043942.D	21 Nov 2024 17:59	vial A pH<2	JC/MD	Ok
29	P4934-08	BPOW6-9-20241119	VX043943.D	21 Nov 2024 18:23	vial A pH<2	JC/MD	Ok
30	P4934-10	BPOW6-8-20241119	VX043944.D	21 Nov 2024 18:46	vial A pH<2	JC/MD	Ok
31	P4934-04	DUP-01-20241118	VX043945.D	21 Nov 2024 19:09	vial A pH<2	JC/MD	Ok
32	P4934-12	DUP-02-20241119	VX043946.D	21 Nov 2024 19:32	vial A pH<2	JC/MD	Ok
33	P4934-05	BPOW6-10-20241119	VX043947.D	21 Nov 2024 19:55	vial A pH<2	JC/MD	Ok
34	VSTDCCC050	VSTDCCC050EC	VX043948.D	21 Nov 2024 20:18		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX112724

Review By	John Carlone	Review On	12/2/2024 10:20:52 AM
Supervise By	Mahesh Dadoda	Supervise On	12/2/2024 10:24:46 AM
SubDirectory	VX112724	HP Acquire Method	HP Processing Method 82X112124W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131824
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131825,VP131826

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044023.D	27 Nov 2024 07:52		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX044024.D	27 Nov 2024 08:46	V13516	JC/MD	Ok,M
3	VX1127MBL01	VX1127MBL01	VX044025.D	27 Nov 2024 09:15		JC/MD	Ok
4	VX1127WBL01	VX1127WBL01	VX044026.D	27 Nov 2024 09:38		JC/MD	Ok
5	VX1127MBS01	VX1127MBS01	VX044027.D	27 Nov 2024 10:01		JC/MD	Ok,M
6	VX1127WBS01	VX1127WBS01	VX044028.D	27 Nov 2024 10:27		JC/MD	Ok,M
7	P5006-18	TW-WTS-02	VX044029.D	27 Nov 2024 10:50		JC/MD	Ok
8	PB165270TB	PB165270TB	VX044030.D	27 Nov 2024 11:13		JC/MD	Ok
9	P5001-01	SPLP-C1-1	VX044031.D	27 Nov 2024 11:36	vial A pH#6.0 Surrogate fail;BSD failed low	JC/MD	ReRun
10	P5001-02	SPLP-C1-2	VX044032.D	27 Nov 2024 11:59	vial A pH#6.0 Surrogate fail;BSD failed low	JC/MD	ReRun
11	P5001-03	SPLP-C1-3	VX044033.D	27 Nov 2024 12:22	vial A pH#6.0 Surrogate fail;BSD failed low	JC/MD	ReRun
12	P5001-04	SPLP-C2-1	VX044034.D	27 Nov 2024 12:45	vial A pH#6.0 Surrogate fail;BSD failed low	JC/MD	ReRun
13	P5001-05	SPLP-C2-2	VX044035.D	27 Nov 2024 13:08	vial A pH#6.0 Surrogate fail;BSD failed low;Need 10X	JC/MD	Dilution
14	P5001-06	SPLP-C2-3	VX044036.D	27 Nov 2024 13:31	vial A pH#6.0 Surrogate fail;BSD failed low;Need 10X	JC/MD	Dilution
15	P5001-08	SPLP-C3-1	VX044037.D	27 Nov 2024 13:54	vial A pH#6.0 E Flag in Previous Sample;Surrogate fail;BSD failed low	JC/MD	ReRun
16	P5001-09	SPLP-C3-2	VX044038.D	27 Nov 2024 14:16	vial A pH#6.0 Surrogate fail;BSD failed low;Need 10X	JC/MD	Dilution

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX112724

Review By	John Carlone	Review On	12/2/2024 10:20:52 AM
Supervise By	Maresh Dadoda	Supervise On	12/2/2024 10:24:46 AM
SubDirectory	VX112724	HP Acquire Method	HP Processing Method 82X112124W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP131824		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131825,VP131826		

17	P5001-10	SPLP-C3-3	VX044039.D	27 Nov 2024 14:39	vial A pH#6.0 Surrogate fail;BSD failed low;Need 10X	JC/MD	Dilution
18	P5001-11	SPLP-C4-1	VX044040.D	27 Nov 2024 15:02	vial A pH#6.0 Surrogate fail;BSD failed low	JC/MD	ReRun
19	VX1127WBSD01	VX1127WBSD01	VX044041.D	27 Nov 2024 15:25	Recovery failed low for comp. #44,60,71 and failed high for comp. #95	JC/MD	Ok,M
20	VX1127MBSD01	VX1127MBSD01	VX044042.D	27 Nov 2024 15:48		JC/MD	Ok,M
21	P4960-05	SW3	VX044043.D	27 Nov 2024 16:11		JC/MD	Ok
22	P5021-02	MW-08	VX044044.D	27 Nov 2024 16:34	vial A pH<2 BSD failed high;Run @ Lower Dilution	JC/MD	ReRun
23	P5021-01	MW-06	VX044045.D	27 Nov 2024 16:57	vial A pH<2 BSD failed high	JC/MD	ReRun
24	P5021-03	MW-10	VX044046.D	27 Nov 2024 17:20	vial A pH<2 BSD failed high	JC/MD	ReRun
25	P5021-04	MW-11	VX044047.D	27 Nov 2024 17:43	vial A pH<2 BSD failed high	JC/MD	ReRun
26	P5021-05	MW-13	VX044048.D	27 Nov 2024 18:06	vial A pH<2 BSD failed high	JC/MD	ReRun
27	P5021-06	MW-12	VX044049.D	27 Nov 2024 18:29	vial A pH<2 BSD failed high;Need 20X	JC/MD	Dilution
28	VSTDCCC050	VSTDCCC050EC	VX044050.D	27 Nov 2024 18:52		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX120224

Review By	John Carlone	Review On	12/3/2024 9:50:50 AM
Supervise By	Maresh Dadoda	Supervise On	12/3/2024 1:39:36 PM
SubDirectory	VX120224	HP Acquire Method	HP Processing Method 82X112124W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP131840
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131841,VP131842

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX044051.D	02 Dec 2024 08:35		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX044052.D	02 Dec 2024 09:45	V13516	JC/MD	Ok,M
3	VX1202MBL01	VX1202MBL01	VX044053.D	02 Dec 2024 10:14		JC/MD	Ok
4	VX1202WBL01	VX1202WBL01	VX044054.D	02 Dec 2024 10:37		JC/MD	Ok
5	VX1202WBS01	VX1202WBS01	VX044055.D	02 Dec 2024 11:00		JC/MD	Ok,M
6	VX1202WBSD01	VX1202WBSD01	VX044056.D	02 Dec 2024 11:27		JC/MD	Ok,M
7	P5021-02	MW-08	VX044057.D	02 Dec 2024 11:49	vial B pH<2	JC/MD	Ok
8	P5021-06DL	MW-12DL	VX044058.D	02 Dec 2024 12:12	vial B pH<2	JC/MD	Ok
9	P5021-01	MW-06	VX044059.D	02 Dec 2024 12:35	vial B pH<2	JC/MD	Ok,M
10	P5021-03	MW-10	VX044060.D	02 Dec 2024 12:58	vial B pH<2	JC/MD	Ok
11	P5021-04	MW-11	VX044061.D	02 Dec 2024 13:21	vial B pH<2	JC/MD	Ok
12	P5021-05	MW-13	VX044062.D	02 Dec 2024 13:44	vial B pH<2	JC/MD	Ok
13	PB165251TB	PB165251TB	VX044063.D	02 Dec 2024 14:07		JC/MD	Ok
14	P4995-02	001	VX044064.D	02 Dec 2024 14:30	vial A pH#5.0	JC/MD	Ok
15	P5001-05DL	SPLP-C2-2DL	VX044065.D	02 Dec 2024 14:52	vial B pH#6.0	JC/MD	Ok
16	P5001-06DL	SPLP-C2-3DL	VX044066.D	02 Dec 2024 15:15	vial B pH#6.0	JC/MD	Ok
17	P5001-09DL	SPLP-C3-2DL	VX044067.D	02 Dec 2024 15:38	vial B pH#6.0	JC/MD	Ok
18	P5001-10DL	SPLP-C3-3DL	VX044068.D	02 Dec 2024 16:01	vial B pH#6.0	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX120224

Review By	John Carlone	Review On	12/3/2024 9:50:50 AM				
Supervise By	Maresh Dadoda	Supervise On	12/3/2024 1:39:36 PM				
SubDirectory	VX120224	HP Acquire Method	HP Processing Method		82X112124W.M		
STD. NAME		STD REF.#					
Tune/Reschk Initial Calibration Stds		VP131840					
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP131841,VP131842					

19	P5001-01RE	SPLP-C1-1RE	VX044069.D	02 Dec 2024 16:24	vial B pH#6.0 Surrogate fail	JC/MD	Confirms
20	P5001-02RE	SPLP-C1-2RE	VX044070.D	02 Dec 2024 16:47	vial B pH#6.0 Surrogate fail	JC/MD	Confirms
21	P5001-03RE	SPLP-C1-3RE	VX044071.D	02 Dec 2024 17:10	vial B pH#6.0 Surrogate fail	JC/MD	Confirms
22	P5001-04RE	SPLP-C2-1RE	VX044072.D	02 Dec 2024 17:32	vial B pH#6.0 Surrogate fail	JC/MD	Confirms
23	P5001-08RE	SPLP-C3-1RE	VX044073.D	02 Dec 2024 17:55	vial B pH#6.0 Surrogate fail	JC/MD	Confirms
24	P5001-11RE	SPLP-C4-1RE	VX044074.D	02 Dec 2024 18:18	vial B pH#6.0 Surrogate fail	JC/MD	Confirms
25	PB165299TB	PB165299TB	VX044075.D	02 Dec 2024 18:41		JC/MD	Ok
26	P5006-01	SPLP-C4-2	VX044076.D	02 Dec 2024 19:04	vial A pH#6.0 Surrogate fail	JC/MD	ReRun
27	P5006-02	SPLP-C4-3	VX044077.D	02 Dec 2024 19:27	vial A pH#6.0 Surrogate fail	JC/MD	ReRun
28	VSTDCCC050	VSTDCCC050EC	VX044078.D	02 Dec 2024 19:50		JC/MD	Ok,M

M : Manual Integration