

Surrogate Summary

SDG No.: **P5021**

Client: **LiRo Engineers, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5021-01	MW-06	1,2-Dichloroethane-d4	50	51.0	102	74	125
		Dibromofluoromethane	50	47.5	95	75	124
		Toluene-d8	50	48.9	98	86	113
		4-Bromofluorobenzene	50	52.0	104	77	121
P5021-02	MW-08	1,2-Dichloroethane-d4	50	48.9	98	74	125
		Dibromofluoromethane	50	48.0	96	75	124
		Toluene-d8	50	48.8	98	86	113
		4-Bromofluorobenzene	50	51.6	103	77	121
P5021-03	MW-10	1,2-Dichloroethane-d4	50	51.8	104	74	125
		Dibromofluoromethane	50	46.4	93	75	124
		Toluene-d8	50	49.3	99	86	113
		4-Bromofluorobenzene	50	50.0	100	77	121
P5021-04	MW-11	1,2-Dichloroethane-d4	50	51.9	104	74	125
		Dibromofluoromethane	50	47.0	94	75	124
		Toluene-d8	50	51.0	102	86	113
		4-Bromofluorobenzene	50	51.6	103	77	121
P5021-05	MW-13	1,2-Dichloroethane-d4	50	53.1	106	74	125
		Dibromofluoromethane	50	46.7	93	75	124
		Toluene-d8	50	50.2	100	86	113
		4-Bromofluorobenzene	50	51.1	102	77	121
P5021-06	MW-12	1,2-Dichloroethane-d4	50	53.2	106	74	125
		Dibromofluoromethane	50	46.4	93	75	124
		Toluene-d8	50	50.2	100	86	113
		4-Bromofluorobenzene	50	52.3	105	77	121
P5021-06DL	MW-12DL	1,2-Dichloroethane-d4	50	51.2	102	74	125
		Dibromofluoromethane	50	46.3	93	75	124
		Toluene-d8	50	49.4	99	86	113
		4-Bromofluorobenzene	50	51.1	102	77	121
VX1127WBL01	VX1127WBL01	1,2-Dichloroethane-d4	50	50.8	102	74	125
		Dibromofluoromethane	50	46.6	93	75	124
		Toluene-d8	50	48.9	98	86	113
		4-Bromofluorobenzene	50	47.2	94	77	121
VX1127WBS01	VX1127WBS01	1,2-Dichloroethane-d4	50	54.7	109	74	125
		Dibromofluoromethane	50	53.5	107	75	124
		Toluene-d8	50	53.1	106	86	113
		4-Bromofluorobenzene	50	53.6	107	77	121
VX1127WBSD0	VX1127WBSD01	1,2-Dichloroethane-d4	50	51.5	103	74	125
		Dibromofluoromethane	50	46.2	92	75	124
		Toluene-d8	50	47.1	94	86	113
		4-Bromofluorobenzene	50	48.5	97	77	121
VX1202WBL01	VX1202WBL01	1,2-Dichloroethane-d4	50	49.6	99	74	125
		Dibromofluoromethane	50	47.3	95	75	124
		Toluene-d8	50	48.6	97	86	113
		4-Bromofluorobenzene	50	46.5	93	77	121
VX1202WBS01	VX1202WBS01	1,2-Dichloroethane-d4	50	54.9	110	74	125
		Dibromofluoromethane	50	53.7	107	75	124
		Toluene-d8	50	50.8	102	86	113
		4-Bromofluorobenzene	50	52.8	105	77	121