

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5021

Client: LiRo Engineers, Inc.

Analytical Method: SW8260-Low

Datafile : VX044041.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1127WBSD01	Methyl tert-butyl Ether	20	19.8	ug/L	99	9		78	114	20
	Benzene	20	17.4	ug/L	87	18		82	109	20
	Toluene	20	18.4	ug/L	92	14		82	110	20
	Ethyl Benzene	20	18.5	ug/L	93	13		83	109	20
	m/p-Xylenes	40	36.1	ug/L	90	16		82	110	20
	o-Xylene	20	18.8	ug/L	94	11		83	109	20
	Isopropylbenzene	20	18.8	ug/L	94	14		83	112	20
	N-propylbenzene	20	18.9	ug/L	95	13		83	112	20
	1,3,5-Trimethylbenzene	20	19.1	ug/L	96	14		85	112	20
	tert-Butylbenzene	20	18.5	ug/L	93	15		83	112	20
	1,2,4-Trimethylbenzene	20	19.3	ug/L	97	13		85	111	20
	Sec-butylbenzene	20	18.5	ug/L	93	15		81	114	20
	p-Isopropyltoluene	20	19.0	ug/L	95	16		78	116	20
	n-Butylbenzene	20	18.5	ug/L	93	15		75	115	20
	Naphthalene	20	32.1	ug/L	160	42	* *	78	119	20