

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5021

Client: LiRo Engineers, Inc.

Analytical Method: SW8260-Low

Datafile : VX044055.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1202WBS01	Methyl tert-butyl Ether	20	20.3	ug/L	102			78	114	
	Benzene	20	19.4	ug/L	97			82	109	
	Toluene	20	19.4	ug/L	97			82	110	
	Ethyl Benzene	20	20.7	ug/L	104			83	109	
	m/p-Xylenes	40	41.5	ug/L	104			82	110	
	o-Xylene	20	21.3	ug/L	106			83	109	
	Isopropylbenzene	20	20.6	ug/L	103			83	112	
	N-propylbenzene	20	20.7	ug/L	104			83	112	
	1,3,5-Trimethylbenzene	20	21.3	ug/L	106			85	112	
	tert-Butylbenzene	20	21.3	ug/L	106			83	112	
	1,2,4-Trimethylbenzene	20	21.2	ug/L	106			85	111	
	Sec-butylbenzene	20	20.8	ug/L	104			81	114	
	p-Isopropyltoluene	20	21.5	ug/L	108			78	116	
	n-Butylbenzene	20	20.1	ug/L	101			75	115	
	Naphthalene	20	20.3	ug/L	102			78	119	