

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LIRO01
Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021
Lab File ID: VX044024.D Date Analyzed: 11/27/2024
Instrument ID: MSVOA_X Time Analyzed: 08:46
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	130547	5.54	218884	6.75	184783	10.05
UPPER LIMIT	261094	6.037	437768	7.251	369566	10.549
LOWER LIMIT	65273.5	5.037	109442	6.251	92391.5	9.549
EPA SAMPLE NO.						
MW-12	112926	5.55	218400	6.76	192264	10.06
VX1127WBL01	125569	5.54	244455	6.75	212718	10.05
VX1127WBS01	106843	5.54	192894	6.76	162801	10.05
VX1127WBSD01	127369	5.54	240268	6.76	205658	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.