

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LIRO01
 Lab Code: CHEM Case No.: P5021 SAS No.: P5021 SDG NO.: P5021
 Lab File ID: VX044052.D Date Analyzed: 12/02/2024
 Instrument ID: MSVOA_X Time Analyzed: 09:45
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	140021	5.54	242098	6.75	204952	10.05
UPPER LIMIT	280042	6.043	484196	7.25	409904	10.549
LOWER LIMIT	70010.5	5.043	121049	6.25	102476	9.549
EPA SAMPLE NO.						
MW-06	137934	5.54	259323	6.76	223099	10.06
MW-08	123344	5.54	229138	6.76	199363	10.05
MW-10	121483	5.54	237218	6.75	202018	10.06
MW-11	120070	5.54	229971	6.76	208817	10.05
MW-13	125796	5.54	244811	6.76	217402	10.05
MW-12DL	125735	5.54	244727	6.76	215420	10.06
VX1202WBL01	137622	5.54	262897	6.76	221008	10.05
VX1202WBS01	133564	5.55	236701	6.76	191057	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.