

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5022

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133665

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10	10	100	90-110	11/18/2024
Chloride	mg/L	3	3	100	90-110	11/18/2024
Fluoride	mg/L	2	2	100	90-110	11/18/2024
Nitrite	mg/L	3	3	100	90-110	11/18/2024
Nitrate	mg/L	2.5	2.5	100	90-110	11/18/2024
Sulfate	mg/L	14.6	15	97	90-110	11/18/2024
Orthophosphate as P	mg/L	5	5	100	90-110	11/18/2024
Sample ID: CCV1						
Bromide	mg/L	10.4	10	104	90-110	11/27/2024
Chloride	mg/L	3.1	3	103	90-110	11/27/2024
Fluoride	mg/L	2	2	100	90-110	11/27/2024
Nitrite	mg/L	3.1	3	103	90-110	11/27/2024
Nitrate	mg/L	2.6	2.5	104	90-110	11/27/2024
Sulfate	mg/L	14.7	15	98	90-110	11/27/2024
Orthophosphate as P	mg/L	5.2	5	104	90-110	11/27/2024
Sample ID: CCV2						
Bromide	mg/L	10.4	10	104	90-110	11/27/2024
Chloride	mg/L	3.1	3	103	90-110	11/27/2024
Fluoride	mg/L	2	2	100	90-110	11/27/2024
Nitrite	mg/L	3.1	3	103	90-110	11/27/2024
Nitrate	mg/L	2.6	2.5	104	90-110	11/27/2024
Sulfate	mg/L	14.7	15	98	90-110	11/27/2024
Orthophosphate as P	mg/L	5.2	5	104	90-110	11/27/2024

Initial and Continuing Calibration Verification**Client:** Weston Solutions**SDG No.:** P5022**Project:** Ft Meade Tipton Airfield Parcel RI - PO 0111169**RunNo.:** LB133690

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.00	7	100	90-110	12/02/2024
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	12/02/2024
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	12/02/2024

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5022

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133779

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
TOC		mg/L	934	1000	93	90-110	11/12/2024
Sample ID:	CCV1						
TOC		mg/L	937	1000	94	90-110	12/06/2024
Sample ID:	CCV2						
TOC		mg/L	999	1000	100	90-110	12/06/2024