

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Weston Solutions **SDG No.:** P5022  
**Contract:** WEST04 **Lab Code:** CHEM **Case No.:** P5022 **SAS No.:** P5022  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** PLASMA-PURE

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|----|------------------|------------------|---------------|
| ICV10     | Mercury | 3.85           | 4.0        | 96            | 90 - 110                  | CV | 12/02/2024       | 15:45            | LB133693      |

---

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** Weston Solutions **SDG No.:** P5022  
**Contract:** WEST04 **Lab Code:** CHEM **Case No.:** P5022 **SAS No.:** P5022  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** PLASMA-PURE

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|----|------------------|------------------|---------------|
| CCV72     | Mercury | 4.91           | 5.0        | 98            | 90 - 110                  | CV | 12/02/2024       | 15:52            | LB133693      |
| CCV73     | Mercury | 5.18           | 5.0        | 104           | 90 - 110                  | CV | 12/02/2024       | 16:20            | LB133693      |
| CCV74     | Mercury | 5.19           | 5.0        | 104           | 90 - 110                  | CV | 12/02/2024       | 16:45            | LB133693      |