

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions

SDG No.: P5022

Contract: WEST04

Lab Code: CHEM

Case No.: P5022

SAS No.: P5022

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165333BS Mercury	mg/Kg	0.25	0.22		88	80 - 124	CV