

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P5026
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5026 **SAS No.:** P5026
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV109	Mercury	3.79	4.0	95	90 - 110	CV	12/03/2024	10:58	LB133702

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P5026
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5026 **SAS No.:** P5026
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV75	Mercury	5.17	5.0	103	90 - 110	CV	12/03/2024	11:02	LB133702
CCV76	Mercury	5.04	5.0	101	90 - 110	CV	12/03/2024	11:33	LB133702
CCV77	Mercury	5.22	5.0	104	90 - 110	CV	12/03/2024	12:08	LB133702
CCV78	Mercury	5.34	5.0	107	90 - 110	CV	12/03/2024	12:29	LB133702

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P5026
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5026 **SAS No.:** P5026
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1000	1000	100	90 - 110	P	12/03/2024	17:37	LB133718
	Barium	511	520	98	90 - 110	P	12/03/2024	17:37	LB133718
	Cadmium	510	510	100	90 - 110	P	12/03/2024	17:37	LB133718
	Chromium	533	520	102	90 - 110	P	12/03/2024	17:37	LB133718
	Lead	1010	1000	101	90 - 110	P	12/03/2024	17:37	LB133718
	Selenium	1030	1000	103	90 - 110	P	12/03/2024	17:37	LB133718
	Silver	254	250	102	90 - 110	P	12/03/2024	17:37	LB133718

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. SDG No.: P5026

Contract: TULL02 Lab Code: CHEM Case No.: P5026 SAS No.: P5026

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	21.1	20.0	106	80 - 120	P	12/03/2024	17:41	LB133718
	Barium	96.2	100	96	80 - 120	P	12/03/2024	17:41	LB133718
	Cadmium	6.12	6.0	102	80 - 120	P	12/03/2024	17:41	LB133718
	Chromium	10.5	10.0	105	80 - 120	P	12/03/2024	17:41	LB133718
	Lead	11.5	12.0	96	80 - 120	P	12/03/2024	17:41	LB133718
	Selenium	20.5	20.0	102	80 - 120	P	12/03/2024	17:41	LB133718
	Silver	10.7	10.0	107	80 - 120	P	12/03/2024	17:41	LB133718

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P5026

Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5026 **SAS No.:** P5026

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4830	5000	97	90 - 110	P	12/03/2024	18:18	LB133718
	Barium	9260	10000	93	90 - 110	P	12/03/2024	18:18	LB133718
	Cadmium	2320	2500	93	90 - 110	P	12/03/2024	18:18	LB133718
	Chromium	944	1000	94	90 - 110	P	12/03/2024	18:18	LB133718
	Lead	4670	5000	93	90 - 110	P	12/03/2024	18:18	LB133718
	Selenium	4840	5000	97	90 - 110	P	12/03/2024	18:18	LB133718
	Silver	1190	1250	95	90 - 110	P	12/03/2024	18:18	LB133718
CCV02	Arsenic	5190	5000	104	90 - 110	P	12/03/2024	18:43	LB133718
	Barium	9530	10000	95	90 - 110	P	12/03/2024	18:43	LB133718
	Cadmium	2420	2500	97	90 - 110	P	12/03/2024	18:43	LB133718
	Chromium	1000	1000	100	90 - 110	P	12/03/2024	18:43	LB133718
	Lead	4860	5000	97	90 - 110	P	12/03/2024	18:43	LB133718
	Selenium	5220	5000	104	90 - 110	P	12/03/2024	18:43	LB133718
	Silver	1250	1250	100	90 - 110	P	12/03/2024	18:43	LB133718
CCV03	Arsenic	4870	5000	98	90 - 110	P	12/03/2024	19:48	LB133718
	Barium	10600	10000	106	90 - 110	P	12/03/2024	19:48	LB133718
	Cadmium	2290	2500	92	90 - 110	P	12/03/2024	19:48	LB133718
	Chromium	943	1000	94	90 - 110	P	12/03/2024	19:48	LB133718
	Lead	4600	5000	92	90 - 110	P	12/03/2024	19:48	LB133718
	Selenium	4890	5000	98	90 - 110	P	12/03/2024	19:48	LB133718
	Silver	1180	1250	94	90 - 110	P	12/03/2024	19:48	LB133718
CCV04	Arsenic	4850	5000	97	90 - 110	P	12/03/2024	20:47	LB133718
	Barium	9130	10000	91	90 - 110	P	12/03/2024	20:47	LB133718
	Cadmium	2300	2500	92	90 - 110	P	12/03/2024	20:47	LB133718
	Chromium	942	1000	94	90 - 110	P	12/03/2024	20:47	LB133718
	Lead	4620	5000	92	90 - 110	P	12/03/2024	20:47	LB133718
	Selenium	4860	5000	97	90 - 110	P	12/03/2024	20:47	LB133718
	Silver	1170	1250	94	90 - 110	P	12/03/2024	20:47	LB133718
CCV05	Arsenic	4790	5000	96	90 - 110	P	12/03/2024	21:20	LB133718
	Barium	9220	10000	92	90 - 110	P	12/03/2024	21:20	LB133718
	Cadmium	2330	2500	93	90 - 110	P	12/03/2024	21:20	LB133718
	Chromium	969	1000	97	90 - 110	P	12/03/2024	21:20	LB133718
	Lead	4700	5000	94	90 - 110	P	12/03/2024	21:20	LB133718
	Selenium	4780	5000	96	90 - 110	P	12/03/2024	21:20	LB133718

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Construction Co., Inc. **SDG No.:** P5026
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5026 **SAS No.:** P5026
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1200	1250	96	90 - 110	P	12/03/2024	21:20	LB133718
