



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Construction Co., Inc.

SDG No.: P5026

Contract: TULL02

Lab Code: CHEM

Case No.: P5026

SAS No.: P5026

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	83	40 - 160	CV	12/03/2024	11:07	LB133702
CRI01	Arsenic	19.9	20.0	100	40 - 160	P	12/03/2024	17:54	LB133718
	Barium	88.2	100	88	40 - 160	P	12/03/2024	17:54	LB133718
	Cadmium	5.63	6.0	94	40 - 160	P	12/03/2024	17:54	LB133718
	Chromium	10.1	10.0	101	40 - 160	P	12/03/2024	17:54	LB133718
	Lead	11.0	12.0	92	40 - 160	P	12/03/2024	17:54	LB133718
	Selenium	20.8	20.0	104	40 - 160	P	12/03/2024	17:54	LB133718
	Silver	10.2	10.0	102	40 - 160	P	12/03/2024	17:54	LB133718