



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: P5026

Contract: TULL02

Lab Code: CHEM

Case No.: P5026

SAS No.: P5026

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB109	Mercury	0.20	+/-0.20	U	0.20	CV	12/03/2024	11:00	LB133702

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: P5026

Contract: TULL02 **Lab Code:** CHEM

Case No.: P5026 **SAS No.:** P5026

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB75	Mercury	0.20	+/-0.20	U	0.20	CV	12/03/2024	11:04	LB133702
CCB76	Mercury	0.20	+/-0.20	U	0.20	CV	12/03/2024	11:35	LB133702
CCB77	Mercury	0.20	+/-0.20	U	0.20	CV	12/03/2024	12:10	LB133702
CCB78	Mercury	0.20	+/-0.20	U	0.20	CV	12/03/2024	12:31	LB133702

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc. **SDG No.:** P5026
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5026 **SAS No.:** P5026

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	12/03/2024	17:50	LB133718
	Barium	100	+/-100	U	100	P	12/03/2024	17:50	LB133718
	Cadmium	6.00	+/-6.00	U	6.00	P	12/03/2024	17:50	LB133718
	Chromium	10.0	+/-10.0	U	10.0	P	12/03/2024	17:50	LB133718
	Lead	12.0	+/-12.0	U	12.0	P	12/03/2024	17:50	LB133718
	Selenium	20.0	+/-20.0	U	20.0	P	12/03/2024	17:50	LB133718
	Silver	10.0	+/-10.0	U	10.0	P	12/03/2024	17:50	LB133718

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tully Construction Co., Inc.</u>		SDG No.: <u>P5026</u>							
Contract: <u>TULL02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5026</u>	SAS No.: <u>P5026</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	12/03/2024	18:22	LB133718
	Barium	100	+/-100	U	100	P	12/03/2024	18:22	LB133718
	Cadmium	6.00	+/-6.00	U	6.00	P	12/03/2024	18:22	LB133718
	Chromium	10.0	+/-10.0	U	10.0	P	12/03/2024	18:22	LB133718
	Lead	12.0	+/-12.0	U	12.0	P	12/03/2024	18:22	LB133718
	Selenium	20.0	+/-20.0	U	20.0	P	12/03/2024	18:22	LB133718
	Silver	10.0	+/-10.0	U	10.0	P	12/03/2024	18:22	LB133718
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	12/03/2024	18:47	LB133718
	Barium	100	+/-100	U	100	P	12/03/2024	18:47	LB133718
	Cadmium	6.00	+/-6.00	U	6.00	P	12/03/2024	18:47	LB133718
	Chromium	10.0	+/-10.0	U	10.0	P	12/03/2024	18:47	LB133718
	Lead	12.0	+/-12.0	U	12.0	P	12/03/2024	18:47	LB133718
	Selenium	20.0	+/-20.0	U	20.0	P	12/03/2024	18:47	LB133718
	Silver	10.0	+/-10.0	U	10.0	P	12/03/2024	18:47	LB133718
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	12/03/2024	19:56	LB133718
	Barium	100	+/-100	U	100	P	12/03/2024	19:56	LB133718
	Cadmium	6.00	+/-6.00	U	6.00	P	12/03/2024	19:56	LB133718
	Chromium	10.0	+/-10.0	U	10.0	P	12/03/2024	19:56	LB133718
	Lead	12.0	+/-12.0	U	12.0	P	12/03/2024	19:56	LB133718
	Selenium	20.0	+/-20.0	U	20.0	P	12/03/2024	19:56	LB133718
	Silver	10.0	+/-10.0	U	10.0	P	12/03/2024	19:56	LB133718
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	12/03/2024	20:51	LB133718
	Barium	100	+/-100	U	100	P	12/03/2024	20:51	LB133718
	Cadmium	6.00	+/-6.00	U	6.00	P	12/03/2024	20:51	LB133718
	Chromium	10.0	+/-10.0	U	10.0	P	12/03/2024	20:51	LB133718
	Lead	12.0	+/-12.0	U	12.0	P	12/03/2024	20:51	LB133718
	Selenium	20.0	+/-20.0	U	20.0	P	12/03/2024	20:51	LB133718
	Silver	10.0	+/-10.0	U	10.0	P	12/03/2024	20:51	LB133718
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	12/03/2024	21:24	LB133718
	Barium	100	+/-100	U	100	P	12/03/2024	21:24	LB133718
	Cadmium	6.00	+/-6.00	U	6.00	P	12/03/2024	21:24	LB133718
	Chromium	10.0	+/-10.0	U	10.0	P	12/03/2024	21:24	LB133718
	Lead	12.0	+/-12.0	U	12.0	P	12/03/2024	21:24	LB133718
	Selenium	20.0	+/-20.0	U	20.0	P	12/03/2024	21:24	LB133718
	Silver	10.0	+/-10.0	U	10.0	P	12/03/2024	21:24	LB133718

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Tully Construction Co., Inc.</u>				SDG No.:	<u>P5026</u>			
Contract:	<u>TULL02</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>P5026</u>	SAS No.:	<u>P5026</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number