

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Construction Co., Inc. **SDG No.:** P5026
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5026 **SAS No.:** P5026
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	-0.12			-20	20	12/03/2024	17:59	LB133718
	Barium	-0.78	6.0	13	-94	106	12/03/2024	17:59	LB133718
	Cadmium	5.59	1.0	559	-5	7	12/03/2024	17:59	LB133718
	Chromium	56.8	52.0	109	42	62	12/03/2024	17:59	LB133718
	Lead	7.43			-12	12	12/03/2024	17:59	LB133718
	Selenium	-13.7			-20	20	12/03/2024	17:59	LB133718
	Silver	-3.39			-10	10	12/03/2024	17:59	LB133718
ICSAB01	Arsenic	106	104	102	88.4	120	12/03/2024	18:13	LB133718
	Barium	478	537	89	437	637	12/03/2024	18:13	LB133718
	Cadmium	966	972	99	826	1120	12/03/2024	18:13	LB133718
	Chromium	537	542	99	460	624	12/03/2024	18:13	LB133718
	Lead	54.8	49.0	112	37	61	12/03/2024	18:13	LB133718
	Selenium	37.9	46.0	82	26	66	12/03/2024	18:13	LB133718
	Silver	188	201	94	170	232	12/03/2024	18:13	LB133718