

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Construction Co., Inc. **SDG No.:** P5026
Contract: TULL02 **Lab Code:** CHEM **Case No.:** P5026 **SAS No.:** P5026

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165310BS							
Arsenic	ug/L	4000	3920		98	80 - 120	P
Barium	ug/L	1000	914		91	80 - 120	P
Cadmium	ug/L	1000	940		94	80 - 120	P
Chromium	ug/L	2000	2010		100	80 - 120	P
Lead	ug/L	5000	4740		95	80 - 120	P
Selenium	ug/L	10000	9810		98	80 - 120	P
Silver	ug/L	380	372		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: P5026

Contract: TULL02

Lab Code: CHEM

Case No.: P5026

SAS No.: P5026

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165334BS Mercury	ug/L	4.0	3.77		94	80 - 120	CV